A-A-51534
8 July 1986

# COMMERCIAL ITEM DESCRIPTION

## SCREW, BONE (82° COUNTERSUNK HEAD)

The General Services Administration has authorized the use of this commercial item description.

This commercial item description covers 82° countersunk head bone pins with pilot points suitable for implanting bones.

Salient characteristics:

Characteristics. Bone screw shall be in accordance with Figure 2 of ASTM F543. Shall be No. 6M bone thread, 32 threads per inch. The bone screw shall be of the double-fluted, self-tapping design, as shown in ASTM F544 for medical screws possessing cruciate oval head. The tapping point shall be as shown in ASTM F544 for the double flute design. Radius of the flutes shall be 0.125 ± .015 inches.

Material. Bone screw shall be fabricated from cold drawn wire conforming to the composition specified in ASTM F138, Grade 2.

Surface preparation. Description of surface characteristics and methods of surface preparation shall be in accordance with ASTM F86.

Finish. The completed bone screw shall have a mirror finish, properly passivated.

Corrosion properties. Bone screws shall show no evidence of corrosion when subjected to the following tests:

Intergranular corrosion. The intergranular corrosion susceptibility test specified in ASTM F138.

Saline. Bone screws shall be completely immersed into Normal Saline Solution (USP), and while immersed, placed into an oven at a temperature of 37° ± 1°C for a period of seven days, removed from the oven and examined for corrosion while still immersed.

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 202610-26T08:57Z
Check the source to verify that this is the current version before use.

Boil. Bone screws shall be scrubbed with soap and water, rinsed thoroughly in distilled water, and dried. After cleaning, the instruments shall be boiled in distilled water in a glass beaker or 30 minutes, allowed to remain submerged for an additional 24 hours.

Workmanship. The bone screws shall be free from defects which detract from their appearance or impair their serviceability.

Product Quality Level.

Visual and dimensional examination. Shall be examined to verify that design, construction and workmanship are in accordance with applicable requirements.

Materials. The supplier shall be responsible for performing tests to assure that materials are in accordance with requirements specified.

Unit. Each (EA). One bone screw, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.



Source: https://assist.dla.mil -- Downloaded: 202640-26T08:57Z
Check the source to verify that this is the currentversion before use.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105. The acceptable quality level (AQL) shall be 1.0 for an inspection level of S-2. The unit of product for test purposes shall be bone screw.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.



Source: https://assist.dla.mil -- Downloaded: 202610-26T08:57Z
Check the source to verify that this is the current version before use.

NOTE: The following National Stock Numbers are covered by this document:

| NSN | ITEM IDENTIFICATION | INCHES LONG |
|---|---|---|
| 6515-00-366-3010 | SCREW, BONE, 82° Countersunk Head | 3/8 |
| 6515-00-366-3015 | SCREW, BONE, 82° Countersunk Head | 1/2 |
| 6515-00-366-3020 | SCREW, BONE, 82° Countersunk Head | 5/8 |
| 6515-00-366-3025 | SCREW, BONE, 82° Countersunk Head | 3/4 |
| 6515-00-366-3030 | SCREW, BONE, 82° Countersunk Head | 7/8 |
| 6515-00-366-3035 | SCREW, BONE, 82° Countersunk Head | 1 |
| 6515-00-366-3040 | SCREW, BONE, 82° Countersunk Head | 1-1/8 |
| 6515-00-366-3045 | SCREW, BONE, 82° Countersunk Head | 1-1/4 |
| 6515-00-366-3050 | SCREW, BONE, 82° Countersunk Head | 1-3/8 |
| 6515-00-366-3055 | SCREW, BONE, 82° Countersunk Head | 1-1/2 |
| 6515-00-366-3060 | SCREW, BONE, 82° Countersunk Head | 1-5/8 |
| 6515-00-366-3065 | SCREW, BONE, 82° Countersunk Head | 1-3/4 |
| 6515-00-366-3070 | SCREW, BONE, 82° Countersunk Head | 1-7/8 |
| 6515-00-366-3075 | SCREW, BONE, 82° Countersunk Head | 2 |
| 6515-00-366-3080 | SCREW, BONE, 82° Countersunk Head | 2-1/4 |

Ordering data (Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order).

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS

05313

Project No. 6515-3517



Source: https://assist.dla.mil -- Downloaded: 202640-26T08:57Z
Check the source to verify that this is the current version before use.

**INSTRUCTIONS:** In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

*NOTE.* This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

*(Fold along this line)*

*(Fold along this line)*

DEFENSE LOGISTICS AGENCY

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATE

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO 4966 WASHINGTON D C

POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
Attn: DPSC-RSTS
2800 South 20th Street
Philadelphia, PA 19101

Source: https://assist.dla.mil -- Downloaded: 202610-26T08:57Z
Check the source to verify that this is the current version before use.

# STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

*(See Instructions – Reverse Side)*

| 1. DOCUMENT NUMBER | 2. DOCUMENT TITLE |
|---|---|
| A-A-51534 | SCREW, BONE (82° COUNTERSUNK HEAD) |

| 3a. NAME OF SUBMITTING ORGANIZATION | 4. TYPE OF ORGANIZATION *(Mark one)* |
|---|---|
| | ☐ VENDOR |
| | ☐ USER |
| b. ADDRESS *(Street, City, State, ZIP Code)* | ☐ MANUFACTURER |
| | ☐ OTHER *(Specify)* _____ |

5. PROBLEM AREAS

a. Paragraph Number and Wording

b. Recommended Wording

c. Reason/Rationale for Recommendation

6. REMARKS

| 7a. NAME OF SUBMITTER *(Last, First, MI)* – Optional | b. WORK TELEPHONE NUMBER *(Include Area Code)* – Optional |
|---|---|
| c. MAILING ADDRESS *(Street, City, State, ZIP Code)* – Optional | 8. DATE OF SUBMISSION *(YYMMDD)* |

DD FORM 82 MAR 1426 PREVIOUS EDITION IS OBSOLETE

Source: https://assist.dla.mil - Downloaded: 2026-10-26T08:57Z
Check the source to verify that this is the current version before use.