A-A-51538
31 July 1986

COMMERCIAL ITEM DESCRIPTION

CUP, DENTAL AMALGAM

The General Services Administration has authorized the use of this commercial item description in preferance to Military Specification MIL-C-36843A.

This commercial item description covers a metal cup, without washer, suitable for holding amalgam.

Salient characteristics:

The metal cup shall be suitable for holding triturated amalgam to facilitate amalgam carrier loading. The design and dimensions of the cup shall conform to figure 1 as specified. The cup shall be cylindrical in shape. The external curved surface shall contain two parallel grooves, equidistant from each end to facilitate handling. The cup shall weigh a minimum of 150 grams.

Material. The cup shall be fabricated of class 3, type 303 stainless steel as specified in GG-I-526.

Finish. The cavity and the outside surface of the cup shall have a mirror or bright finish. The cavity shall be smoothly finished and free of grind marks to preclude entrapment of amalgam. The cup shall be furnished clean and ready for use.

Identification marking. The metal cups shall be permanently and legibly marked with the name or registered trademark of the manufacturer. Marking shall appear on the flat bottom surface of the cup.


Workmanship. The cup shall have no rough edges, burrs or cracks which detract from its appearance or impair its serviceability.

Unit. Each (EA). One cup, as specified, constitutes one unit.

AMSC N/A FSC 6520

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 202610-26T08:57Z
Check the source to verify that this is the current version before use.



NOTES: Unless otherwise specified:

(1) Break all sharp edges.

(2) Dimensions are in inches.

(3) Tolerances on fractions, ± 1/64 inch.

Figure 1. Cup, Dental Amalgam

Source: https://assist.dla.mil -- Downloaded: 202610-26T08:57Z
Check the source to verify that this is the current version before use.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Sampling for inspection. Sampling for inspection shall be conducted in accordance with MIL-STD-105. Where feasible, the same sample shall be used for the determination of two or more inspection characteristics. The unit of product for inspection purposes shall be one amalgam cup. Unless otherwise specified, sampling for examination shall be conducted at inspection level II, AQL 2.5, and sampling for tests (when applicable) shall be conducted at inspection level S-1, AQL 1.0.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Source: https://assist.dla.mil -- Downloaded: 202610-26T08:57Z
Check the source to verify that this is the current version before use.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractor/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packing, packaging, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

National Stock Number	Item Identification
6520-00-138-9748	CUP, DENTAL AMALGAM, Corrosion-Resisting Steel

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be as specified in the contract and/or order.

MILITARY INTERESTS:

Custodians:

Army-MD
Navy-MS
Air Force -03

PREPARING ACTIVITY:

DoD-MB

Agent:
DLA-DM

CIVIL AGENCY COORDINATING ACTIVITY

VA-OSS

Project No. 6520-1953

Source: https://assist.dla.mil -- Downloaded: 202640-26T08:57Z
Check the source to verify that this is the currentversion before use.