A-A-51542
2 August 1986

COMMERCIAL ITEM DESCRIPTION

TEST SET, ELECTRONIC CIRCUIT-COMPONENT

The General Services Adminstration has approved the use of this commercial item description.

This commercial item description covers an electronic circuit-component test set designed to test unpowered circuits and devices in maintenance support of medical equipment by detecting open circuits, short circuits, leakage and bonding problems, and unusual failure modes.

Salient characteristics:

The electronic circuit-component test set shall be comprised of the minimum following characteristics.

1. Unit shall operate on a selectable 120 or 230 Volts AC at a frequency of 50-60 Hertz, single phase. The selectable voltage switch shall be accessible with minimum effort by the operator.

2. Unit shall test resistance circuits/components for short and/or open circuits.

3. Unit shall test capacitors for short and/or open circuits and capacitor leakage. Shall test capacitors at a range from 1 picofarad to 25 microfarads (1×10^{-9} to 2.5×10^{-5} farads).

4. Unit shall test semi-conductor devices to include zenec diodes, bipolar junction transistors, triacs and opto-electronic devices for leakage, shorts and open bonding.

5. Unit shall test integrated circuits for component failure.

6. Unit shall have at least (3) selectable impedance ranges with automatic range scanning capabilities.

7. Unit shall nave at least (3) selectable frequency ranges.

AMSC N/A FSC 6625

DISTRIBUTION STATEMENT A. Approved for public release; distribution unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:57Z
Check the source to verify that this is the current version before use.

8. Unit shall have pulse generation capabilities with a selectable range from 100 to 4000 Hertz. Output amplitude shall be at least 0 to 5 volts peak, with a current up to 50 milli-amperes and an impedance of 100 ohms.

9. Unit selector switch(s)/terminals shall describe function, range and/or test. All controls shall be located on the front panel.

10. Unit shall be furnished with a protective cover or carrying case.

11. The units comparison display shall be of the catnode ray tube (CRT) type.

12. Unit shall not exceed 500 cubic inches and weigh no more than 10 pounds including the protective cover or case.

13. Unit shall have an operating temperature of 0 degrees to 50 degrees Celsius, a storage temperature of -50 degrees to +60 degrees Celsius and 0 to 70% relative humidity rating for storage and operation.

Workmanship. The electronic test set shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Each (EA). One test set, as specified, constitutes one unit.

QUALITY ASSURANCE PROVISIONS

Responsibility for inspection. Unless otherwise specified in the contract, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract, the contractor may use his own or any other facilities suitable for the performance of the inspection requirements specified herein unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:57Z
Check the source to verify that this is the current version before use.

Inspection. Inspection, as used herein, is defined as both examination (such as visual and auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item. Unless otherwise specified, sampling for inspection shall be conducted at inspection level II with an AQL of 2.5.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified herein. The unit of product for tests purposes shall be one Electronic Circuit-Component Test Set. Unless otherwise specified, sampling for tests shall be conducted at special inspection level S-1, with an AQL of 2.5.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more tests.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices . The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offer or/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:57Z
Check the source to verify that this is the current version before use.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Notes. The following National Stock Number is covered by this document:

NSN	ITEM IDENTIFICATION
6625-01-217-7608(W)	TEST SET, ELECTRONIC CIRCUIT-COMPONENT, 120/230 Volts, 50-60 Hertz, AC

Ordering data: (Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order).

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD-MP

Agent

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITY:

VA-OSS

Project No. 6625-0662

Location: CID-test-set-7608(W)
NCR-deb, disk 6

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:57Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS: In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE. This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

DEFENSE LOGISTICS AGENCY

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO 4986 WASHINGTON D C

POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
Attn: DPSC-RSTS
2800 South 20th Street
Philadelphia, PA 19101

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:57Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1. DOCUMENT NUMBER	2. DOCUMENT TITLE
A-A-51542	TEST SET, ELECTRONIC CIRCUIT-COMPONENT

3a. NAME OF SUBMITTING ORGANIZATION	4. TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR
	☐ USER
b. ADDRESS *(Street, City, State, ZIP Code)*	☐ MANUFACTURER
	☐ OTHER *(Specify)* _____

5. PROBLEM AREAS

a. Paragraph Number and Wording

b. Recommended Wording

c. Reason/Rationale for Recommendation

6. REMARKS

7a. NAME OF SUBMITTER *(Last, First, MI)* – Optional	b. WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
c. MAILING ADDRESS *(Street, City, State, ZIP Code)* – Optional	8. DATE OF SUBMISSION *(YYMMDD)*

(FOLD BACK THIS FORM, CUT ALONG THIS LINE.)

DD FORM 82 MAR 1426 PREVIOUS EDITION IS OBSOLETE

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:57Z
Check the source to verify that this is the current version before use.