A-A-51544
18 July 1986

COMMERCIAL ITEM DESCRIPTION

SYRINGE, FOUNTAIN

The General Services Administration has authorized the use of this commercial item description.

This commercial item description covers a disposable, empty plastic enema bag, without barium sulfate, with tubing and an air retention tip with cuff suitable for use in radiographic barium enema studies.

Salient characteristics:

The fountain syringe shall be a clean, disposable, transparent plastic enema bag, with tubing and air retention tip, capable of holding a minimum of 64 fluid ounces (1920 cc).

The bag shall have a leak-proof, screw-on cap, and provisions for hanging the bag during use. The bag shall be continuously heat sealed and leak-proof.

The bag shall have an attached, approximately 86 inches (218 cm) long, nominal 1/2 inch lumen transparent pliable tubing attached at the bottom and flush with the inside of the bag. The other end of the tubing shall be attached to a Miller style retention tip with a 100 cc capacity cuff. There shall be no obstruction of the bore of the tubing.

The Miller tip shall have an attached nominal 1/4 inch outside diameter tubing approximately 22-1/2 inches long, designed for patient air insufflation and an attached nominal 1/8 inch outside diameter tubing, approximately 20-1/2 inches long, designed to introduce air into the air retention cuff. The tubes shall be designed to accept an air bulb patient insufflator and a one-shot cuff inflator, respectively. The inflation devices shall not be attached to these tubes and should be easily recognized for the purpose intended by shape and color.

Both the nominal 1/4 inch outside diameter and the 1/2 inch lumen tubing shall be individually equipped with slide-type clamps, made of non-corrosive material, and capable of single-handed activation for cessation of flow without breakage or distortion of the clamps or incision of the tubing.

AMSC N/A FSC 6530

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:57Z
Check the source to verify that this is the current version before use.

All components shall contain no sharp edges that may injure the body during use or any ingredient known to be a skin irritant.

One complete fountain syringe shall be individually packaged in a sealed plastic bag, free from cuts and tears. The contents of the sealed plastic bag shall be clean and contain no foreign matter.

Each sealed plastic bag containing one syringe shall contain complete and clear instructions for use.

Workmanship. The fountain syringe and components shall be free from defects which detract from their appearance or impair their serviceability.

Unit. Package (PG). One package containing 24 individually packaged fountain syringes, plus two one-shot cuff inflators and one air bulb patient insufflator, as specified, constitutes one unit.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:57Z
Check the source to verify that this is the current version before use.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

6530-01-168-7864

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order.

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD - MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS

PGC 54233

Project No. 6530-1856

Location: CID/SYRINGE-FOUNTAIN/H13

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:57Z
Check the source to verify that this is the current version before use.