A-A-0051545(DM)
14 July 1986

INTERIM COMMERCIAL ITEM DESCRIPTION

HUMIDIFIER, INHALATION THERAPY APPARATUS

The General Services Administration has authorized the use of this interim commercial item description which was developed by the Defense Personnel Support Center, Directorate of Medical Materiel, DPSC-RST, 2800 South 20th Street, Philadelphia, PA 19101, based upon currently available technical information.

This interim commercial item description covers a single-patient-use bubble humidifier for use in the administration of inhalation therapy.

Salient characteristics:

The humidifier shall be disposable for single patient use, bubble type, intended for use with oxygen. It shall consist of, at least, a liquid reservoir/jar, humidifier head/cap, and oxygen/gas diffuser.

The unit shall be shatter-resistant. Nylon shall not be employed in the fabrication of components which are intended to be in direct communication with oxygen. Allowed materials are plastic type materials, corrosion resistant metal, and metal treated for corrosion resistance. All materials shall be suitable for intended use. The unit shall be capable of withstanding storage in uncontrolled environmental conditions (temperature and humidity) for extended periods of time (5 years); or, if the humidifier can only be stored in a controlled environment, the offeror shall submit information with the bid/proposal which indicates the environmental limits within which the unit should be stored.

The reservoir shall have an operating capacity of 400 ml. Markings on the reservoir shall conform with ANSI Z79.9. The reservoir shall permit visual observation of liquid level.

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:57Z
Check the source to verify that this is the current version before use.

The humidifier head shall include a D.I.S.S. gas inlet conforming to CGA 1240 for oxygen, a 2 psi relief valve conforming to ANSI Z79.9, and a combination 3/16 - 1/4 inch tubing connector as the aerosol outlet. The head shall be capable of being removed/attached easily for purpose of refilling. The relief valve shall incorporate an audible warning device which shall provide an audible warning signal should the aerosol tubing circuit become occluded when oxygen/gas flow is as low as 3 LPM. This audible signal shall not sound when the outlet tube circuitry is not obstructed.

The unit shall be intended for use with a noncomponent flowmeter placed between a compressed gas supply (45 to 60 psig) and the unit, but shall be operational when used without a flowmeter. The unit shall be intended to humidify oxygen with oxygen flows from a minimum flow of, at most, 3 LPM to a maximum flow, at least, 10 LPM. The humidifier, and its components, shall be in compliance with ANSI Z79.9.

Workmanship. The humidifier shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Package (PG). One package containing 50 humidifiers, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contact or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Source: https://assist.dla.mil -- Downloaded: 202640-26T08:57Z
Check the source to verify that this is the currentversion before use.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105, and as indicated in table I. The unit of product for examination purposes shall be one humidifier.

Table I. Sampling for Examination.

	Inspection level	AQL (Percent defects)
Major defects	II	1.0
Total defects	II	3.0

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified in table II. Where feasible, the same sample shall be used for the determination of two or more test characteristics. The unit of product for test purposes shall be one humidifier.

Table II. Sampling for tests.

Characteristic	Inspection Level	AQL (Percent defects)
Reservoir capacity	1 unit/contract Qty	0.0
Overpressure relief	S-1	1.5
Warning Signal (aerosol circuitry occlusion)	S-1	1.5
Flow rates	S-1	1.5
Fluid outputs	S-1	1.5

Examination. The humidifier shall be examined to determine compliance with the salient characteristics specified herein. Where feasible, the same sample shall be used for the determination of two or more test characteristics. Examination shall include, but shall not be limited to, the classifications of the defects listed in table III.

Source: https://assist.dla.mil -- Downloaded: 202610-26T08:57Z
Check the source to verify that this is the current version before use.

Table III. Classification of defects.

Categories	Defects
Major	
101	Humidifier not bubble type.
102	Absence of relief valve.
103	Head not easily removed/installed.
104	Reservoir marking not as specified.
105	Gas inlet not D.I.S.S., as specified.
106	Aerosol outlet does not accept 3/16 inch and 1/4 inch ld tubing.
107	Components not free of cracks, burrs, sharp edges, corrosion.
Minor	
201	Components not free of pits or unintended mold marks.

Tests. Tests shall be conducted to determine compliance with salient characteristics specified herein. Where feasible, the same sample shall be used for the determination of two or more test characteristics. A test failure shall be a major defect. All tests shall be conducted in accordance with procedures specified by ANSI Z79.9-1979. The warning signal shall be audible only when the aerosol outlet is occluded and shall be audible when flow rate is set at 3 LPM while the aerosol outlet is occluded.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:57Z
Check the source to verify that this is the current version before use.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

6515-01-098-8264

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order.

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD - MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

PGC 59492

Project No. 6515-3771

Location: PC0005.H12-3

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:57Z
Check the source to verify that this is the current version before use.