A-A-51545B
24 March 1989
SUPERSEDING
A-A-51545A
31 August 1987

COMMERCIAL ITEM DESCRIPTION

HUMIDIFIER, INHALATION THERAPY APPARATUS

The General Services Administration has authorized the use of this commercial item description.

This commercial item description covers a single-patient-use bubble humidifier for use in the administration of inhalation therapy.

Salient characteristics:

The humidifier shall be disposable for single patient use, bubble type, and intended for use with oxygen. It shall consist of, at least, a liquid reservoir/jar, humidifier head/cap, and oxygen/gas diffuser.

Allowed materials are plastic type materials, corrosion resistant metal, and metal treated for corrosion resistance. All materials shall be suitable for intended use. The unit shall be capable of withstanding storage in uncontrolled environmental conditions (temperature and humidity) for extended periods of time (5 years).

The reservoir shall have an operating capacity of 400 ml. Markings on the reservoir shall conform with ANSI Z79.9. The reservoir shall permit visual observation of liquid level.

The humidifier head shall include a D.I.S.S. gas inlet conforming to CGA 1240 for oxygen, a 2 psi relief valve conforming to ANSI Z79.9, and a combination 3/16 - 1/4 inch tubing connector as the outlet. The head shall be capable of being removed/attached easily for purpose of refilling. The 2 psi relief valve shall provide an audible noise to warn if the tubing circuit becomes occluded. This relief valve shall not activate when the outlet tubing circuitry is not obstructed and 10 LPM oxygen/gas is flowing.

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:57Z
Check the source to verify that this is the current version before use.

The humidifier shall be intended for use with a noncomponent flowmeter placed between a compressed gas supply (45 to 60 psig) and the unit. The unit shall be intended to humidify oxygen with flows of 3 LPM to 10 LPM. The fluid output shall include a range from 10 mg H_2O/liter at 10 LPM to 15 mg H_2O/liter at 3 LPM. The mean value of the specified fluid output shall occur within a flow range, 5.5-7.5 LPM. The humidifier, and its components, shall be in compliance with ANSI Z79.9.

Workmanship. The humidifier shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Package (PG). One package containing 50 humidifiers, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contact or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:57Z
Check the source to verify that this is the current version before use.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105, with an AQL of 1.0 (percent defective) and an inspection level of II. The unit of product for examination purposes shall be one humidifier as specified in this description.

Defects.

Humidifier not bubble type.

Absence of relief valve.

Head not easily removed/installed.

Reservoir marking not as specified.

Aerosol outlet does not accept 3/16 inch and 1/4 inch ID tubing.

Components not free of cracks, burrs, pits sharp edges, or corrosion.

Components not free of pits or unintended mold marks.

Sampling for tests. Sampling for the following tests shall be conducted in accordance with MIL-STD-105 with an AQL of 1.5 (percent defective) and an inspection level of S-1:

Overpressure relief
Warning signal
(aersol circuitry occlusion)
Flow rates
Fluid outputs

Sampling for the reservior capacity test shall also be in accordance with MIL-STD-105, with no defects permitted.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:57Z
Check the source to verify that this is the current version before use.

Tests. Tests shall be conducted to determine compliance with salient characteristics specified herein. Where feasible, the same sample shall be used for the determination of two or more test characteristics. A test failure shall be a defect. All tests shall be conducted in accordance with procedures specified by ANSI Z79.9. The warning signal shall be audible only when the outlet is occluded and shall be audible when flow rate is set at 3 LPM while the outlet is occluded.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch/pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:57Z
Check the source to verify that this is the current version before use.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended, and as implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

6515-01-098-8264

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order.

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD - MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project No. 6515-4403

Location: CID/HUMIDIFIER/H12-3

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:57Z
Check the source to verify that this is the current version before use.