A-A-0051546 (DM)
10 July 1986

# INTERIM COMMERCIAL ITEM DESCRIPTION

## DRAIN, CLOSED SUCTION (Jackson-Pratt)

The General Services Administration has authorized the use of this interim commercial item description which was developed by the Defense Personnel Support Center, Directorate of Medical Materiel, DPSC-ATT, 2800 South 20th Street, Philadelphia, PA, 19101, based upon currently available technical information.

This interim commercial item description covers two sizes of Jackson-Pratt style drains for use with a hand-held aspirator-type reservoir in providing suction of closed wounds.

Salient characteristics.

The drain shall consist of a length of round tubing bonded to a flattened section. Drains shall be of the following sizes, as specified:

| Sizes | Flat Portion Width (± 1 mm) |
|---|---|
| 1 | 7 mm |
| 2 | 10 mm |

The drain shall be manufactured from compounded silicone rubber. The tubing shall be 79 ± 8 cm in length and shall have an inside diameter of 2.4 - 3.2 mm (3/32 - 1/8 inch). The flattened section shall have 3-4 rows of perforations into the lumen of the drain with a minimum of 60 perforations. The perforations shall be 1-2 mm in diameter. The rows of perforations shall be placed within the distal 75 percent of the flattened section. The flattened section shall have a radiopaque stripe throughout its length and shall have ribs on the inside surface. The juncture of the tubing to the flattened section shall be smoothly tapered to facilitate atraumatic withdrawal of the drain from a wound.

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 202610-26T08:57Z
Check the source to verify that this is the current version before use.

Physical properties. Tubing shall be tested for breaking strength in accordance with Federal Test Method 4111 of Federal Test Method Standard 601. Breaking strength shall be the value of breaking force as defined in Method 4111. Drains shall have a minimum breaking strength of 32 pounds force.

In addition, drains shall be subjected to accelerated aging in accordance with Federal Test Method 7221 of Standard Number 601. The aged drains shall be tested for breaking strength as specified above. The minimum breaking strength of the aged drains shall be 26 pounds force.

Sterility and freedom from pyrogens. Drains shall be sterile and free from pyrogens. Drains shall be tested for sterility and freedom from pyrogens in accordance with the United States Pharmacopeia (USP).

Biocompatibility. Material used to manufacture the drains shall be tested in accordance with the Intracataneous and Intramuscular Implantation tests of the USP.

Workmanship. The drains shall be free from defects which detract from their appearance.

Unit. Package (PG). One package, containing ten individually packaged drains, constitutes one unit. Each drain shall be packaged sterile in a sealed peel-open type container capable of maintaining sterility of contents unless the container is damaged or opened.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.



Source: https://assist.dla.mil -- Downloaded: 202640-26T08:57Z
Check the source to verify that this is the currentversion before use.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Certificates of Biocompatibility. A certificate of compliance, stipulating that offered items meet the biocompatibility test, shall be acceptable, provided that the contractor has supplied acceptable items to the Government within the past three years, and that there have been no changes to material formulation and method of processing since actual testing was performed. The certificate shall include the date of actual testing, the test method and results, the name and title of the person evaluating results, the current contract number and NSN, a statement to the effect that material processing and formulation have not changed since actual testing was performed, and the signature and title of the person making the certification.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105, inspection Level II, AQL 2.5 (% defective). The unit of the product shall be one drain.

Examination. The drain shall be examined to determine compliance with all requirements contained in this specification.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for physical properties test. Sampling for physical properties test shall be conducted in accordance with MIL-STD-105, sampling level S-1, AQL 1.5 and as specified herein. The unit of product for test purposes shall be one drain.



Source: https://assist.dla.mil -- Downloaded: 202610-26T08:57Z
Check the source to verify that this is the current version before use.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Numbers are covered by this document:

| NSN | ITEM IDENTIFICATION |
|---|---|
| 6515-01-216-3368 | DRAIN, CLOSED, SUCTION, Jackson-Pratt, Silicone, Radiopaque, 7 mm Wide, Flat, Perforations 3/4 Length |
| 6515-01-219-2756 | DRAIN, CLOSED, SUCTION, Jackson-Pratt, Silicone, Radiopaque, 10 mm Wide, Flat, Perforations 3/4 Length |



Source: https://assist.dla.mil -- Downloaded: 202610-26T08:57Z
Check the source to verify that this is the current version before use.

Ordering data. (Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order).

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS

PGC SF006

Project No. 6515-3782



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:57Z
Check the source to verify that this is the current version before use.