A-A-0051551 (DM)
16 July 1986

# INTERIM COMMERCIAL ITEM DESCRIPTION

## BLOOD RECIPIENT SET, INDIRECT TRANSFUSION

The General Services Administration has authorized the use of this interim commercial item description which was developed by the Defense Personnel Support Center, Directorate of Medical Materiel, DPSC-ATT, 2800 South 20th Street, Philadelphia, PA, 19101, based upon currently available technical information.

This interim commercial item description covers a sterile indirect transfusion blood recipient set for use with standard blood collecting, dispensing bags and flexible and rigid parenteral solution containers.

Salient characteristics.

Shall be a non toxic, disposable "Y" blood transfusion set. The set shall consist of two non-vented piercing devices each connected by 11 inches of tubing to a filter and drip chamber which shall be connected by tubing to a distal latex flashback/injection site. Overall length shall be approximately 82 inches. The flashback/injection site shall be connected to a male needle adapter. Roller-flow control devices shall be provided in the tubing of each leg of the "Y" arrangement. Easily removed plastic end protectors shall be provided on the spikes and the male needle adapter to maintain flow path sterility. Tubing and clamps shall be fabricated from PVC and polypropylene respectively.

Shall be capable of being used with standard blood collecting and dispensing bags, and all standard flexible and semi rigid parenteral solution containers.

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 202610-26T08:57Z
Check the source to verify that this is the current version before use.

The piercing spike for blood bags shall be formed or molded in such a way to function as an automatic stop. The tip shall be designed with either a side opening, distal opening, or three ribbed openings, and shall be of sufficient length to rupture the inlet seal on the existing blood bags; however, not of such length as to reach far enough into the bag as to perforate the side walls during insertion. Piercing spike shall be fabricated from modified acrylic.

The set shall require approximately 22 ml for priming and shall deliver approximately 10 drops per ml.

The drip chamber shall be a single unit consisting of a drip device and filter unit or, as an alternate, the drip and filter chambers may be separate. The filter shall be fabricated of nylon mesh and shall have a minimum filtering area of 4.33 sq. inches (27.93 sq cm). Additionally, the filter shall conform to the requirements of Standard Sieve Numbers 70 to 100 of the current USP filtering particles 170 mirons or larger. Drip chamber end caps shall be fabricated from modified acrylic.

Tubing shall not separate from attached components when subjected to an axial force of 3 pounds for 2 minutes.

Shall meet the requirements for transfusion and infusion assemblies and the added requirements for Class II plastics as stated in the current United Stated Pharmacopeia.

Shall have a sterile, pyrogen-free fluid path completely sealed to maintain sterility. Shall meet the sterility requirements when tested in accordance with the USP.

The male needle adapter shall be fabricated of a suitable translucent material and shall be tapered to fit a needle having a standard Luer hub in accordance with ANSI/HIMA Standard MD70.1.

Each set shall be packaged in a commercial available immediate container acceptable to the Food and Drug Administration (FDA) for the product contained therein and shall be of such compatibility that neither the contents nor the container are altered in any way by each other.

Workmanship. The blood recipient set shall be free from defects which detract from its appearance or impair its serviceability. Set components shall be free from cuts, tears, flash or weak spots.



Source: https://assist.dla.mil -- Downloaded: 202610-26T08:57Z
Check the source to verify that this is the currentversion before use.

Unit. Package (PG). One package containing forty-eight individually packaged sets along with instructions for use, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified herein. The unit of product for test purposes shall be one blood recipient set.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:57Z
Check the source to verify that this is the current version before use.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

| NSN | ITEM IDENTIFICATION |
|---|---|
| 6515-01-151-2337 | BLOOD RECIPIENT SET, Indirect Transfusion Type, Non-Vented, With Filter and Drip Chamber |

Ordering data. (Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order).



Source: https://assist.dla.mil -- Downloaded: 202640-26T08:57Z
Check the source to verify that this is the currentversion before use.

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS

Project No. 6515-3784

☆U.S. GOVERNMENT PRINTING OFFICE: 1986 — 605-036/44776



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:57Z
Check the source to verify that this is the current version before use.