A-A-51551A
20 February 1987
SUPERSEDING
A-A-0051551(DM)
16 July 1986

COMMERCIAL ITEM DESCRIPTION

BLOOD RECIPIENT SET, INDIRECT TRANSFUSION

The General Services Administration has authorized the use of this Commercial Item Description.

This commercial item description covers a sterile indirect transfusion blood recipient set for use with standard blood collecting, dispensing bags and flexible and rigid parenteral solution containers.

Salient characteristics.

Materials. Rubber injection sites shall be of the best quality and suitable in all respects for injection and resealing after needle withdrawal. The tubing shall be clear, plasticized polyvinyl chloride with a Shore Durometer A hardness of 78 ± 5. The drip chamber shall be clear and flexible polyvinyl chloride with a Shore Durometer A hardness of 96 + 5, -10. All plastic and rubber components shall be sterile and pyrogen-free. The male Luer taper and end protectors shall be made of suitable rigid plastic. Clamps shall be fabricated from polypropylene and the spike from modified acrylic or plexiglas DR. Drip chamber end caps shall be fabricated from modified acrylic and shall be properly sealed to filter housing.

Construction. Shall be a non toxic, disposable "Y" blood transfusion set. The set shall consist of two non-vented piercing devices each connected by 11 inches of tubing to a filter and drip chamber which shall be connected by tubing to a distal latex flashback/injection site. Overall length shall be 82 plus 3 minus 1 inches. The flashback/injection site shall be connected to a male needle adapter. Roller-flow control devices shall be provided in the tubing of each leg of the "Y" arrangement. Easily removed plastic end protectors shall be provided on the spikes and the male needle adapter to maintain flow path sterility. The tubing shall have an inside diameter of 0.120 ± .003 inch and an outside diameter of 0.164 ± .003 inch. The tubing from the two spikes shall blend into a single drip orifice in the drip chamber.

AMSC N/A FSC 6515
DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 202610-26T08:57Z
Check the source to verify that this is the current version before use.

The set shall be capable of being used with standard blood collecting and dispensing bags, and all standard flexible and semi rigid parenteral solution containers.

The piercing spikes shall be formed or molded in such a way to function as an automatic stop. The tip shall be designed with either a side opening, distal opening, or three ribbed openings, and shall be of sufficient length to rupture the inlet seal on the existing blood bags; however, not of such length as to reach far enough into the bag as to perforate the side walls during insertion.

The spike shall be 0.908 $\pm$ .03 inches long from the tip to the shoulder and shall have a diameter of 0.195 $\pm$ .005 inch just below the tip and a diameter of 0.210 $\pm$ .005 inch just above the shoulder. The spike shall be made with a sharp point and shall withstand penetration through rubber or plastic stoppers of parenteral solution containers or blood bag seals without damage.

The set shall require 40 $\pm$ 5 ml for priming and shall deliver 10 $\pm$ 2 drops per ml.

The drip chamber shall be a single unit consisting of a drip device and filter unit or, as an alternate, the drip and filter chambers may be separate. The filter shall be fabricated of nylon or polyethylene mesh and shall have a minimum filtering area of 4.33 sq. inches (27.93 sq cm). Additionally, the filter shall conform to the requirements of Standard Sieve Numbers 70 to 100 of the current USP filtering particles 170 microns or larger.

Tubing shall not leak or separate from attached components when subjected to an axial force of 3 pounds for 2 minutes.

The complete set shall withstand a 20 psig water pressure for five minutes without demonstrating leakage of more than one falling drop.

The roller shall not be forced out of the clamp body under a 3.65 kg load applied to the top of the roller and toward the open end of the clamp.

The injection site shall reseal without a falling drop after puncture and withdrawal of an 18-gage unlubricated needle. The resealing shall occur within 10 seconds of needle withdrawal. An internal water pressure of 4 psig shall be used in the conduct of this test.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:57Z
Check the source to verify that this is the current version before use.

The set shall meet the requirements for transfusion and infusion assemblies and the added requirements for Class II plastics as stated in the current United States Pharmacopeia.

The set shall have a sterile, pyrogen-free fluid path completely sealed to maintain sterility. Shall meet the sterility requirements and be free from pyrogens when tested in accordance with the USP.

The male needle adapter shall be tapered to fit a needle having a standard Luer hub in accordance with ANSI/HIMA Standard MD70.1.

Each set shall be packaged in a commercial available immediate container acceptable to the Food and Drug Administration (FDA) for the product contained therein and shall be of such compatibility that neither the contents nor the container are altered in any way by each other. The package shall maintain the sterility of the set until the package is opened.

Workmanship. The blood recipient set shall be free from defects which detract from its appearance or impair its serviceability. Set components shall be free from cuts, tears, flash or weak spots.

Unit. Package (PG). One package containing forty-eight to fifty individually packaged sets along with instructions for use, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Source: https://assist.dla.mil -- Downloaded: 20261O-26T08:57Z
Check the source to verify that this is the currentversion before use.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified herein. The unit of product for test purposes shall be one blood recipient set.

Sampling Criteria for Tests.

	Inspection	AQL
Tubing separation test	S-3	0.65
Pressure test for leakage	S-3	0.65
Security of roller in clamp body	S-3	0.65
Resealing of injection sites	S-3	0.65

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Source: https://assist.dla.mil -- Downloaded: 202640-26T08:57Z
Check the source to verify that this is the currentversion before use.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

NSN	ITEM IDENTIFICATION
6515-01-151-2337	BLOOD RECIPIENT SET, Indirect Transfusion Type, Non-Vented, With Filter and Drip Chamber

Ordering data. (Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order).

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project No. 6515-3864

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:57Z
Check the source to verify that this is the current version before use.