METRIC

A-A-51552A
15 January 1993
SUPERSEDING
A-A-51552
16 July 1986

COMMERCIAL ITEM DESCRIPTION

AIRWAY, PHARYNGEAL (Guedel)

The General Services Administration has authorized the use of this Commercial Item Description.

This Commercial Item Description covers two sizes of Guedel design disposable, plastic pharyngeal airways.

Salient characteristics:

General. Shall be Guedel design, disposable pharyngeal airways to facilitate pharyngeal suctioning and to accommodate endotracheal tube positioning.

Style and design. Shall be of the following sizes:

Size 6 - 60 millimeter (mm), overall length

Size 10 - 100 millimeter (mm), overall length

Both airway configurations shall be in accordance with ANSI Z79.3 Standards (Figure 1a) for Anesthetic Equipment - Oropharyngeal And Nasopharyngeal Airways.

Materials. Airways shall be fabricated from polyethylene (LDPE) with EVAC (Ethylene Vinyl Acetate) or from polyvinylchloride (PVC) material.

Identification marking. Airway size and manufacturer's name shall be marked on the outer face of the buccal flange using raised letters. Marking shall conform to ANSI Z79.3.

Finish. Airways shall have a smooth, uniform finish conforming to ANSI Z79.3. All edges (internal and external) shall be suitably rounded or dulled.

AMSC N/A **FSC 6515**

DISTRIBUTION STATEMENT A: Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:57Z
Check the source to verify that this is the current version before use.

Workmanship Airways shall be free from warpage, weld-lines, shorts, pits, pimples, fish-eyes, blisters, shrink-marks, blooms and any other defects which detract from their appearance or impair their serviceability Definitions of defects listed above shall be in accordance with ASTM D883, Standard Definitions of Terms Relating to Plastics'

Airways shall be supplied clean and dry

Unit Package (PG) One package containing thirty (30) size 10 individually packaged airways or fifty (50) size 6 individually packaged airways as specified, constitutes one unit

QUALITY ASSURANCE PROVISIONS

Responsibility for inspection Unless otherwise specified in the contract of purchase order the contractor is responsible for the performance of all inspection requirements as specified herein Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government The Government reserves the right to perform any of the inspections set forth in the specification where such inspection are deemed necessary to assure supplies and services conform to prescribed requirements.

Records Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate

Inspection Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Sampling for inspection Sampling for inspection of all material properties shall be conducted in accordance with ASTM D1898, except that the Acceptable Quality Level (AQL) and Inspection Level shall be 1.0 percent and S-2 per MIL-STD-105.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:57Z
Check the source to verify that this is the current version before use.

Sampling for inspection of all other salient characteristics shall be conducted in accordance with MIL-STD-105 at Level II, AQL 1.0 percent. For this portion of the inspection, the unit of product for inspection purposes shall be one airway.

Examination Examination shall be conducted to determine compliance with specification requirements

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105 and as specified. Unit of product for examination purposes shall be one airway as specified. Sampling shall be general inspection level II with an Acceptable Quality Level (AQL) of 1.0.

Contractor certification The contractor shall certify and maintain substantiating evidence that the product offered meets the salient characteristics of this Commercial Item Description, and that the product conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerance specified using conversion tables contained in the latest revision of Federal Standard No 376, and all other requirements of this document are met

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch-pound units, a request should be made to the contracting officer to determine if the product is acceptable

The contracting officer has the option of accepting or rejecting the product

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:57Z
Check the source to verify that this is the current version before use.

Regulatory requirements.

Federal Food, Drug, and Cosmetic Act If the product covered by this document has been determined by the U S Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23 4 to the maximum extent practical

Preservation, packaging, packing, labeling, and marking Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc , involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity and shall conform to applicable carrier's rules and regulations Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order

Ordering data Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order

NOTE The following National Stock Numbers, Item Identifications and their respective Sizes are covered by this document:

NATIONAL STOCK NUMBERS	ITEM IDENTIFICATIONS	SIZES
6515-01-164-8885	AIRWAY, PHARYNGEAL, Guedel, Plastic, Disposable, 6.0 cm, 50's	6
6515-01-164-9637	AIRWAY, PHARYNGEAL, Guedel, Plastic, Disposable, 10.0 cm, 30's	10

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:57Z
Check the source to verify that this is the current version before use.

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
USPHS
FDA-MPQAS

Project Number: 6515-3957
Location: c:\word\doc\seto\AIRGCID

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:57Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

INSTRUCTIONS

1 The preparing activity must complete blocks 1, 2, 3, and 8 In block 1, both the document number and revision letter should be given

2 The submitter of this form must complete blocks 4, 5, 6, and 7

3 The preparing activity must provide a reply within 30 days from receipt of the form

NOTE This form may not be used to request copies of documents, nor to request waivers, or clarification of requirements on current contracts Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements

I RECOMMEND A CHANGE:	1 DOCUMENT NUMBER A-A-51552A	2 DOCUMENT DATE (YYMMDD) 15 January 1993

3 DOCUMENT TITLE
AIRWAY, PHARYNGEAL (Guedel)

4 NATURE OF CHANGE *(Identify paragraph number and include proposed rewrite if possible Attach extra sheets as needed)*

5 REASON FOR RECOMMENDATION

6 SUBMITTER

a NAME (Last, First, Middle Initial)	b ORGANIZATION	
c ADDRESS (Include Zip Code)	d TELEPHONE (Include Area Code) (1) Commercial (2) AUTOVON (If applicable)	7 DATE SUBMITTED (YYMMDD)

8 PREPARING ACTIVITY

a NAME Defense Personnel Support Center ATTN: DPSC-MSE	b TELEPHONE (Include Area Code) (1) Commercial 215-737-2870	(2) AUTOVON 444-2870
c ADDRESS (Include Zip Code) 2800 South 20th Street Philadelphia, PA 19145	IF YOU DO NOT RECEIVE A REPLY WITHIN 45 DAYS, CONTACT Defense Quality and Standardization Office 5203 Leesburg Pike Suite 1403, Falls Church, VA 22041-3466 Telephone (703) 756-2340 AUTOVON 289-2340	

DD Form 1426, OCT 89 *Previous editions are obsolete* '98

Source: https://assist.dla.mil - Downloaded: 2026-10-26T08:57Z
Check the source to verify that this is the current version before use.