A-A-51553
28 July 1986

COMMERCIAL ITEM DESCRIPTION

SPLINT, WOOD

The General Services Administration has authorized the use of this commercial item description in preference to Federal Specification GG-S-621C.

This Commercial Item Description covers wood splints intended to hold injured or fractured limbs in place.

Salient characteristics:

The splints shall be made of basswood, either white wood (sapwood) or heartwood. The wood shall be clean, sound, free from checking, splits, shakes, pitchpockets, fungus or dry rot. Mineral stain is permissible.

Pin knots over 1/8 inch in diameter shall not exceed four per linear foot. Knots under 1/8 inch in diameter shall not exceed twelve per linear foot.

The splints shall be sawn flat wood of the following sizes and dimensions (inches):

Size 1 - 18 $\pm$ 1/8 long by 4 $\pm$ 1/8 wide by 3/16 $\pm$ 1/64 thick.
Size 2 - 72 $\pm$ 1/8 long by 4 $\pm$ 1/8 wide by 3/16 $\pm$ 1/64 thick.

Deviation from flatness in the longitudinal direction shall not exceed 1/16 inch for Size 1 and 1/4 inch for Size 2 splints. Deviation from flatness in the transverse direction up to 1/16 inch shall be permitted.

All edges of splints (sides and ends) shall be smooth and free of splinters. Splints shall be dressed on two sides to a clean, smooth, flat surface.

The basswood shall have a moisture content of between 8 and 15 percent when tested per ASTM-D-2016.

Reference specification. Where ASTM-D-2016 and Mil-Std-105 are referenced herein, the issue in effect on the date of solicitation shall be used to determine compliance with stated requirements.

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:57Z
Check the source to verify that this is the current version before use.

Workmanship. The splints shall be free from defects which detract from their appearance or impair their serviceability.

Unit. Size 1: Package (PG). One package containing twelve splints, as specified, constitutes one unit.

Size 2: Each (EA). One splint, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Inspections and tests. Inspections and tests shall be conducted to determine compliance with specification requirements.

Sampling for inspection. Sampling for inspection shall be conducted in accordance with commercial practice. The unit of product for inspection purposes shall be one splint.

Sampling for tests. Sampling for moisture test shall be in accordance with MIL-STD-105, Inspection Level S-2, except that a minimum of five samples shall be taken and the acceptable quality level shall be 1.0. The test specimens shall be selected from different splints.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:57Z
Check the source to verify that this is the current version before use.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Numbers are covered by this document:

6515-00-372-1200	SPLINT, WOOD, 18 inch X 4 inch X 3/16 inch, 12s
6515-00-372-1310	SPLINT, WOOD, 6 feet X 4 inch X 3/16 inch

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:57Z
Check the source to verify that this is the current version before use.

Ordering data (Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order).

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITY:

VA-OSS

Project No. 6515-3690
Location: CID/SPLINT, WOOD
NCR, disk 6

☆U.S. GOVERNMENT PRINTING OFFICE: 1988 — 606-036/45081

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:57Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1. DOCUMENT NUMBER	2. DOCUMENT TITLE
A-A-51553	SPLINT, WOOD

3a. NAME OF SUBMITTING ORGANIZATION	4. TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR ☐ USER ☐ MANUFACTURER ☐ OTHER *(Specify)*: _____
b. ADDRESS *(Street, City, State, ZIP Code)*	

5. PROBLEM AREAS

a. Paragraph Number and Wording:

b. Recommended Wording:

c. Reason/Rationale for Recommendation:

6. REMARKS

7a. NAME OF SUBMITTER *(Last, First, MI)* – Optional	b. WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
c. MAILING ADDRESS *(Street, City, State, ZIP Code)* – Optional	8. DATE OF SUBMISSION *(YYMMDD)*

(TO DETACH THIS FORM, CUT ALONG THIS LINE.)

DD FORM 62 MAR 1426

PREVIOUS EDITION IS OBSOLETE.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:57Z
Check the source to verify that this is the current version before use.