A-A-51562
28 July 1986

COMMERCIAL ITEM DESCRIPTION

FORMER SET, DENTAL

The General Services Administration has authorized the use of this commercial item description.

This commercial item description covers two sizes of former sets suitable for forming the base and art part of an adult maxillary and mandibular diagnostic model.

Sizes - Medium
Large

Salient characteristics:

Materials. The former set shall be fabricated from an elastic and flexible material such as a natural rubber compound, so that model can be easily removed and have a smooth finish.

Style, design, and dimensions: The former set shall be a dental diagnostic model former set suitable for forming the base and art part of an adult maxillary and mandibular diagnostic model. Shall be capable of producing formed bases such that occluded maxillary and mandibular models may be positioned upright and laterally upon a smooth, flat surface without disarticulation so as to permit study of centric occlusion from both frontal and lateral aspects. The former set shall conform to the dimensions in table I, for the size indicated.

TABLE I.

MEDIUM

	Inside dimension @ heel (inches)	Anterior-posterior inside (inches) dimension (at mid-line of base)
Maxillary	3-1/4 ± 1/16	2-3/8 ± 1/16
Mandibular	3-3/16 ± 1/16	2-1/4 ± 1/16

LARGE

	Inside dimension @ heel (inches)	Anterior-posterior inside (inches) dimension (at mid-line of base)
Maxillary	3-3/4 ± 1/16	2-5/8 ± 1/16
Mandibular	3-5/8 ± 1/16	2-1/2 ± 1/16

AMSC N/A FSC 6520

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 202640-26T08:57Z
Check the source to verify that this is the current version before use.

Both the maxillary and mandibular bases shall have a minimum wall height of 7/16 inch anterior to posterior middle line-section. The wall thickness for all model bases shall be a minimum of 1/8 inch.

Markings. Former set shall be marked with catalog number.

Finish. The former set shall have a finish such that all models formed are smooth, without granular deformities.

Instructions for use shall be furnished with each set.

Workmanship. The former set shall be of uniform thickness. The former set shall be free of defects, such as holes, tears, or excess rubber around edges, which will detract from its appearance or impair its serviceability.

Unit. Set (SE). One former set, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspections. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Sampling for inspection. Sampling for inspection shall be conducted in accordance with MIL-STD-105. Where feasible, the same sample shall be used for the determination of two or more inspection characteristics. The unit of product for inspection purposes shall be one set. Unless otherwise specified, sampling for examination shall be conducted at inspection level II, AQL 2.5, and sampling for tests (when applicable) shall be conducted at inspection level S-1, AQL 1.0.

Source: https://assist.dla.mil -- Downloaded: 202640-26T08:57Z
Check the source to verify that this is the currenversion before use.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Numbers are covered by this document:

National Stock Number	Item Identification
6520-01-182-7858	FORMER SET, DENTAL MODEL Adult, Medium
6520-01-182-7859	FORMER SET, DENTAL, MODEL Adult, Large

Source: https://assist.dla.mil -- Downloaded: 202610-26T08:57Z
Check the source to verify that this is the current version before use.

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be as specified in the contract and/or order.

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Civil Agency Coordinating Activity:
VA-OSS

PREPARING ACTIVITY:

DoD-MR

Agent:

DLA-DM

Project No. 6520-1897

Location: NCR

Source: https://assist.dla.mil -- Downloaded: 202640-26T08:57Z
Check the source to verify that this is the currentversion before use.