A-A-51563
28 July 1986

COMMERCIAL ITEM DESCRIPTION

ARBOR AND BAND SET, Dental Handpiece

The General Services Administration has authorized the use of this commercial item description in preference to Interim Federal Specification GG-A-00671C.

This commercial item description covers an arbor and band set suitable for use with straight type dental handpieces or with dental lathes, in the grinding and finishing of dental prosthetic appliances.

Salient characteristics:

The arbor and band set shall be an assembly consisting of an arbor, sleeve, band with outer surface of silicon carbide, and screw. The component parts of the arbor and band set shall conform to figure 1. Washers shall be used only with the large sleeve in assembly. The arbors, screws and washers shall be made from corrosion-resisting steel, or carbon steel, nickel or chromium plated. The arbor, screws and washers shall be of proper temper and suitable for the purpose intended.

The sleeve shall be made from natural or synthetic rubber of suitable grade with a hardness of 50A to 60A durometer. The resiliency of the rubber sleeve shall remain suitable for the use intended when subjected to the accelerated aging test in accordance with method 7221 of FED-STD-601. As an alternate, the sleeves shall remain flexible and suitable for the use intended for not less than (5) years from date of delivery to the Government when stored at 50° to 90° Fahrenheit. The sleeve shall be capable of expanding when the arbor screw is tightened, in order to hold the abrasive band firmly in place.

The abrasive band shall consist of an inner and outer band fastened together, by a suitable adhesive, to form a closed circular band. The inner band shall be jeans cloth. The outer band shall be jeans cloth having its outer surface closed coated with 60 or 80 grit silicon carbide. The jeans cloth shall have a thread count of 96 warp by 64 filling per inch. The joints of both the inner and outer bands shall be of the 45 ± 5 degree angle, butt type, bonded with a good quality animal hide glue or equivalent adhesive. The joint of the inner band shall be positioned diagonally and diametrically opposite the joint of the outer band. The abrasive band shall be of such flexibility as to withstand the required operating procedures.

AMSC N/A FSC 6520

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:57Z
Check the source to verify that this is the current version before use.

The arbor and band set shall consist of the following assortment:

Six arbors.
Six screws.
Six washers.
Three sleeves, 1/4 inch O.D.
Three sleeves, 3/8 inch O.D.
One hundred abrasive bands, 60 grit, 1/4 inch I.D.
One hundred abrasive bands, 60 grit, 3/8 inch I.D.
One hundred abrasive bands, 80 grit, 1/4 inch I.D.
One hundred abrasive bands, 80 grit, 3/8 inch I.D.

Workmanship. The arbor and band set shall be free from defects which detract from its appearance or may impair its serviceability. The arbors, with sleeve and band attached, shall run true. The accessories shall interface with each other properly to provide a uniform ground surface on the devices being prepared. The sleeves shall be properly glued, to prevent raveling. The abrasive coating shall be of uniform thickness.

Unit. Set (SE). One arbor and band set, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:57Z
Check the source to verify that this is the current version before use.




Figure 1. Arbor and Band Set

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:57Z
Check the source to verify that this is the current version before use.

Sampling for inspection. Sampling for inspection shall be conducted in accordance with MIL-STD-105. Where feasible, the same sample shall be used for the determination of two or more inspection characteristics. The unit of product for inspection purposes shall be one arbor and band set. Unless otherwise specified, sampling for examination shall be conducted at inspection level II, AQL 2.5, and sampling for tests (when applicable) shall be conducted at inspection level S-1, AQL 1.0.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conform to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractors shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:57Z
Check the source to verify that this is the current version before use.

NOTE: The following National Stock Number is covered by this document:

National Stock Number	Item Identification
6520-00-500-9500	ARBOR AND BAND SET, Dental Handpiece, Silicon Carbide, 60 and 80 Grit.

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be as specified in the contract and/or order.

MILITARY INTERESTS:

Custodians:

Army-MD
Navy-MS
Air Force - 03

PREPARING ACTIVITY:

DoD-MB

AGENT:

DLA-DM

Civil Agency Coordinating Activity:
VA-OSS

Project No. 6520-1931

47807

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:57Z
Check the source to verify that this is the current version before use.