A-A-51565
18 August 1986

COMMERCIAL ITEM DESCRIPTION

HOLDER, RUBBER DAM, DENTAL (Nylon)

The General Services Administration has authorized the use of this commercial item description.

This commercial item description covers a nylon holder for retaining rubber dam.

Salient characteristics:

The holder shall be adult size, "U" shaped, lightweight and radiolucent for use in supporting and stabilizing a sheet of rubber dam over the mouth of a patient during endodontic therapy.

The holder shall be made of natural nylon 6-6 plastic. The holder shall conform to the design and dimensions shown in figure 1. The holder shall have a minimum of seven tines to engage the rubber dam. The tines shall be distributed so that one tine is in the center of the base. The horizontal portion of the "U" shape shall be contoured so that a rising arc is formed from the base plane of "U" shape to permit close approximation by the vertical projections to the cheek of the patient. The rubber dam holder shall have a smooth finish and shall have no flash and projections except for the tines. The holder shall be of a lightweight rigid plastic weighing 18 ± 2 grams. The holder shall be unbreakable in normal use. The holder shall not discolor with age and shall be chemically resistant to chloroform, germicides and irrigating solutions. The rubber dam shall withstand repeated autoclaving at temperatures up to 270°F without showing any deterioration such as warping or brittleness.

Markings. The holder shall be marked with manufacturer's name and applicable item name. Each unit of issue shall include a statement that holder is steam autoclavable up to 270°F.


Workmanship. The holder shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Each (EA). One rubber dam holder as specified, constitutes one unit.

AMSC N/A FSC 6520

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:57Z
Check the source to verify that this is the current version before use.



Notes: All dimensions are in inches.

Unless otherwise specified, tolerance shall be ± 1/16 inch.

Break all sharp edges.

FIGURE 1. Holder, Rubber Dam, Dental (Nylon)

Source: https://assist.dla.mil -- Downloaded: 202610-26T08:57Z
Check the source to verify that this is the current version before use.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Sampling for inspection. Sampling for inspection shall be conducted in accordance with MIL-STD-105. Where feasible, the same sample shall be used for the determination of two or more inspection characteristics. The unit of product for inspection purposes shall be one holder. Unless otherwise specified, sampling for examination shall be conducted at inspection level II, AQL 2.5, and sampling for tests (when applicable) shall be conducted at inspection level S-1, AQL 1.0.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specification, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Source: https://assist.dla.mil -- Downloaded: 202640-26T08:57Z
Check the source to verify that this is the currenversion before use.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

National Stock Number	Item Identification
6520-01-049-3597	HOLDER, RUBBER DAM, DENTAL, Radiolucent, "U" Shaped, Adult.

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be as specified in the contract and/or order.

MILITARY INTERESTS:

Custodians:

Army-MD
Navy-MS
Air Force - 03

CIVIL AGENCY COORDINATING ACTIVITY:
VA-OSS

49129

Preparing Activity:

DoD-MB

Agent:

DLA-DM

Project No. 6520-1938

Location: NCR Disk 2

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:57Z
Check the source to verify that this is the current version before use.