COMMERCIAL ITEM DESCRIPTION

IMPRESSION MATERIAL, DENTAL (Agar Type)

The General Services Administration has authorized the use of this commercial item description in preference to Interim Federal Specification U-I-00496b (DSA-DM).

This commercial item description covers two types of dental agar tray impression material in gel form, suitable for taking oral impressions.

Type I - Regular Bodied
Type II - Heavy Bodied

Salient characteristics:

The impression material shall be of the type specified in the procurement document for the applicable National Stock Number and Item Identification. The impression material shall be in accordance with ANSI/ADA Specification No. 11 and the following additional requirements:

1. Type I material shall be red in color. Type II material shall be green in color.

2. The compressive strength shall be not less than 5,600 g/cm^2 in lieu of 2,500 g/cm^2 for type II material.

3. The strain in compression shall be not less than 4.0 nor more than 7.0 percent in lieu of 4.0 and 15.0 percent, respectively, for type II material.

4. The viscosity shall be 1.70-1.80 millimeters per gram for type I material and 1.50-1.65 millimeters per gram, for type II material, when tested as follows: The material (tubes) to be tested shall be placed in a liquefying bath and brought to a rolling boil. Reset timer and boil 10 additional minutes. The material shall be immediately transferred to a storage tank, held exactly within a 155°F to 160°F range, and cooled for exactly 60 minutes. Quickly deposit 40 to 60 grams of unkneaded material from each tube on clean glass plates. Allow the material to seek its own final diameter. Measure the minimum and maximum diameters of each patty or disk in millimeters and weigh (grams) each disk. The viscosity in mm/g shall be the sum of the minimum and maximum diameters, divided by the weight.

AMSC N/A FSC 6520

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:57Z
Check the source to verify that this is the current version before use.

Workmanship. The impression material shall be clean and free from defects which detract from its appearance or impair its function.

Unit. Package (PG). One package containing 12 tubes of impression material of one type only, as specified, constitutes one unit. Each tube shall contain sufficient material to take a full size upper or lower impression (approximately 2 ounces net weight).

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Sampling for inspection procedures. Sampling and inspection procedures of ANSI/ADA Specification No. 11 shall apply for each lot of impression material furnished.

Contractor certification. The contractor shall certify that the product offered is ADA certified, meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:57Z
Check the source to verify that this is the current version before use.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Numbesr are covered by this document:

NSN	ITEM IDENTIFICATION	Type
6520-00-536-2450	IMPRESSION MATERIAL, DENTAL, Hydrocolloid Tray Material, Paste, Regular Bodied, Agar Type, 12s.	Type I
6520-01-120-5020	IMPRESSION MATERIAL, DENTAL, Hydrocolloid Tray Material, Paste, Heavy Bodied, Agar Type, 12s.	Type II

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be as specified in the contract and/or order.

MILITARY INTERESTS:

Custodians:

Army-MD
Navy-MS
Air Force - 03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

Civil Agency Coordinating Activity:
VA-OSS

Project No. 6520-1869

48000
49124

Location: NCR

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:57Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS: In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE: This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

DEFENSE LOGISTICS AGENCY

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO 4966 WASHINGTON D C

POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
Attn: DPSC-ATTS
2800 South 20th Street
Philadelphia, PA 19101

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:57Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1. DOCUMENT NUMBER	2. DOCUMENT TITLE
A-A-51566	Impression Material, Dental (Agar Type)

3a. NAME OF SUBMITTING ORGANIZATION	4. TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR
	☐ USER
b. ADDRESS *(Street, City, State, ZIP Code)*	☐ MANUFACTURER
	☐ OTHER *(Specify)* _____

5. PROBLEM AREAS

a. Paragraph Number and Wording:

b. Recommended Wording

c. Reason/Rationale for Recommendation

6. REMARKS

7a. NAME OF SUBMITTER *(Last, First, MI)* – Optional	b. WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
c. MAILING ADDRESS *(Street, City, State, ZIP Code)* – Optional	8. DATE OF SUBMISSION *(YYMMDD)*

DD FORM 82 MAR 1426 PREVIOUS EDITION IS OBSOLETE

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:57Z
Check the source to verify that this is the current version before use.