A-A-51577
25 July 1986

COMMERCIAL ITEM DESCRIPTION

VIAL, BLOOD DILUTING (Disposable, Clear)

This commercial item description covers a disposable plastic blood dilution vial with a flat projecting optical base, for use with the Coulter Electronics Model ZBI automatic blood cell counter.

Salient characteristics:

Shall be a disposable clear polystyrene vial having a 20 milliliter minimum capacity. Shall be supplied with a fitted polyethylene snap-on cap.

Shall be for use as a blood diluting vial in the Coulter Electronics Model ZBI automatic blood cell counter and other blood cell counters that are stock listed under NSN 6630-01-016-9432 (Counter, Blood Cells, Automatic, Digital, 115 Volt, 50-60 Hz, AC).

Shall have a configuration that is in the shape of a round cylinder with rounded inner seams and a projecting square bottom base. The inner surfaces shall be leakproof. The base shall have optically clear surfaces on four sides. Frosted surface(s) on the cylindrical portion shall be included for convenient marking and labeling.

Shall have nominal dimensions that include a top diameter of 34 millimeters (mm) and overall height of 60 mm. Vials having minimum dimensions of 32.0 mm diameter and 56.0 mm overall height shall be acceptable.

Shall be supplied free of dust and any other particle matter, and shall be packaged in a shrink wrapped tray(s) and in such a manner as to remain clean until used.

Workmanship. The vial and closure cap shall be free from defects which detract from their appearance or impair their serviceability.

Unit. Package (PG). One package containing five hundred capped vials, as specified, constitutes one unit.

AMSC N/A FSC 6640

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 202610-26T08:58Z
Check the source to verify that this is the current version before use.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Inspection and Tests. Inspection and tests shall be conducted in accordance with commercial practice to determine compliance with specification requirements. Sampling for inspections and tests shall also be conducted in acordance with commercial practice.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:58Z
Check the source to verify that this is the current version before use.

Federal Food, Drug and Cosmetic Act cont'd. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment form the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

The following National Stock Number is covered by this number:

6640-01-111-1158

Ordering data (Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order).

MILITARY INTERESTS:

Custodians:

Army-MD
Navy-MS
Air Force-03

PREPARING ACTIVITY:

DoD-MB

Agent:
DLA-DM

CIVIL AGENCY COORDINATING ACTIVITY:

VA-OSS

Project No. 6640-1382

Location : CID-vial-blood-1158, NCR, disk 6

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:58Z
Check the source to verify that this is the current version before use.