MIL-A-16288/3
28 November 1973
SUPERSEDING
MS39003
27 November 1968

MILITARY SPECIFICATION SHEET

ADAPTER, COMPRESSED GAS CYLINDER VALVE CONNECTIONS:

U. S. REGULATOR TO JAPANESE OXYGEN VALVE

This specification sheet is approved for use by all Departments and Agencies of the Department of Defense.

The complete requirements for procuring the adapter described herein shall consist of this document and the issue in effect of MIL-A-16288.

REQUIREMENTS:

1. The adapter shall be as specified in figure 1. Unless otherwise specified herein, dimensions are in inches, linear tolerances shall be plus or minus 1/64 inch, and angular tolerance shall be plus or minus 1/2 degree.

2. Permanently mark on surface in 1/8-inch-minimum characters: "U.S.-JAPAN O_2".

3. Maximum service pressure 3000 psi.

INTENDED USE: The adapter is intended to connect a U. S. oxygen regulator to a Japanese oxygen valve.

Custodians:

Army - ME
Navy - YD
Air Force - 68

Preparing activity:

Army - ME

Project No. 8120-0341

Review activities:

Army - MD
Navy - SH, MS

User activities:

Army - AV
Navy - MC, AS

FSC 8120

Source: https://assist.dla.mil -- Downloaded: 202610-18T08:28Z
Check the source to verify that this is the current version before use.




FIGURE 1. ADAPTER, U.S. REGULATOR TO JAPANESE VALVE, OXYGEN.

☆ U. S. GOVERNMENT PRINTING OFFICE: 1973 -713-149/2505

X-2379

Source: https://assist.dla.mil -- Downloaded: 202610-18T08:28Z
Check the source to verify that this is the current version before use.