MIL-B-16924C
Amendment-4
9 October 1974
SUPERSEDING
Amendment-3
30 March 1974

MILITARY SPECIFICATION

BLOOD RECIPIENT SET, INDIRECT TRANSFUSION

This Amendment forms a part of Military Specification MIL-B-16924C, dated 4 April 1972 and is approved for use by all Departments and Agencies of the Department of Defense.

Page 4:

3.1.2.2 Hubs. Delete entirely and substitute:

"3.1.2.2 Hubs shall be fabricated from a suitable plastic (containing no exudable toxic materials), aluminum alloy, nickel plated brass, or stainless steel class 3, type 302, 303, or 304 of specification GG-I-526. Aluminum hubs shall be degreased and then either anodized, or passivated with nitric acid solution, resulting in a matte, nontinted finish.

Page 5:

3.2.1.2 Line 3 delete "98 plus 2, minus 4 Shore A at 25° C." and substitute "94 ± 3 Shore A at 25° C."

Page 15:

* 4.4.5 Line 4, delete "100° F" and substitute "70° F".

Page 16:

4.4.8 Reseal test (molded rubber part). Delete last sentence entirely.

Custodians:
Army - MD
Navy - MS
Air Force - 03

Review activities:
Army - MD
Navy - MS
Air Force - 03

Preparing activity:
Defense Supply Agency - DM

Project No. 6515-1872

Review information is current as of the date of this document. For future coordination of changes to this document, draft circulation should be based on the information in the current DODISS.

FSC 6515

Source: https://assist.dla.mil -- Downloaded: 2026-10-18T09:37Z
Check the source to verify that this is the current version before use.