


Example Comparison Report

Disagreement case: Swin miss, YOLO partial support

Case: a8750c349b5dac834473304bad0f2877

Review status: uncertain Confidence: moderate

Image Comparison



Model Comparison

Reference	Swin	YOLO
Aortic enlargement, Cardiomegaly, Pleural thickening	Predicted: No finding Correct: none Missed: Aortic enlargement, Cardiomegaly, Pleural thickening	Predicted: Cardiomegaly Correct: Cardiomegaly Missed: Aortic enlargement, Pleural thickening

Findings

- The Swin classifier indicates no acute findings with high confidence (98.9%).
- However, the YOLO detector identifies a potential cardiomegaly in the lower mediastinal region with low confidence (26.5%).

Impression

- Discordant AI model outputs regarding cardiac size assessment.
- While the image-level classifier strongly favors no acute abnormality, the object detection model suggests possible cardiomegaly with localization but low confidence.
- The chest X-ray appears technically adequate for interpretation.
- No other acute thoracic abnormalities are identified by either model.

Agentic Workflow Summary

The agentic report favors No finding with a moderate confidence band and a uncertain review status. Primary AI-supported findings are none, while the remaining uncertain findings are Cardiomegaly. Relative to radiologist labels, Swin captured none and YOLO localized Cardiomegaly. Overall pattern emphasis is No acute abnormality.

Supporting Evidence

Top Swin Probabilities

- - No finding: 0.989
- - Cardiomegaly: 0.010
- - Pleural thickening: 0.003
- - Other lesion: 0.002
- - Pulmonary fibrosis: 0.001
- - Aortic enlargement: 0.001

Top YOLO Detections

- - Cardiomegaly: 0.26 @ [189.0, 244.9, 398.9, 352.8]

Takeaways

- Swin correct labels: 0 / 3
- YOLO correct labels: 1 / 3
- Swin key miss: Aortic enlargement, Cardiomegaly, Pleural thickening
- YOLO key miss: Aortic enlargement, Pleural thickening

Confidence Rules Used

- Swin positive label cutoff: 0.50
- YOLO detection confidence cutoff: 0.25
- Agentic AI confidence band is qualitative and selected by the reviewer.

This is an AI-generated decision-support summary that requires human radiologist review for clinical use. The conflicting model outputs regarding cardiac size assessment necessitate expert interpretation, particularly given the clinical importance of cardiomegaly detection and the technical factors that can affect cardiac silhouette assessment on chest radiography.