

Presentation Comparison Report


Single localized abnormality Case: 804bcde30d36e32d9429f00bed7a388d

Review status: supported Confidence: high


Original CXR



Ground-truth boxes



YOLO support overlay



Ground Truth

- Aortic enlargement

Swin Classification

Predicted: Aortic enlargement
Correct: Aortic enlargement
Missed: None
Overcalled: None

YOLO Labels

Predicted: Aortic enlargement
Correct: Aortic enlargement
Missed: None
Overcalled: None

Agentic Workflow Summary

Supported findings: Aortic enlargement
Uncertain findings: None
Final report labels: Aortic enlargement
Pattern buckets: Cardiomeastinal abnormality

Evidence and Report Summary

Findings

The chest X-ray demonstrates aortic enlargement. Both the Swin classifier (probability 0.89) and YOLO detector (confidence 0.26) identified this finding, with localization support provided by the detection model in the expected mediastinal region. While the Swin classifier also suggested moderate probabilities for pleural thickening (0.30) and lower probabilities for cardiomegaly (0.17) and calcification (0.09), these findings lack localization support from the detection model and are not definitively supported by the available evidence.

Impression

Findings consistent with aortic enlargement. This represents a cardiomediastinal abnormality that warrants clinical correlation and potential further evaluation as clinically indicated.

Top Swin Probabilities

- Aortic enlargement: 0.890
- Pleural thickening: 0.303
- No finding: 0.237
- Cardiomegaly: 0.167
- Calcification: 0.091
- Other lesion: 0.081

Ground-Truth Boxes Top YOLO Detections

- | | |
|---|---|
| - Aortic enlargement @ [272.8, 110.0, 328.7, 211.0] | - Aortic enlargement: 0.26 @ [272.6, 116.3, 329.1, 190.2] |
|---|---|

Presentation Takeaway

Swin correct labels: 1 / 1
YOLO correct labels: 1 / 1
Key miss for Swin: None
Key miss for YOLO: None

Thresholds Used

Swin positive label cutoff: 0.50
YOLO detection confidence cutoff: 0.25
Agentic AI confidence band is qualitative, not a fixed numeric threshold.

This is an AI-generated decision-support summary based on automated image analysis. Human radiologist review is required for clinical

Confidence Calculation

How this case's confidence was calculated

Per-finding band cutoffs: high ≥ 0.80 , moderate ≥ 0.50 , low < 0.50 .

Swin label threshold = 0.50. YOLO detection threshold = 0.25.

Aortic enlargement: high from score = $\max(\text{Swin}(0.89), \min(1.00, 0.6 \cdot 0.89 + 0.4 \cdot 0.26)) = 0.89$

Agentic AI report band reference: `high` occurred in 236 / 300 cases with exact-match accuracy 90.7%.

Interpretation note: the report-level agentic AI confidence band is constrained to high / moderate / low in the schema, but the model chooses that band. The per-finding confidence lines above are deterministic and computed from the Swin probability, YOLO support, and the label taxonomy.

This is an AI-generated decision-support summary based on automated image analysis. Human radiologist review is required for clinical