Examiners：_____

Examinee：_____

Date： ____/____/____ *(MM/DD/YEAR)*

R605 EEG Experiment Procedure Checklist

- **EEG Recording Software:** Curry ☐ 8 ☐ 9
- **NeuroScan Quick Cap EEG Cap:** 32-channel ☐ S ☐ M ☐ L ☐Others：
- **Does the EEG pre-recording preparation include examiner assistance?**

  ☐No ☐Yes, assisted parts: _____
- **Scheduled Experiment Time:** _____
- **Experiment Timeline:**

Time	Event	Duration
	Experiment starts, participants arrives at the lab.	Hair washing: ____ minutes (Estimate maximum 10 minutes)
	Participant finish pre-experiment hair washing.	
	The start of EEG pre-recording preparation	EEG pre-recording preparation : ____ minutes (Approximately around 10-15 minutes for 14 channels)
	Cleaning & Exfoliating EOG & mastoids positions, and completing the positioning of CZ channel of EEG cap	
	Completing gel insertion for all channels (The impedances are between 5~15 kΩ; < 5 kΩ is optimal), and signal quality is stable.	
	Experiment starts	Overall experiment duration : ____ minutes
	Experiment ends	
	Participant finish post-experiment hair washing.	Hair washing: ____ minutes (Estimate maximum 10 minutes)
	The completion of the experiment, participants leave the lab.	

  - **Overall Duration:** ___hours ____minutes。

- **Procedure Checklist:**

Item	Details	Assessment
Gel Making	Gel viscosity	☐Pass ☐ Room for improvement ☐ failed, reasons:
	Preparation amount (The number of syringes)	☐Pass ☐ Room for improvement ☐ failed, reasons:

EEG Recording Software	Connect the cap to the amplifier before opening the software	☐Pass ☐ Room for improvement ☐ failed, reasons:
	Channels Configuration	☐Pass ☐ Room for improvement ☐ failed, reasons:
	Impedance and signal quality check	☐Pass ☐ Room for improvement ☐ failed, reasons:
Elastic Cap	Size selection (S / M / L)	☐Pass ☐ Room for improvement ☐ failed, reasons:
	Cz center positioning	☐Pass ☐ Room for improvement ☐ failed, reasons:
	Chin strip position & timing	☐Pass ☐ Room for improvement ☐ failed, reasons:
Gel Insertion	Channel Order ground & reference	☐Pass ☐ Room for improvement ☐ failed, reasons:
	EOG & mastoids Channels positioning	☐Pass ☐ Room for improvement ☐ failed, reasons:
	Gel Insertion (Move the hair prior to inserting gel)	☐Pass ☐ Room for improvement ☐ failed, reasons:
	Post gel insertion: Electrode adjustment	☐Pass ☐ Room for improvement ☐ failed, reasons:
EEG	Remind to sit still and avoid head movements during the experiment	☐Pass ☐ Room for improvement ☐ failed, reasons:
	Constantly check the subject's condition and EEG signal quality	☐Pass ☐ Room for improvement ☐ failed, reasons:
	Cap takeoff after the software is shut down	☐Pass ☐ Room for improvement ☐ failed, reasons:
EEG Cleaning	Restore the original R605 lab setup	☐Pass ☐ Room for improvement ☐ failed, reasons:
	The cleanness of cap, gel preparation tools, syringe, and needle	☐Pass ☐ Room for improvement ☐ failed, reasons:
Others		☐Pass ☐ Room for improvement ☐ failed, reasons:

		☐Pass ☐ Room for improvement ☐ failed, reasons:

■ **Assessment final result:**