

HPLC Analytical Test Report

Report Type: High Performance Liquid Chromatography Analysis

Instrument ID: 2

Instrument Type: HPLC with UV Detection

Analysis Date: 2026-01-07

Operator: jane.smith

Method Name: HPLC_USP_Assay_Method_v2.0

CALIBRATION RESULTS

Correlation Coefficient R: 0.99995

R-Squared (R^2): 0.9999

Slope: 14.58

Intercept: -2.35

Standard Error: 45.23

Relative Standard Deviation: 0.45%

QUALITY CONTROL RESULTS

Total QC Results: 9

Passed: 9

Failed: 0

Pass Rate: 100.0%

Acceptance Criteria: 80-120% Recovery

Overall Status: ACCEPTED

SAMPLE ANALYSIS RESULTS

Sample BIO-002: 424.8 ug/mL - Status: VALID

Sample BIO-003: 438.5 ug/mL - Status: VALID

Sample API-001: 401.8 ug/mL - Status: VALID

Sample PHM-001: 388.4 ug/mL - Status: VALID

CONCLUSION

All acceptance criteria have been met. Analysis results are acceptable.

Analyst: jane.smith | Date: 2026-01-07 | Time: 14:15:20