Product Catalog

5 Products

Generated: 10/23/2026

DLinRT.eu Database

Table of Contents

1. Limbus Contour (Limbus AI)
2. Contour+ (MVision AI)
3. DLC Expert (Mirada Medical)
4. Deep Learning Segmentation (RaySearch Laboratories)
5. ART-Plan Segmentation (Therapanacea)

For information only. Not a clinical validation or endorsement; verify with the manufacturer and local commissioning.

1. Limbus Contour

Limbus AI

General Information

Product Name:	Limbus Contour
Company:	Limbus AI
Category:	Auto-Contouring
Version:	3.0
Release Date:	6/15/2023
Last Updated:	6/15/2026
Description:	Limbus Contour is a software-only medical device intended for use by trained radiation oncologists, dosimetrists and physicists to derive optimal contours for input to radiation treatment planning. Supported image modalities are Computed Tomography and Magnetic Resonance. The Limbus Contour Software assists in the following scenarios: Operates in conjunction with radiation treatment planning systems or DICOM viewing systems to load, save, and display medical images and contours for treatment evaluation and treatment planning; Creation, transformation, and modification of contours for applications including, but not limited to: transferring contours to radiotherapy treatment planning systems, aiding adaptive therapy and archiving contours for patient follow-up; Localization and definition of healthy anatomical structures. Limbus Contour is not intended for use with digital mammography. (Source: FDA 510(k) K241837 Summary, accessed 2026-05-30)

Evidence Classification

Rigor E2 — External, independent or multi-center validation **Efficacy L3 — Decision/workflow efficacy**

Readiness R4 — High implementation readiness

+ Vendor-independent **+ Multi-centre** **+ Multi-national** **- Prospective** **+ External validation**

24 scored publication(s) · 24 listed · score origin: papers
Rigor: Multiple independent multi-institutional peer-reviewed studies (Starke 2024, Radici 2022, Grossi 2025, Doolan 2023, Tabita 2025) support E2. No prospective/RCT evidence — insufficient for E3.
Efficacy: Independent workflow-adoption and time-saving evidence (Radici 2022) plus consistency benefits supports I3.
Readiness: Derived from E3 + CE + FDA 510(k): mature peer-reviewed multi-center evidence with regulatory clearance; minor local commissioning and user training expected.

For information only. Not a clinical validation or endorsement; verify with the manufacturer and local commissioning.

Evidence by Source (24)

An efficient automated approach for accumulated dose estimation in prostate cancer radiotherapy
Grohmann M, Krug D, Baehr A, et al. · Phys Imaging Radiat Oncol · 2026 · DOI 10.1016/j.phro.2026.100942
E1 L2

Trade off of different deep learning based auto segmentation approaches for t
Thibodeau Antonacci et al. · Medical Physics · 2025 · DOI 10.1002/mp.17782
E1 L2

Validation of different automated segmentation models for target volume contouring in…
Meixner E et al. · Clin Transl Radiat Oncol · 2024 · DOI 10.1016/j.ctro.2024.100855
E1 L2

Automated segmentation in pelvic radiotherapy: A comprehensive evaluation of ATLAS-, machine…
Bordigoni B et al. · Phys Med · 2024 · DOI 10.1016/j.ejmp.2024.104486
E1 L3 Sets product level

A student trained convolutional neural network competing with a commercial AI software and experts…
Bürkle et al. · Scientific Reports · 2024 · DOI 10.1038/s41598-024-76288-y
E1 L2

Artificial Intelligence–Based Autosegmentation: Advantages in Delineation, Absorbed Dose-Distribution,…
Sarria G et al. · Advances in Radiation Oncology · 2024 · DOI 10.1016/j.adro.2023.101394
E1 L3

Clinical Implementation and Evaluation of Auto-Segmentation Tools for Multi-Site Contouring…
Heilemann G et al. · Phys Imaging Radiat Oncol · 2023 · DOI 10.1016/j.phro.2023.100515
E1 L2

Clinical Use of a Commercial Artificial Intelligence-Based Software for Autocontouring in Radiation…
Hoque S et al. · Cancers · 2023 · DOI 10.3390/cancers15245735
E1 L3

Clinical Validation of a Deep-Learning Segmentation Software in Head and Neck: An Early Analysis in…
D'Aviero A et al. · International Journal of Environmental Research and Public Health · 2022 · DOI 10.3390/ijerph19159057
E1 L2

Clinical validation of AI-assisted contouring in prostate radiation therapy treatment planning:…
Arjmandi N et al. · J Appl Clin Med Phys · 2026 · DOI 10.1002/acm2.70425
E1 L3

Comparative analysis of AI-generated and deformed image registration contours on daily CBCT in…
Ashburner M et al. · Phys Eng Sci Med · 2026 · DOI 10.1007/s13246-025-01686-z
E1 L2

For information only. Not a clinical validation or endorsement; verify with the manufacturer and local commissioning.

Implementation of deep learning-based auto-segmentation for radiotherapy planning structures: a…
Wong J et al. · Radiat Oncol · 2021 · DOI 10.1186/s13014-021-01831-4
E1 L3

Integrating Artificial Intelligence Into Radiation Oncology: Can Humans Spot AI?
Shanbhag NM et al. · Cureus · 2023 · DOI 10.7759/cureus.50486
E1 L1c

Training and Validation of Deep Learning-Based Auto-Segmentation Models for Lung Stereotactic…
Wong J et al. · Frontiers in Oncology · 2021 · DOI 10.3389/fonc.2021.626499
E1 L2

Workflow Efficiency in Vaginal Cuff High Dose Rate Brachytherapy Using Artificial Intelligence-Based…
Walter YA et al. · Cancers (Basel) · 2025 · DOI 10.3390/cancers17172751
E1 L3

Clinical evaluation of the efficacy of limbus artificial intelligence software to augment contouring for…
Starke A et al. · British Journal of Radiology · 2024 · DOI 10.1093/bjr/tqae077
E2 L3 Sets product level

Workflow impact of Limbus AI auto-contouring
Radici L et al. · Life · 2022 · DOI 10.3390/life12122088
E1 L3

Can Deep Learning-Based Auto-Contouring Software Achieve Accurate Pelvic Volume Delineation…
Grossi C et al. · Current Oncology · 2025 · DOI 10.3390/curroncol32060321
E2 L2

A clinical evaluation of the performance of five commercial artificial intelligence contouring systems…
Doolan PJ et al. · Front. Oncol. · 2023 · DOI 10.3389/fonc.2023.1213068
E2 L3

Evaluation and failure analysis of four commercial deep learning based autose
Fan M et al. · J Applied Clin Med Phys · 2025 · DOI 10.1002/acm2.70010
E2 L2

Robustness evaluation of an artificial intelligence-based automatic contouring software in daily…
Fontaine J et al. · Physica Medica · 2025 · DOI 10.1016/j.ejmp.2025.105065
E1 L1t

An evaluation of rectum contours generated by artificial intelligence automatic contouring software…
McLaughlin O et al. · Biomed. Phys. Eng. Express · 2025 · DOI 10.1088/2057-1976/adf8f2
E1 L2

For information only. Not a clinical validation or endorsement; verify with the manufacturer and local commissioning.

Investigation on performance of multiple AI-based auto-contouring systems in organs at risks…
Kim YW et al. · Phys Eng Sci Med · 2024 · DOI 10.1007/s13246-024-01434-9
E2 L2

Comparison of Two Auto-Contouring Systems for Head and Neck Organs at Risk to Institutional…
Bayley C et al. · Applied Sciences · 2026 · DOI 10.3390/app16115681
E1 L2

Key Features

- Automated contouring
- Runs locally
- DICOM compatibility

Technical Specifications

Input Formats:	DICOM
Output Formats:	DICOM-RTSTRUCT
Processing Time:	Minutes per case
Integration:	Standalone
Deployment:	Local
Modality Support:	CT; MRI; CBCT

Regulatory Information

CE: cleared FDA: 510k_cleared

CE Status:	cleared
FDA Status:	510k_cleared
CE Details:	cleared (Class Class IIa)
FDA Details:	510k_cleared (K241837)

Contact Information

Website:	https://limbus.ai/
Product URL:	https://limbus.ai/
Company URL:	https://limbus.ai/

For information only. Not a clinical validation or endorsement; verify with the manufacturer and local commissioning.

2. Contour+

MVision AI

General Information

Product Name:	Contour+
Company:	MVision AI
Category:	Auto-Contouring
Version:	1.3.1
Release Date:	10/27/2025
Last Updated:	9/8/2026
Description:	Contour+ (MVision AI Segmentation) is a software system for image analysis algorithms to be used in radiation therapy treatment planning workflows. The system includes processing tools for automatic contouring of CT and MR images using machine learning based algorithms. The produced segmentation templates for regions of interest must be transferred to appropriate image visualization systems as an initial template for a medical professional to visualize, review, modify and approve prior to further use in clinical workflows. The system creates initial contours of pre-defined structures of common anatomical sites, i.e., Head and Neck, Brain, Breast, Lung and Abdomen, Male Pelvis, and Female Pelvis. Contour+ (MVision AI Segmentation) is not intended to detect lesions or tumors. The device is not intended for use with real-time adaptive planning workflows. (Source: FDA 510(k) K241490 Summary, accessed 2026-05-30)

Evidence Classification

Rigor E2 — External, independent or multi-center validation **Efficacy L3 — Decision/workflow efficacy**

Readiness R4 — High implementation readiness

+ Vendor-independent **+ Multi-centre** **+ Multi-national** **- Prospective** **+ External validation**

16 scored publication(s) · 16 listed · score origin: papers
Rigor: Independent peer-reviewed studies (Doolan 2023, Miura 2025 x2) demonstrate multi-site validation. Multi-national HARMONY study pending peer review. No prospective/RCT so E3 not met.
Efficacy: Independent literature documents geometric accuracy plus workflow and consistency benefits, including reduced editing effort; consistent with I2. No prospective outcome study identified.
Readiness: Derived from E2 + CE + FDA 510(k): mature peer-reviewed multi-center evidence with regulatory clearance; minor local commissioning and user training expected.

For information only. Not a clinical validation or endorsement; verify with the manufacturer and local commissioning.

Evidence by Source (16)

How smart is artificial intelligence in organs delineation? Testing a CE and FDA-approved…
Strolin S et al. · Front Oncol · 2023 · DOI 10.3389/fonc.2023.1089807
E1 L3 Sets product level

Safety and efficiency of a fully automatic workflow for auto-segmentation in radiotherapy using…
Cavus H et al. · Physics and Imaging in Radiation Oncology · 2024 · DOI 10.1016/j.phro.2024.100627
E1 L3

Validation of different automated segmentation models for target volume contouring in…
Meixner E et al. · Clin Transl Radiat Oncol · 2024 · DOI 10.1016/j.ctro.2024.100855
E1 L2

Automated segmentation of target volumes in breast cancer radiotherapy, impact on target size and dose…
Tang V et al. · Clin Transl Radiat Oncol · 2025 · DOI 10.1016/j.ctro.2025.100986
E1 L2

Autosegmentation based on different-sized training datasets of consistently-curated volumes and impact…
Elisabeth Olsson C et al. · Phys Imaging Radiat Oncol · 2022 · DOI 10.1016/j.phro.2022.04.007
E1 L2

Deep-learning magnetic resonance imaging-based automatic segmentation for organs-at-risk in the…
Turcas A et al. · Phys Imaging Radiat Oncol · 2023 · DOI 10.1016/j.phro.2023.100454
E1 L3

Evaluation of the accuracy of automated segmentation based on deep learning for prostate cancer patients
Miura H et al. · Med Dosim · 2025 · DOI 10.1016/j.meddos.2024.09.002
E1 L2

Potential of E-Learning Interventions and Artificial Intelligence-Assisted Contouring Skills in…
Rasmussen ME et al. · JCO Glob Oncol · 2024 · DOI 10.1200/GO.24.00173
E1 L3

Automated segmentation in pelvic radiotherapy: A comprehensive evaluation of ATLAS-, machine…
Bordigoni B et al. · Phys Med · 2024 · DOI 10.1016/j.ejmp.2024.104486
E1 L3

A clinical evaluation of the performance of five commercial artificial intelligence contouring systems…
Doolan PJ et al. · Front. Oncol. · 2023 · DOI 10.3389/fonc.2023.1213068
E2 L3 Sets product level

Performance evaluation of MVision AI Contour+ in gastric MALT lymphoma segmentation
Miura H et al. · Rep Pract Oncol Radiother. · 2025 · DOI 10.5603/rpor.104144
E1 L3

For information only. Not a clinical validation or endorsement; verify with the manufacturer and local commissioning.

Multicentre evaluation of deep learning CT autosegmentation of the head and neck region for radiotherapy
Pang EPP et al. · npj Digit. Med. · 2025 · DOI 10.1038/s41746-025-01624-z
E2 L3

Geometric, dosimetric and psychometric evaluation of three commercial AI software solutions for…
Podobnik G et al. · Sci Rep · 2025 · DOI 10.1038/s41598-025-18598-3
E2 L2

Artificial intelligence contouring in radiotherapy for organs-at-risk and lymph node areas
Meyer C et al. · Radiat Oncol · 2024 · DOI 10.1186/s13014-024-02554-y
E1 L2

A Deep Learning-Based Automated CT Segmentation of Prostate Cancer Anatomy for Radiation…
Kiljunen T et al. · Diagnostics · 2020 · DOI 10.3390/diagnostics10110959
E2 L3

Evaluation of AI-based segmentation and synthetic CT generation for MRI-only prostate…
Östensson A, Jonsson J · J Appl Clin Med Phys · 2026 · DOI 10.1002/acm2.70765
E1 L2

Key Features

- AI-powered segmentation
- Fast processing
- Clinical workflow integration
- Cloud or on-premise deployment

Technical Specifications

Input Formats:	DICOM
Output Formats:	DICOM-RTSTRUCT
Processing Time:	Minutes per case
Integration:	TPS integration; PACS integration
Deployment:	Cloud-based; On-premise
Modality Support:	CT; MRI

Regulatory Information

CE: cleared FDA: 510k_cleared

CE Status:	cleared
FDA Status:	510k_cleared

For information only. Not a clinical validation or endorsement; verify with the manufacturer and local commissioning.

CE Details:	cleared (Class Class IIa)
FDA Details:	510k_cleared (K241490)

Contact Information

Website:	https://mvision.ai/contour/
Product URL:	https://mvision.ai/contour/
Company URL:	https://mvision.ai/

For information only. Not a clinical validation or endorsement; verify with the manufacturer and local commissioning.

3. DLC Expert

Mirada Medical

General Information

Product Name:	DLC Expert
Company:	Mirada Medical
Category:	Auto-Contouring
Version:	2.2
Release Date:	5/20/2023
Last Updated:	6/15/2026
Description:	Workflow Box is a software system designed to allow users to route DICOM-compliant data to and from automated processing components. Supported modalities include CT, MR, RTSTRUCT. Workflow Box includes processing components for automatically contouring imaging data using deformable image registration to support atlas based contouring, re-contouring of the same patient and machine learning based contouring. Workflow Box is a data routing and image processing tool which automatically applies contours to data which is sent to one or more of the included image processing workflows. Contours generated by Workflow Box may be used as an input to clinical workflows including, but not limited to, radiation therapy treatment planning. Workflow Box must be used in conjunction with appropriate software to review and edit results generated automatically by Workflow Box components, for example image visualization software must be used to facilitate the review and edit of contours generated by Workflow Box component applications. Workflow Box is intended to be used by trained medical professionals. (Source: FDA 510(k) K181572 Summary [Workflow Box, which includes DLCExpert], accessed 2026-05-30)

Evidence Classification

Rigor E2 — External, independent or multi-center validation

Efficacy L3 — Decision/workflow efficacy

Readiness R3 — Enhanced implementation readiness

+ Vendor-independent

+ Multi-centre

- Multi-national

- Prospective

+ External validation

7 scored publication(s) · 7 listed · score origin: papers
Rigor: 2026-08-27 Wave 6 citation audit: the Brouwer 2020, Choi 2020 and Zhu 2021 entries were removed as scored papers — their stored DOIs resolve to unrelated articles on Crossref and no matching publication could be verified. Only Doolan 2023 (doi:10.3389/fonc.2023.1213068), an independent multi-system benchmark with geometric endpoints, remains, so the effective score is E2/I1.
Efficacy: Raised to I2 in the 2026-09-23 HAIR cross-check: a multi-centre UK study reports contouring time savings

For information only. Not a clinical validation or endorsement; verify with the manufacturer and local commissioning.

(doi:10.1016/j.phro.2022.11.003) and a replanning study reports dosimetric endpoints (doi:10.1016/j.meddos.2025.04.004). Product now discontinued.
Readiness: Derived from E2 + CE + FDA 510(k): moderate implementation effort — local validation, interface testing and workflow confirmation required before adoption.

Evidence by Source (7)

Evaluation of measures for assessing time-saving of automatic organ-at-risk segmentation in radiotherapy
Vaassen F, Hazelaar C, Canters R, et al. · Phys Imaging Radiat Oncol · 2020 · DOI 10.1016/j.phro.2019.12.001
E1 L3 Sets product level

An evaluation of MR based deep learning auto-contouring for planning head and neck radiotherapy
Hague C et al. · Radiother Oncol · 2021 · DOI 10.1016/j.radonc.2021.02.018
E1 L2

Assessment of contour accuracy in head and neck replanning: Deep learning trained model compared…
Yuen J et al. · Med Dosim · 2025 · DOI 10.1016/j.meddos.2025.04.004
E1 L2

Clinical evaluation of deep learning and atlas-based auto-segmentation for critical organs at risk in…
Gibbons E et al. · J Med Radiat Sci · 2023 · DOI 10.1002/jmrs.618
E1 L3

Evaluating the Effectiveness of Deep Learning Contouring across Multiple Radiotherapy Centres
Walker Z et al. · Phys Imaging Radiat Oncol · 2022 · DOI 10.1016/j.phro.2022.11.003
E2 L3 Sets product level

External validation of deep learning-based contouring of head and neck organs at risk
Brunenberg E et al. · Physics and Imaging in Radiation Oncology · 2020 · DOI 10.1016/j.phro.2020.06.006
E1 L2

A clinical evaluation of the performance of five commercial artificial intelligence contouring systems…
Doolan PJ et al. · Front. Oncol. · 2023 · DOI 10.3389/fonc.2023.1213068
E2 L3

Key Features

- Deep learning algorithms
- Batch processing
- Clinical workflow integration

Technical Specifications

Input Formats:	DICOM
Output Formats:	DICOM-RTSTRUCT

For information only. Not a clinical validation or endorsement; verify with the manufacturer and local commissioning.

Processing Time:	Minutes per patient
Integration:	TPS integration; PACS integration; Workflow manager
Deployment:	On-premises; Cloud-based
Modality Support:	CT

Regulatory Information

CE: withdrawn FDA: withdrawn

CE Status:	withdrawn
FDA Status:	withdrawn
CE Details:	withdrawn (Class Class IIa)
FDA Details:	withdrawn (K181572)

Contact Information

Website:	https://mirada-medical.com/product/mirada-dlcexpert-ai-autocontouring/
Product URL:	https://www.mirada-medical.com/radiation-oncology
Company URL:	https://mirada-medical.com/radiation-oncology/

For information only. Not a clinical validation or endorsement; verify with the manufacturer and local commissioning.

4. Deep Learning Segmentation

RaySearch Laboratories

General Information

Product Name:	Deep Learning Segmentation
Company:	RaySearch Laboratories
Category:	Auto-Contouring
Version:	2026
Release Date:	12/1/2023
Last Updated:	10/15/2026
Description:	RayStation is a software system for radiation therapy and medical oncology. Based on user input, RayStation proposes treatment plans. After a proposed treatment plan is reviewed and approved by authorized intended users, RayStation may also be used to administer treatments. The system functionality can be configured based on user needs. (Source: FDA 510(k) K240398 Summary [RayStation, which includes Deep Learning Segmentation], accessed 2026-05-30)

Evidence Classification

Rigor E2 — External, independent or multi-center validation

Efficacy L4 — Therapeutic/patient management efficacy

Readiness R3 — Enhanced implementation readiness

+ Vendor-independent + Multi-centre - Multi-national - Prospective + External validation

18 scored publication(s) · 18 listed · score origin: papers
Rigor: 2026-08-27 Wave 6 citation audit: the Sharma 2025 DOI was corrected to doi:10.1016/j.clon.2025.103796 (Geometric and dosimetric evaluation of a RayStation deep learning model); the Ng 2022 entry was removed because DOI 10.3390/app12147210 resolves to an unrelated computer-science article. Remaining independent single-centre studies support E1.
Efficacy: Raised to I2 in the 2026-08-27 Wave 6 backfill: Mikalsen et al. Acta Oncol 2023 reports editing effort and clinical acceptance for the thorax and breast models. Rigor E2 follows the independent five-system benchmark (Doolan 2023).
Readiness: Derived from E2 + CE + FDA 510(k): moderate implementation effort — local validation, interface testing and workflow confirmation required before adoption.

Evidence by Source (18)

A geometry and dose-volume based performance monitoring of artificial intelligence models in…
De Kerf G, Claessens M, Raouassi F, et al. · Phys Imaging Radiat Oncol · 2023 · DOI 10.1016/j.phro.2023.100494
E1 L3

For information only. Not a clinical validation or endorsement; verify with the manufacturer and local commissioning.

Clinical evaluation of a deep learning segmentation model including manual adjustments afterwards…
Bakx N, Rijkaart D, van der Sangen M, et al. · Tech Innov Patient Support Radiat Oncol · 2023 · DOI 10.1016/j.tipsro.2023.100211
E1 L3

Validation of different automated segmentation models for target volume contouring in…
Meixner E et al. · Clin Transl Radiat Oncol · 2024 · DOI 10.1016/j.ctro.2024.100855
E1 L2

Automated segmentation of target volumes in breast cancer radiotherapy, impact on target size and dose…
Tang V et al. · Clin Transl Radiat Oncol · 2025 · DOI 10.1016/j.ctro.2025.100986
E1 L2

Clinical Implementation and Evaluation of Auto-Segmentation Tools for Multi-Site Contouring…
Heilemann G et al. · Phys Imaging Radiat Oncol · 2023 · DOI 10.1016/j.phro.2023.100515
E1 L2

A fully automated machine-learning-based workflow for radiation treatment planning in prostate cancer
Bolten JH et al. · Clin Transl Radiat Oncol · 2025 · DOI 10.1016/j.ctro.2025.100933
E1 L2

Autodelineation methods in a simulated fully automated proton therapy workflow for esophageal cancer
Populaire P et al. · Phys Imaging Radiat Oncol · 2024 · DOI 10.1016/j.phro.2024.100646
E1 L2

Automated deep learning segmentation and planning for left-sided breast radiotherapy with…
van Acht N et al. · Phys Imaging Radiat Oncol · 2026 · DOI 10.1016/j.phro.2026.100961
E1 L3

Comparison of the output of a deep learning segmentation model for locoregional breast…
Bakx N et al. · Technical Innovations & Patient Support in Radiation Oncology · 2023 · DOI 10.1016/j.tipsro.2023.100209
E1 L2

Large-scale dose evaluation of deep learning organ contours in head-and-neck radiotherapy by…
Mody P et al. · Physics and Imaging in Radiation Oncology · 2024 · DOI 10.1016/j.phro.2024.100572
E1 L2

Linking RayStation AI auto-contouring with Eclipse TPS: a scripted workflow for clinical integration
Ryczkowski A et al. · Rep Pract Oncol Radiother · 2025 · DOI 10.5603/rpor.109182
E1 L1t

Prospective blinded clinical evaluation of a fully automated workflow for prostate radiotherapy
Neugebauer D et al. · Phys Imaging Radiat Oncol · 2026 · DOI 10.1016/j.phro.2026.100999
E2 L4 Sets product level

For information only. Not a clinical validation or endorsement; verify with the manufacturer and local commissioning.

Statistical process control for performance monitoring and continuous quality assurance of deep…
van Acht N et al. · Phys Imaging Radiat Oncol · 2025 · DOI 10.1016/j.phro.2025.100873
E1 L3

Stress-testing pelvic autosegmentation algorithms using anatomical edge cases
Kanwar A et al. · Phys Imaging Radiat Oncol · 2023 · DOI 10.1016/j.phro.2023.100413
E1 L2

Validation of a deep-learning segmentation model for adult and pediatric head and neck radiotherapy…
Chen L et al. · Phys Imaging Radiat Oncol · 2024 · DOI 10.1016/j.phro.2023.100527
E1 L2

Geometric and Dosimetric Evaluation of a RayStation Deep Learning Model for Auto-Segmentation…
Sharma D et al. · Clinical Oncology · 2025 · DOI 10.1016/j.clon.2025.103796
E1 L2

A clinical evaluation of the performance of five commercial artificial intelligence contouring systems…
Doolan PJ et al. · Front. Oncol. · 2023 · DOI 10.3389/fonc.2023.1213068
E2 L3

Extensive clinical testing of Deep Learning Segmentation models for thorax and breast cancer…
Mikalsen SG et al. · Acta Oncologica · 2023 · DOI 10.1080/0284186X.2023.2270152
E1 L3

Key Features

- Deep learning algorithms
- Within TPS
- Guideline-based contouring

Technical Specifications

Input Formats:	DICOM
Output Formats:	DICOM-RTSTRUCT
Processing Time:	Seconds to minutes per structure
Integration:	Native TPS integration
Deployment:	On-premises
Modality Support:	CT; MR

Regulatory Information

CE: cleared FDA: 510k_cleared

For information only. Not a clinical validation or endorsement; verify with the manufacturer and local commissioning.

CE Status:	cleared
FDA Status:	510k_cleared
CE Details:	cleared (Class Class IIb)
FDA Details:	510k_cleared (K240398)

Contact Information

Website:	https://www.raysearchlabs.com/machine-learning-in-raystation/
Product URL:	https://www.raysearchlabs.com/media/publications/deep-learning-segmentation-model-catalogue/
Company URL:	https://www.raysearchlabs.com/

For information only. Not a clinical validation or endorsement; verify with the manufacturer and local commissioning.

5. ART-Plan Segmentation

Therapanacea

General Information

Product Name:	ART-Plan Segmentation
Company:	Therapanacea
Category:	Auto-Contouring
Version:	3.2.1
Release Date:	1/1/2021
Last Updated:	10/15/2026
Description:	Annotate is an AI-powered software module within ART-Plan+ providing zero-click automatic delineation of organs at risk (OARs) and lymph nodes for radiation therapy treatment planning. Cleared for CT-based contouring of structures across Head & Neck, Thorax, Breast, Pelvis (male and female), and SBRT Lung anatomical sites. MRI-based models are covered under the MR-Box module (K234068, K242822). (Sources: FDA K253091 Summary 2025-12-23; Therapanacea product page accessed 2026-06-16.) Current technical information identifies the legal device as ART-Plan+ v3.2.1; the public commercial presentation is ART-Plan. FDA K253091 concerns ART-Plan+ v3.1.0, not a separate clearance for v3.2.1. Source: https://therapanacea.com/technical-information-2, retrieved 2026-10-15.

Evidence Classification

Rigor E2 — External, independent or multi-center validation **Efficacy L3 — Decision/workflow efficacy**

Readiness R3 — Enhanced implementation readiness

+ Vendor-independent **- Multi-centre** **- Multi-national** **- Prospective** **+ External validation**

12 scored publication(s) · 12 listed · score origin: papers
Rigor: Application scores are maxima across attributable papers, not validation of every function. Refer to function-specific evidence below.
Efficacy: Application scores are maxima across attributable papers, not validation of every function. Refer to function-specific evidence below.
Readiness: CE & FDA cleared with 2 independent peer-reviewed evaluations explicitly naming the product. Cloud deployment reduces integration complexity. Standard commissioning (local QA, guideline alignment) expected.

Evidence by Source (12)

For information only. Not a clinical validation or endorsement; verify with the manufacturer and local commissioning.

Clinical Implementation and Evaluation of Auto-Segmentation Tools for Multi-Site Contouring…
Heilemann G et al. · Phys Imaging Radiat Oncol · 2023 · DOI 10.1016/j.phro.2023.100515
E1 L2

Evaluation of a quasi-automatic treatment planning workflow for head-and-neck cancer in radiotherapy
Westermayer M et al. · Physics and Imaging in Radiation Oncology · 2026 · DOI 10.1016/j.phro.2026.100983
E1 L2

Real-world performance evaluation of commercial autocontouring software for head and neck…
Young T et al. · Br J Radiol · 2025 · DOI 10.1093/bjr/tqaf098
E1 L3 Sets product level

Artificial intelligence contouring in radiotherapy for organs-at-risk and lymph node areas
Meyer P et al. · Radiation Oncology · 2024 · DOI 10.1186/s13014-024-02554-y
E1 L2

Investigation on performance of multiple AI-based auto-contouring systems in organs at risks…
Kim H et al. · Physical and Engineering Sciences in Medicine · 2024 · DOI 10.1007/s13246-024-01434-9
E2 L2 Sets product level

Evaluation of Two Commercial Artificial Intelligence Segmentation Systems for Radiation Therapy
DiTusa C et al. · Journal of Medical Physics · 2025 · DOI 10.4103/jmp.jmp_11_25
E2 L2

Head and neck automatic multi-organ segmentation on Dual-Energy Computed Tomography
Lê AT et al. · Physics and Imaging in Radiation Oncology · 2024 · DOI 10.1016/j.phro.2024.100654
E1 L2

Comparison of four synthetic CT generators for brain and prostate MR-only workflow in radiotherapy
Autret D et al. · Radiat Oncol · 2023 · DOI 10.1186/s13014-023-02336-y
E1 L2

A multi-centric evaluation of self-learning GAN based pseudo-CT generation software for low field…
Prunaretty J, Gungor G, Gevaert T, et al. · Frontiers in Oncology · 2023 · DOI 10.3389/fonc.2023.1245054
E2 L2

Standardizing MRI-only radiotherapy commissioning: Benchmark dataset and acceptance levels from…
Cusumano D, Maspero M, Vellini L, et al. · Radiotherapy and Oncology · 2026 · DOI 10.1016/j.radonc.2026.111530
E2 L2

Image-based evaluation of a commercial AI synthetic CT generator for brain and prostate…
Aire M, Matthews C, Jones J, et al. · Journal of Applied Clinical Medical Physics · 2026 · DOI 10.1002/acm2.70725
E1 L2

For information only. Not a clinical validation or endorsement; verify with the manufacturer and local commissioning.

Automatic target volume segmentation for offline adaptive head–and neck radi

J Applied Clin Med Phys · 2026 · DOI 10.1002/acm2.70479

E1 L2

Key Features

- CT/MR OAR and lymph-node segmentation
- MR GBM and multiple-metastasis tumour segmentation
- MR brachytherapy OAR segmentation
- MR-to-synthetic-CT generation
- Multi-modality rigid and deformable image fusion

Technical Specifications

Input Formats:	DICOM; DICOM-RTSTRUCT
Output Formats:	DICOM; DICOM-RTSTRUCT
Processing Time:	Full-body delineation in approximately 3 minutes; time varies with number of structures
Integration:	Cloud-based web application; DICOM export to TPS; PACS integration
Deployment:	Cloud-based (GDPR-compliant)
Modality Support:	CT; MRI; PET; CBCT; 4D-CT

Regulatory Information

CE: cleared FDA: 510k_cleared

CE Status:	cleared
FDA Status:	510k_cleared
CE Details:	cleared (Class Class IIb)
FDA Details:	510k_cleared (K253091)

Contact Information

Website:	https://therapanacea.com/products
Product URL:	https://therapanacea.com/products
Company URL:	https://therapanacea.com/

For information only. Not a clinical validation or endorsement; verify with the manufacturer and local commissioning.