

## Aurora Observatory Mission Report

The salinity array is calibrated every 45 days.

Calibration drift above 0.3 PSU triggers a maintenance ticket.

Section 1 covers the shoreline sensor cluster.

## Section 2: Thermal probes

Thermal probes are recalibrated after every 120 days of deployment.

Probe T-7 failed its March calibration and was replaced.