

ID	Name	Description	Rationale	Design Validation	Verification Method
SYS-023	Calibrated Field Uniformity of 96% of active volume	Calibration correction matrix for the Field Uniformity of 96% of the active volume shall be accurate to 1%	Modeling of the as-delivered electric field needs to be better than FD.	Full Scale Demonstrator	MIF Integrated Testing