

**KVP Healthcare Inc.**

Service Report

Service Call

PACIFIC GLOBAL MEDICAL CENTER

Schedule Date : Jun 20, 2025	Service Date : Jun 20, 2025	Location : NOVALICHES, QUEZON CITY
Machine: GENORAY DMX - 600	Reported Problem (if any) : SERVICE CALL	
Service Done: *PERFORMED SYSTEM RESTORE TO A PREVIOUS DATE RESTORE POINT, *COMPUTER RESTART. SYSTEM RESTORE INITIALIZING - DONE, *PERFORMED CONFIGURATION , *CHECK DEVICE INITIALIZATION, *FUNCTIONAL TEST - OKAY, *PERFORMED MECHANICAL RESET OF EXPOSURE HAND SWITCH.		
Remarks: *SYSTEM RUNNING & READY TO USE , *RECOMMEND THAT THE AIRCON SHOULD BE ALWAYS POWERED ON (24/7), *RECOMMEND ALSO THAT THE DEHUMIDIFIER IS ALWAYS ON & MONITORED	Status: Done	Service By: Jeffrey Disonglo, Cloyd Lojuco, Chad Laquindanum,

Generated at June 23, 2025