

(a) DPV staircase: 20 steps of 5 mV, 25 mV pulses
scope clipping level
Change from start (mV)
Time (s)
(b) One DPV cycle
width 48.0 ms
pulse
25.14 mV
step
4.76 mV
period 496.1 ms
Change from start (mV)
Time (ms)
(c) Step heights vs DAC LSB
10 LSB
5 mV set
9 LSB
Step height (mV)
Step number