



台灣福祉科技有限公司

儀器履歷表

名	稱	標準型雙向扭力扳手 (2103238898)	品 牌 / 型 號 / 規 格	FULLER/ 5-25 型號：6474470 序號：2101233998			
編	號	SM26	允	差	±5.0%		
精	度	0.5N.m	保	管	單	位	廠務單位-江冠毅
校 驗 類 別	☐ 內校 ☒ 外校		校 驗 週 期		一年		
	☐ 免校		汰 換 週 期		-		
送修紀錄							
送修日	送修原因		完成日	維修單位／人員			
校驗紀錄							
校 驗 日	校驗單位	校驗內容	合格判定	判定人	預計下次校 驗日期	備註	
2021.7.12	-	新購	OK	黃嘉鈺	2022.7.12	專屬 SIETNTA 使用	
2022.9.13	MTC	F11-09-048-01	OK	黃嘉鈺	2023.9.13	TAF	
2023.10.6	MTC	F12-09-263-01	OK	黃嘉鈺	2024.10.6	TAF	
移轉紀錄							
日 期	移 交 人	移 交 單 位	交 接 人	交 接 單 位	原因備註		
2021/7/12	新購	新購	江冠毅	廠務單位			



量測科技股份有限公司
Measurement Technology Co., Ltd.
CALIBRATION REPORT



1734

NO. F12-09-263-01

Issue Date

發佈日期 2023/10/6

儀器校正報告書

Page 1 of 2

Applicant 申請者		台灣福祉科技有限公司	
Address 地址		新北市樹林區八德街 230 號	
Manufacturer 製造廠商	FULLER	Model No. 型號	6474470
		Serial No. 序號	2101233998
Description 儀器名稱	扭力扳手		Received Date 收件日期
			2023/9/22
Procedure used 校正程序	詳見內頁說明	Calibration Date 校正日期	2023/10/3
Condition of calibration 校正時之環境條件	Temp 溫度	21 °C ~ 23 °C	R.H. 相對濕度
		55 % ~ 65 %	
Standards Employed 校正時使用之標準器			
Equipment 儀器名稱	Manufacturer 製造廠商	Model 型號	Serial Number 序號
TORQUE CALIBRATOR	Norbar	50593.LOG/43228	73031/72563
TORQUE CALIBRATOR	Norbar	50772.LOG/43228	73013/72563
Traceability 追溯機構	Report No. 報告號碼	Calibration Date 校正日期	Due Date 有效日期
MTC(TAF 1734) MTC(TAF 1734)	M12-08-247-03 M12-08-247-04	2023/9/5 2023/9/5	2024/9/4 2024/9/4

MTC in hereby certifies that the equipment noted herein has been compared with the above listed standards. The standards used to perform this calibration are traceable to the National Measurement Laboratory (NML) of ROC or National Time and Frequency Standard Laboratory(NTFSL) of ROC. The MTC laboratories are in compliance with ISO/IEC 17025.

量測科技股份有限公司特此證明本受校儀器已與上列標準器實施比較校正，上述之標準器均可追溯至中華民國國家度量衡標準實驗室或國家時間與頻率標準實驗室。本公司所屬實驗室之運作與管理均符合 ISO/IEC 17025 之要求



Approved Signatory
報告簽署人



Lab. Head
實驗室主管





一. 校正結果：

順時方向

標準值(N.m)	器示值(N.m)	器 差(N.m)	擴充不確定度(%)
48.1	50.0	1.9	0.50
77.2	80.0	2.8	0.44
117.4	120.0	2.6	0.30
158.6	160.0	1.4	0.30
200.8	200.0	-0.8	0.29

二、說 明：

- (一) 本校正報告書僅對此校正件有效，且未獲得本實驗室同意，此校正報告書不得摘錄複製，但全文複製除外。
- (二) 器 差 = 器示值 - 標準值
- (三) 擴充不確定度：信賴水準約為 95 %，依照有效自由度計算結果查t分配表得涵蓋因子(k)等於2。
- (四) 本校正方法係參照本公司制訂之扭力扳手校正程序(MT-C-99-011)
- (五) 順時或逆時方向以對待測件施力的方向為準。
- (六) 參考資料：
 - (1) 扭力扳手校正程序(MT-C-99-011)十版，2021年。
 - (2) 扭力扳手校正系統評估報告(MT-S-99-012)五版，2021年。
 - (3) 扭力國際標準規範(ISO 6789-2:2017)。

(以 下 空 白)