

Measurements, and Analog to Digital conversion

ENGR 217 – LECTURE 1

8/24/2023

1 Measurements

Many of the measurements that will be taken pertain to electrical characteristics changing due to environmental factors. Most of the time what we measure is changes in voltage. We need to convert changes in voltage which are analog (continuous) values to digital (discrete) values, since computers are digital.

Electrical Properties of Matter

- voltage
- Resistance
- Capacitance
- Inductance
- others

Voltage

It is a measure of electric potential. The, electric potential energy of a charged particle in an electric field, divided by the amount of charge (**Joules/Coulomb**). Volts are a difference in potential between two points (some point and ground, some point and infinity, two different points, etc.)

THings to think about with an ADC

- Speed
- Input Range
- Resolution
- Differential or single ended

Differential v. Single Ended

Single-ended are taken relative to the DAQ electrical ground (in your case, the air table is plugged in to the building, so the building ground.) These can be done with a single wire.

Differential looks at the voltage difference between two input wires, which may or may not have anything to do with the DAQ ground.

- differential is more accurate
- Harder though: need two wires and need to isolate the ADC system from the ground

Speed

This is the measure of samples per second(Hertz). For our lab we get by with thousands of samples per second – kilo Hertz - kHz.

Range

Maximum Voltage difference →Values can only be read in this range. If a value is too high or too low, it will be "clipped" to the crossed extrema.

- You can often use signal conditioning to make readings of out-of-range values, or make readings which use more of the range