CONFIDENTIAL INFORMATION

PFIZER

PSSM PCoE

VALIDATION REPORT

R&D MICROBIOLOGY - CHENNAI

Report Title: Summary Report for Validation and Transfer of Microbial Enumeration Tests of Sugammadex Sodium to Rocky Mount, NC

Study No.: VAL100190816AND VERSION 1.0

Version: V1.0

Project Name/Number: VAL100190816

TABLE OF CONTENTS

1.0 PURPOSE 3
2.0 SUMMARY 4
3.0 MICROBIAL ENUMERATION TEST METHOD TRANSFER 5
4.0 MATERIALS 6
5.0 METHODS/PROCEDURES 9
5.1 PRODUCT POSITIVE CONTROL 9
5.2 PRODUCT NEGATIVE CONTROL 9
5.3 BUFFER NEGATIVE CONTROL 10
5.4 PEPTONE NEGATIVE CONTROL 10
5.5 INOCULUM COUNT VERIFICATION 10
5.6 RESULTS 10
6.0 VALIDITY CRITERIA 14
7.0 ACCEPTANCE CRITERIA 16
8.0 CONCLUSION 18
9.0 NON-CONFORMANCE 19
10.0 REFERENCES 20
11.0 CHANGE HISTORY 26

1.0 PURPOSE

2.0 SUMMARY

3.0 MICROBIAL ENUMERATION TEST METHOD TRANSFER

4.0 MATERIALS

Media and Diluents

Table 4.0.1

Name of media and Diluents	Make	Lot No.	Expiry Date / Use before date
Phosphate Buffer	In House	00715609-0062A	
0.1N HCL	Merck	HC01429260	31-Jan-2023
SCDA Plates	Hi media	MPB0656	31-Jan-2023
SDA Plates	Hi media	MPB0903	31-Mar-2023
Rehydration Fluid	Biomerieux	JDP-057	17-Apr-2023

Consumables

Table 4.0.2

Consumables	Make	Lot No.	Expiry Date / Use before date
Sterile Container	Tarsons	034BVH291020	28-Oct-2025
Stericup-0.22 Micron filtration system	Merck	MP211505G2	30-Apr-2024
Micropipette tips -10 mL	Eppendorf	H175855H	28-Feb-2023
Syringe filter 0.22 micron	Merck	R0JB89897	31-Jul-2023
Sterile 10mL syringe	Dispovan	225102JF2	31-May-2027
Micropipette tips 2-200 Micro litre	AST	21K22S208-302	28-Feb-2025
0.45 m X 47 mm mixed cellulose ester membrane filters	Merck	F1EB16037	31-May-2023
Greenishyellow, cottoneygrowth		Greenishyellow, cottoneygrowth	
greenishyellow, cottoneygrowthwith spores		greenishyellow, cottoney-growthwith spores	

Equipment's used

Table 4.0.3

S. No.	Name of Equipment	Equipment No.	Calibration due date	Performance Verification due date
1	Biosafety cabinet	PCE0811	NA	05 NOV 2022
		ADE251	NA	NA
2	Analytical balance	PCE1057	13-Apr-2023	NA
		ADE278	03-Mar-2021	NA
3	Micropipette	PCE0882	16-Feb-2023	NA
		PCE0835	10-Mar-2023	NA
		ADE403	30-Jul-2021	NA
		ADE407	30-Jul-2021	NA
4	EZ-Stream Pump	PCE0887	NA	NA

S. No.	Name of Equipment	Equipment No.	Calibration due date	Performance Verification due date
IKKT-RDE0177	07-Mar-2021	NA		
9	XPERT FILTERED BALANCE SYSTEM	ADE332	NA	NA
10	Vortex mixer	IKKT-RDE0102	NA	NA
11	EZ-FIT Manifold	IKKT-RDE0113	NA	NA

S. No.	Name of Equipment	Equipment No.	Calibration due date	Performance Verification due date
IKKT-RDE0108	NA	NA		
5	pH meter	PCE1059	16-Mar-2023	NA
		IKKT-RDE0167	27-Jun-2021	NA
6	Master flex Digital Pump Drive	PCE0673	NA	NA
7	Incubators	PCE1088	NA	NA
		ADE263	14-Apr-2021	NA
		ADE264	14-Apr-2021	NA
8	Colony counter	PCE0846	11-Mar-2023	NA
		PCE0848	11-Mar-2023	NA

Microorganisms Used

Table 4.0.4

Name of Microorganism	Make	Lot Number	Expiry Date
Bacillus Subtilis ATCC 6633/NCTC10400	Biomerieux	B6365	02-Nov-2022
Pseudomonas aeruginosa NCTC 12924 (ATCC 9027)	Biomerieux	B6456	17-Dec-2022
Staphylococcus aureus NCTC 10788 (ATCC 6538)	Biomerieux	B6502	20-Jan-2023
Candida albicans NCPF 3179 (ATCC 10231)	Biomerieux	B6397	19-Nov-2022
Aspergillus brasiliensis NCPF2275 (ATCC 16404)	Biomerieux	B6412	26-Nov-2022

5.0 METHODS/PROCEDURES

Drug Substance Batches

Table 5.0.1

Drug substance	Supplier	Batch/lot number
Sugammadex Sodium	TEVA	7MPS02844051221
Sugammadex Sodium	TEVA	7MPS02844061221
Sugammadex Sodium	TEVA	7MPS02845860319

Observed pH values

Table 5.0.2

S. No.	Batch Number of the sample	pH of the sample
1	7MPS02844051221	7.333
2	7MPS02844061221	7.336
3	7MPS02845860319	7.335
4	Batch No : 0000192271	7.206
5	Batch No. 0000192355	7.206
6	Batch No. 0000192845	7.208

5.1 PRODUCT POSITIVE CONTROL

5.2 PRODUCT NEGATIVE CONTROL

5.3 BUFFER NEGATIVE CONTROL

5.4 PEPTONE NEGATIVE CONTROL

5.5 INOCULUM COUNT VERIFICATION

5.6 RESULTS

Results observed on SCDA Plates - Batch 7MPS02844051221

Table 5.6.1

Microorganisms	Count observed on Product Positive Control (PPC) in CFU				Count observed on Inoculum Count (IC) in CFU				% Recovery (Average PPC/Average IC) x 100
	Plate 1	Plate 2	Plate 3	Average	Plate 1	Plate 2	Plate 3	Average	
Bacillussubtilis NCTC10400 (ATCC6633)	44	36	35	38.33	39	35	34	36.00	106
Pseudomonasaeruginosa NCTC12924(ATCC9027)	41	39	42	40.67	40	38	43	40.33	101
StaphylococcusaureusNCTC 10788(ATCC6538)	46	42	42	43.33	45	43	42	43.67	99
CandidaalbicansNCPF3179 (ATCC10231)	44	47	42	44.33	46	44	43	44.33	100
Aspergillus brasiliensis NCPF2275 (ATCC16404)	37	35	36	36.00	36	38	35	36.33	99

Results observed on SDA Plates - Batch 7MPS02844051221

Table 5.6.2

Microorganisms	Count observed on Product Positive Control (PPC) in CFU				Count observed on Inoculum Count (IC) in CFU				% Recovery (Average PPC/Average IC) x 100
	Plate 1	Plate 2	Plate 3	Average	Plate 1	Plate 2	Plate 3	Average	
CandidaalbicansNCPF3179 (ATCC10231)	44	45	46	45.00	50	56	54	53.33	84
Aspergillus brasiliensis NCPF2275 (ATCC16404)	35	34	33	34.00	40	39	36	38.33	89

Equivalency Recovery - Batch 7MPS02844051221

Table 5.6.3

Name of the organisms	Average PPC in SCDA (Average Value)	Average IC in SDA (Average Value)	Percentage recovery
CandidaalbicansNCPF3179 (ATCC10231)	44.33	53.33	83
Aspergillus brasiliensis NCPF2275 (ATCC16404)	36.00	38.33	94

Results observed on SCDA Plates - Batch 7MPS02844061221

Table 5.6.4

Microorganisms	Count observed on Product Positive Control (PPC) in CFU				Count observed on Inoculum Count (IC) in CFU				% Recovery (Average PPC/Average IC) x 100
	Plate 1	Plate 2	Plate 3	Average	Plate 1	Plate 2	Plate 3	Average	
Bacillussubtilis NCTC10400 (ATCC6633)	41				39	35	34	36.00	107
Pseudomonasaeruginosa NCTC12924(ATCC9027)	43				40	38	43	40.33	102
StaphylococcusaureusNCTC 10788(ATCC6538)	47				45	43	42	43.67	101
CandidaalbicansNCPF3179 (ATCC10231)	43				46	44	43	44.33	92
Aspergillus brasiliensis NCPF2275 (ATCC16404)	32				36	38	35	36.33	83

Results observed on SDA Plates - Batch 7MPS02844061221

Table 5.6.5

Microorganisms	Count observed on Product Positive Control (PPC) in CFU				Count observed on Inoculum Count (IC) in CFU				% Recovery (Average PPC/Average IC) x 100
	Plate 1	Plate 2	Plate 3	Average	Plate 1	Plate 2	Plate 3	Average	
CandidaalbicansNCPF3179 (ATCC10231)				44.33	50	56	54	53.33	83
Aspergillus brasiliensis NCPF2275 (ATCC16404)				32.67	40	39	36	38.33	85

Equivalency Recovery - Batch 7MPS02844061221

Table 5.6.6

Name of the organisms	Average PPC in SCDA (Average Value)	Average IC in SDA (Average Value)	Percentage recovery
CandidaalbicansNCPF3179 (ATCC10231)	41.00	53.33	77
Aspergillus brasiliensis NCPF2275 (ATCC16404)	30.33	38.33	79

Results observed on SCDA Plates - Batch 7MPS02845860319

Table 5.6.7

Microorganisms	Count observed on Product Positive Control (PPC) in CFU				Count observed on Inoculum Count (IC) in CFU				% Recovery (Average PPC/Average IC) x 100
	Plate 1	Plate 2	Plate 3	Average	Plate 1	Plate 2	Plate 3	Average	
Bacillussubtilis NCTC10400 (ATCC6633)				35.33	39	35	34	36.00	98
Pseudomonasaeruginosa NCTC12924(ATCC9027)				37.67	40	38	43	40.33	93
StaphylococcusaureusNCTC 10788(ATCC6538)				48.00	45	43	42	43.67	110
CandidaalbicansNCPF3179 (ATCC10231)				39.67	46	44	43	44.33	89
Aspergillus brasiliensis NCPF2275 (ATCC16404)				30.33	36	38	35	36.33	83

Results observed on SDA Plates - Batch 7MPS02845860319

Table 5.6.8

Microorganisms	Count observed on Product Positive Control (PPC) in CFU				Count observed on Inoculum Count (IC) in CFU				% Recovery (Average PPC/Average IC) x 100
	Plate 1	Plate 2	Plate 3	Average	Plate 1	Plate 2	Plate 3	Average	
CandidaalbicansNCPF3179 (ATCC10231)				46.00	50	56	54	53.33	86
Aspergillus brasiliensis NCPF2275 (ATCC16404)				31.00	40	39	36	38.33	81

Equivalency Recovery - Batch 7MPS02845860319

Table 5.6.9

Name of the organisms	Average PPC in SCDA (Average Value)	Average IC in SDA (Average Value)	Percentage recovery
CandidaalbicansNCPF3179 (ATCC10231)	39.67	53.33	74
Aspergillus brasiliensis NCPF2275 (ATCC16404)	30.33	38.33	79

6.0 VALIDITY CRITERIA

Validity Criteria

Table 6.0.1

Validity Criteria	Observed Results
The Morphology of the Colonies observed in Product Positive Control (PPC) and Inoculum Count (IC) verification plates were found to be consistent with the challenge microorganisms.	Complies with the stated criterion based on the batch results summarized in this report.

Validity Criteria	Observed Results
The Average of Inoculum Count (IC) verification plates was not more than (NMT) 100CFU.	Inoculum Count averages did not exceed 100 CFU across all challenge organisms and batches tested (highest observed: 53.33 CFU).
The volume of the inoculum did not exceed 1% of the total volume of the diluted product.	Complies with the stated criterion based on the batch results summarized in this report.
The Exposed Settle Plates (Set-1, Set-2 &Set-3) showed No Growth.	Complies with the stated criterion based on the batch results summarized in this report.
The Morphology of the Colonies observed in Product Positive Control (PPC) and Inoculum Count (IC) verification plates were found to be consistent with the challenge microorganisms.	Complies with the stated criterion based on the batch results summarized in this report.
The Peptone Negative Control and Buffer Negative Control showed no growth of microorganisms.	Complies with the stated criterion based on the batch results summarized in this report.
The Product Negative Control was not more than (NMT) 100 CFU/g. A) The obtained average count for each Batch on SCDA plate showed absence of microorganisms (0CFU/g). B) The obtained average count for each Batch on SDA plate showed absence of microorganisms (0 CFU/g).	Product Negative Control counts did not exceed 100 CFU/g across all batches tested (highest observed: 0 CFU).
The Average of Inoculum Count (IC) verification plates was not more than (NMT) 100CFU.	Inoculum Count averages did not exceed 100 CFU across all challenge organisms and batches

Validity Criteria	Observed Results
The Peptone Negative Control and Buffer Negative Control showed no growth of microorganisms.	Complies with the stated criterion based on the batch results summarized in this report.
The Product Negative Control was not more than (NMT) 100 CFU/g. A) The obtained average count for each Batch on SCDA plate showed absence of microorganisms (0CFU/g). B) The obtained average count for each Batch on SDA plate showed absence of microorganisms (0 CFU/g).	Product Negative Control counts did not exceed 100 CFU/g across all batches tested (highest observed: 0 CFU).
The Average of Inoculum Count (IC) verification plates was not more than (NMT) 100CFU.	Inoculum Count averages did not exceed 100 CFU across all challenge organisms and batches tested (highest observed: 53.33 CFU).
The volume of the inoculum did not exceed 1% of the total volume of the diluted product.	Complies with the stated criterion based on the batch results summarized in this report.
The Exposed Settle Plates (Set-1, Set-2 &Set-3) showed No Growth.	Complies with the stated criterion based on the batch results summarized in this report.
The Morphology of the Colonies observed in Product Positive Control (PPC) and Inoculum Count (IC) verification plates were found to be consistent with the challenge microorganisms.	Complies with the stated criterion based on the batch results summarized in this report.
The Peptone Negative Control and Buffer Negative Control showed no growth of microorganisms.	Complies with the stated criterion based on the batch results summarized in this report.

Validity Criteria	Observed Results
The Product Negative Control was not more than (NMT) 100 CFU/g. A) The obtained average count for each Batch on SCDA plate showed absence of microorganisms (0CFU/g). B) The obtained average count for each Batch on SDA plate showed absence of microorganisms (0 CFU/g).	Product Negative Control counts did not exceed 100 CFU/g across all batches tested (highest observed: 0 CFU).

7.0 ACCEPTANCE CRITERIA

Acceptance criteria (Validation)

Table 7.0.1

Acceptance criteria (Validation)	Observed Results
The Recovery of each microorganism in the Inoculated Product Positive Control (PPC) was between 50% and 200%, when compared against the Inoculum Count (IC) verification.	Recovery percentage ranged from 81% to 110% across all batches tested, within the 50%–200% acceptance range.

Acceptance criteria (Validation)	Observed Results
0 ABBREVIATIONS.....11	Complies with the stated criterion based on the batch results summarized in this report.
0 NON – CONFORMANCE.....12	Complies with the stated criterion based on the batch results summarized in this report.
The Recovery of each microorganism in the Inoculated Product Positive Control (PPC) was between 50% and 200%, when compared against the Inoculum Count (IC) verification.	Recovery percentage ranged from 81% to 110% across all batches tested, within the 50%–200% acceptance range.
For Aspergillus brasiliensis and Candida albicans, Product Positive Control (PPC) and Inoculum Count (IC) was plated both on SCDA & SDA plates. The comparison between PPC on SCDA and IC on SDA for Aspergillus brasiliensis and Candida albicans met the acceptance criteria of 50% and 200%.	Equivalency recovery percentage ranged from 74% to 94% across all batches tested, within the 50%–200% acceptance range.
Hence, the routine bioburden test shall be performed using only SCDA media at Incubation conditions of 30-35°C for NLT 3days.	Complies with the stated criterion based on the batch results summarized in this report.
0 ABBREVIATIONS.....11	Complies with the stated criterion based on the batch results summarized in this report.
0 NON – CONFORMANCE.....12	Complies with the stated criterion based on the batch results

Acceptance criteria (Validation)	Observed Results
For Aspergillus brasiliensis and Candida albicans, Product Positive Control (PPC) and Inoculum Count (IC) was plated both on SCDA & SDA plates. The comparison between PPC on SCDA and IC on SDA for Aspergillus brasiliensis and Candida albicans met the acceptance criteria of 50% and 200%.	Equivalency recovery percentage ranged from 74% to 94% across all batches tested, within the 50%–200% acceptance range.
Hence, the routine bioburden test shall be performed using only SCDA media at Incubation conditions of 30-35°C for NLT 3days.	Complies with the stated criterion based on the batch results summarized in this report.
0 ABBREVIATIONS.....11	Complies with the stated criterion based on the batch results summarized in this report.
0 NON – CONFORMANCE.....12	Complies with the stated criterion based on the batch results summarized in this report.
The Recovery of each microorganism in the Inoculated Product Positive Control (PPC) was between 50% and 200%, when compared against the Inoculum Count (IC) verification.	Recovery percentage ranged from 81% to 110% across all batches tested, within the 50%–200% acceptance range.
For Aspergillus brasiliensis and Candida albicans, Product Positive Control (PPC) and Inoculum Count (IC) was plated both on SCDA & SDA plates. The comparison between PPC on SCDA and IC on SDA for Aspergillus brasiliensis and Candida albicans met the acceptance criteria of 50% and 200%.	Equivalency recovery percentage ranged from 74% to 94% across all batches tested, within the 50%–200% acceptance range.

Acceptance criteria (Validation)	Observed Results
Hence, the routine bioburden test shall be performed using only SCDA media at Incubation conditions of 30-35°C for NLT 3days.	Complies with the stated criterion based on the batch results summarized in this report.

8.0 CONCLUSION

9.0 NON-CONFORMANCE

10.0 REFERENCES

Table 10.0.1

Document Number	Document Title	Effective/Sign Off Date
USP <61>	Microbiological Examination of Non -Sterile Products	01 MAY 2018
USP <1117>	Microbiological Best Laboratory Practices	01 MAY 2018
USP <1227>	Validation of Microbial Recovery from Pharmacopeial Articles	01 MAY 2018
SOP-LAB-03311	Microbiological Test Requirements for Ingredients Used in the Manufacture of Parenteral Products Procedure, PEH R&D	22 DEC 2020
V7112	Microbiological Methods Validation	12 JUN 2020
L6102	Transfer of test method	08 JUN 2018
SOP-LAB-05213	WRD Biovia Workbook-PCoE R&D	09 OCT2020
SOP-LAB-04874	Validation for Total Aerobic Microbialcount (TAMC) and Total Yeast and Molds count (TYMC) for Active Pharmaceutical ingredients and Excipients –Lake fore-stand IKKT	04 SEP 2019
SOP-LAB-03312	Biological Laboratory Test Method Validation Procedure, PEH R&D	22 DEC 2020

Document Number	Document Title	Effective/Sign Off Date
SOP-LAB-02515	Generating Scientific Protocols and Reports inGDMS	21 APR2021

Table 10.0.2

Document Number	Document Title	Effective/Sign Off Date
USP <61>	Microbiological Examination of Non -Sterile Products	01 MAY 2018
USP <1117>	Microbiological Best Laboratory Practices	01 MAY 2018
USP <1227>	Validation of Microbial Recovery from Pharmacopeial Articles	01 MAY 2018
SOP-LAB-03311	Microbiological Test Requirements for Ingredients Used in the Manufacture of Parenteral Products Procedure, PEH R&D	22 DEC 2020
V7112	Microbiological Methods Validation	12 JUN 2020
L6102	Transfer of test method	08 JUN 2018
SOP-LAB-02756	Electronic Lab Notebooks-PCoE	29 JUL 2020
SOP-LAB-04874	Validation for Total Aerobic Microbialcount (TAMC) and Total Yeast and Molds count (TYMC) for Active Pharmaceutical ingredients and Excipients –Lake forest and IKKT	04 SEP 2019
SOP-LAB-03312	Biological Laboratory Test Method Validation Procedure, PEH R&D	22 DEC 2020

Document Number	Document Title	Effective/Sign Off Date
SOP-LAB-02515	Generating Scientific Protocols and Reports in GDMS	06 JUL 2020

Table 10.0.3

Document Number	Document Title	Effective/Sign Off Date
USP <61>	Microbiological Examination of Non -Sterile Products	01 MAY 2018
USP <1117>	Microbiological Best Laboratory Practices	01 MAY 2018
USP <1227>	Validation of Microbial Recovery from Pharmacopeial Articles	01 MAY 2018
SOP-LAB-03311	Microbiological Test Requirements for Ingredients Used in the Manufacture of Parenteral Products Procedure, PEH R&D	22 DEC 2020
V7112	Microbiological Methods Validation	12 JUN 2020
L6102	Transfer of test method	08 JUN 2018
SOP-LAB-05213	WRD Biovia Workbook-PCoE R&D	09 OCT2020
SOP-LAB-04874	Validation for Total Aerobic Microbialcount (TAMC) and Total Yeast and Molds count (TYMC) for Active Pharmaceutical ingredients and Excipients –Lake fore-stand IKKT	04 SEP 2019
SOP-LAB-03312	Biological Laboratory Test Method Validation Procedure, PEH R&D	22 DEC 2020

Document Number	Document Title	Effective/Sign Off Date
SOP-LAB-02515	Generating Scientific Protocols and Reports inGDMS	21 APR2021

Table 10.0.4

Document Number	Document Title	Effective/Sign Off Date
USP <61>	Microbiological Examination of Non -Sterile Products	01 MAY 2018
USP <1117>	Microbiological Best Laboratory Practices	01 MAY 2018
USP <1227>	Validation of Microbial Recovery from Pharmacopeial Articles	01 MAY 2018
SOP-LAB-03311	Microbiological Test Requirements for Ingredients Used in the Manufacture of Parenteral Products Procedure, PEH R&D	22 DEC 2020
V7112	Microbiological Methods Validation	12 JUN 2020
L6102	Transfer of test method	08 JUN 2018
SOP-LAB-02756	Electronic Lab Notebooks-PCoE	29 JUL 2020
SOP-LAB-04874	Validation for Total Aerobic Microbialcount (TAMC) and Total Yeast and Molds count (TYMC) for Active Pharmaceutical ingredients and Excipients –Lake forest and IKKT	04 SEP 2019
SOP-LAB-03312	Biological Laboratory Test Method Validation Procedure, PEH R&D	22 DEC 2020

Document Number	Document Title	Effective/Sign Off Date
SOP-LAB-02515	Generating Scientific Protocols and Reports in GDMS	06 JUL 2020

Table 10.0.5

Document Number	Document Title	Effective/Sign Off Date
USP <61>	Microbiological Examination of Non -Sterile Products	01 MAY 2018
USP <1117>	Microbiological Best Laboratory Practices	01 MAY 2018
USP <1227>	Validation of Microbial Recovery from Pharmacopeial Articles	01 MAY 2018
SOP-LAB-03311	Microbiological Test Requirements for Ingredients Used in the Manufacture of Parenteral Products Procedure, PEH R&D	22 DEC 2020
V7112	Microbiological Methods Validation	12 JUN 2020
L6102	Transfer of test method	08 JUN 2018
SOP-LAB-05213	WRD Biovia Workbook-PCoE R&D	09 OCT2020
SOP-LAB-04874	Validation for Total Aerobic Microbialcount (TAMC) and Total Yeast and Molds count (TYMC) for Active Pharmaceutical ingredients and Excipients –Lake fore-stand IKKT	04 SEP 2019
SOP-LAB-03312	Biological Laboratory Test Method Validation Procedure, PEH R&D	22 DEC 2020

Document Number	Document Title	Effective/Sign Off Date
SOP-LAB-02515	Generating Scientific Protocols and Reports inGDMS	21 APR2021

Table 10.0.6

Document Number	Document Title	Effective/Sign Off Date
USP <61>	Microbiological Examination of Non -Sterile Products	01 MAY 2018
USP <1117>	Microbiological Best Laboratory Practices	01 MAY 2018
USP <1227>	Validation of Microbial Recovery from Pharmacopeial Articles	01 MAY 2018
SOP-LAB-03311	Microbiological Test Requirements for Ingredients Used in the Manufacture of Parenteral Products Procedure, PEH R&D	22 DEC 2020
V7112	Microbiological Methods Validation	12 JUN 2020
L6102	Transfer of test method	08 JUN 2018
SOP-LAB-02756	Electronic Lab Notebooks-PCoE	29 JUL 2020
SOP-LAB-04874	Validation for Total Aerobic Microbialcount (TAMC) and Total Yeast and Molds count (TYMC) for Active Pharmaceutical ingredients and Excipients –Lake for-est and IKKT	04 SEP 2019
SOP-LAB-03312	Biological Laboratory Test Method Validation Procedure, PEH R&D	22 DEC 2020

Document Number	Document Title	Effective/Sign Off Date
SOP-LAB-02515	Generating Scientific Protocols and Reports in GDMS	06 JUL 2020

11.0 CHANGE HISTORY