

PRESSURE TRANSDUCER
FINAL CALIBRATION

0.00 - 7.00 BARG
Excitation 28.000 Vdc

Job: WHS0059544 Serial: BG042821D034
Model: PXM309-007GI Tested By: TG
Date: 3/30/2022 Temperature Range: -20 to +85C
Calibrated: 0.00 - 7.00 BARG Specfile: PXM309-I+=1bar

Pressure BARG	Unit Data mA
0.00	4.004
3.50	12.047
7.00	20.061
3.50	12.044
0.00	4.017

Balance 4.004 mA
Sensitivity 16.057 mA

ELECTRICAL LEAKAGE: PASS

PRESSURE CONNECTION/FITTING: G 1/4-19 Male

ELECTRICAL WIRING/CONNECTOR: BLACK = -EXCITATION

RED = +EXCITATION

This Calibration was performed using Instruments and Standards that are traceable to the United States National Institute of Standards Technology.

S/N	Description	Range	Reference	Cal Cert
41000FUN	AUTO Mensor 1200	0 - 82.74 BARG	C-3051	C-3051
US37046202	HP 34970A DMM	Unit Under Test	C-1315	N/A

Q.A. Representative : *Torri Granel*

Date: 3/30/2022

This transducer is tested to & meets published specifications. This product is controlled in a bonded storage location until sold. The recalibration date of this product should not exceed either (A) one year from the date of shipment or (B) five years from the date on this certificate, whichever is less. Calibration cycles less than the provided guidelines should be managed per the customer's quality management system.

Omega Engineering Inc., 800 Connecticut Ave., Norwalk, CT 06854
<http://www.omega.com> email: info@omega.com phone (800) 826-6342

PRESSURE TRANSDUCER
FINAL CALIBRATION0.00 - 7.00 BARG
Excitation 28.000 Vdc

Job: WHS0059544 Serial: BG042821D041
 Model: PXM309-007GI Tested By: TG
 Date: 3/30/2022 Temperature Range: -20 to +85C
 Calibrated: 0.00 - 7.00 BARG Specfile: PXM309-I+=1bar

Pressure BARG	Unit Data mA
0.00	3.995
3.50	12.022
7.00	19.996
3.50	12.024
0.00	4.007

Balance 3.995 mA
 Sensitivity 16.001 mA

ELECTRICAL LEAKAGE: PASS

PRESSURE CONNECTION/FITTING: G 1/4-19 Male

ELECTRICAL WIRING/CONNECTOR: BLACK = -EXCITATION

RED = +EXCITATION

This Calibration was performed using Instruments and Standards that are traceable to the United States National Institute of Standards Technology.

S/N	Description	Range	Reference	Cal Cert
41000FUN	AUTO Mensor 1200	0 - 82.74 BARG	C-3051	C-3051
US37046202	HP 34970A DMM	Unit Under Test	C-1315	N/A

Q.A. Representative : *Toni Granel*

Date: 3/30/2022

This transducer is tested to & meets published specifications. This product is controlled in a bonded storage location until sold. The recalibration date of this product should not exceed either (A) one year from the date of shipment or (B) five years from the date on this certificate, whichever is less.

Calibration cycles less than the provided guidelines should be managed per the customer's quality management system.

Omega Engineering Inc., 800 Connecticut Ave., Norwalk, CT 06854
<http://www.omega.com> email: info@omega.com phone (800) 826-6342