

NATIONAL LEGAL METROLOGY EVALUATION CENTRE

Accreditation No: NABL-MET-2026-0982 | Metrology Complex, CSIR Road, New Delhi 110012, India

TYPE EVALUATION TEST REPORT — OIML R 76

Report No: LM/2026/00001 | Evaluation ID: TEST-2026-0001 | Standard: OIML R 76-1:2006

1. INSTRUMENT UNDER EVALUATION

Manufacturer: Avery Weigh-Tronix India Ltd.	Model Name: WeighMaster Bench Scale	Serial Number: SN-2026-WM-98101	Instrument Type: Electronic Platform Scale
Accuracy Class: Class III (Medium)	Max Capacity (Max): 30 kg	Min Capacity (Min): 0.1 kg	Scale Interval (e): 0.01 kg

2. LABORATORY & ENVIRONMENTAL CONDITIONS

Ambient Temperature: 21.50 °C	Relative Humidity: 45.00 %	Atmospheric Pressure: 1012.50 hPa	Standards Used: E2 Mass Set
---	--------------------------------------	---	---------------------------------------

3. METROLOGICAL TEST RESULTS (OIML R 76-1:2006)

• Weighing Performance Test (Clause A.4.4) [Result: PASS]

#	Applied Load	Indication	Changeover (L)	Error (E)	MPE Limit	Status
1	0 kg	0 kg	-	0 kg	±0.005 kg	PASS
2	5 kg	5.002 kg	0.005	+0.002 kg	±0.005 kg	PASS
3	10 kg	10.003 kg	0.005	+0.003 kg	±0.01 kg	PASS
4	20 kg	20.004 kg	0.005	+0.004 kg	±0.01 kg	PASS
5	30 kg	30.005 kg	0.005	+0.005 kg	±0.015 kg	PASS
6	20 kg	20.004 kg	0.005	+0.004 kg	±0.01 kg	PASS
7	0 kg	0 kg	-	0 kg	±0.005 kg	PASS

• Repeatability Test (Clause A.4.10) [Result: PENDING]

• Eccentric Loading Test (Clause A.4.7) [Result: PENDING]

FINAL METROLOGICAL VERDICT: INCOMPLETE

Based on the results of the evaluation carried out under official OIML R 76-1:2006 requirements, the Non-Automatic Weighing Instrument WeighMaster Bench Scale (SN: SN-2026-WM-98101) DOES NOT COMPLY with metrological requirements.

Tested By (Technician):

Priya Verma (Metrology Engineer)

Date: 2026-09-19

Approved By (Technical Reviewer):

Arun K. Patel (Senior Verification Officer)

Date: 2026-09-19

Signature / Verification Stamp

Official Metrology Seal