

NATIONAL LEGAL METROLOGY EVALUATION CENTRE

Accreditation No: NABL-MET-2026-0982 | Metrology Complex, CSIR Road, New Delhi 110012, India

TYPE EVALUATION TEST REPORT — OIML R 76

Report Number: LM/2026/00001 | Test ID: TEST-2026-0001 | Generated: 2026-09-30

Manufacturer Avery Weigh-Tronix India Ltd.	Model Name WeighMaster Bench Scale	Serial Number SN-2026-WM-98101	Accuracy Class Class III (Medium)
Max Capacity (Max) 30.000000 kg	Min Capacity (Min) 0.100000 kg	Verification Scale (e) 0.010000 kg	Actual Scale (d) 0.010000 kg

1. Environmental Test Conditions (Clause 3.9)

Ambient Temperature 21.50 °C	Relative Humidity 45.00 %	Atmospheric Pressure 1012.50 hPa	Standards Used E2 Mass Set
--	-------------------------------------	--	--------------------------------------

2. Weighing Performance Test (Clause A.4.4) — Status: COMPLETED

Pt #	Direction	Applied (kg)	Indicated (kg)	Error (kg)	MPE (kg)	Verdict
1	INCREASING	0	0	0	±0.005	PASS
2	INCREASING	5	5.002	0.002	±0.005	PASS
3	INCREASING	10	10.003	0.003	±0.01	PASS
4	INCREASING	20	20.004	0.004	±0.01	PASS
5	INCREASING	30	30.005	0.005	±0.015	PASS
6	DECREASING	20	20.004	0.004	±0.01	PASS
7	DECREASING	0	0	0	±0.005	PASS

3. Repeatability Test (Clause A.4.10) — Status: PENDING

4. Eccentric Loading Test (Clause A.4.7) — Status: PENDING

Evidence & Supporting Attachments

Attachment Category	File & Description	Visual Evidence / Details
INSTRUMENT PHOTO Size: 0 KB	test_instrument_photo.png Front view of weighing terminal showing model badge and load receiver Uploaded by: Arun K. Patel (Senior Verification Officer)	

FINAL METROLOGICAL VERDICT: INCOMPLETE

Based on the results of the evaluation carried out under official OIML R 76-1:2006 requirements, the Non-Automatic Weighing Instrument WeighMaster Bench Scale (SN: SN-2026-WM-98101) DOES NOT FULLY COMPLY with metrological tolerances or is incomplete.

OFFICIAL LEGAL METROLOGY DIGITAL SIGNATURE & VERIFICATION SEAL

"I hereby certify under official Legal Metrology authority that this evaluation has been verified in strict compliance with OIML R 76-1:2006."

Digitally Signed By Arun K. Patel (Senior Verification Officer) (Chief Metrology Evaluation Officer)	Certificate Reference SIG-LM-2026-A2F295B1 Serial: CERT-OIML-2026-IN-0001	Signed Timestamp 2026-09-30 15:25:22.514864 IST Algorithm: SHA256withRSA / OIML Metrology Digital Seal
---	--	--

Cryptographic Seal (SHA-256 Digest): 84b6b98e7ba0d7c9104ab73a758db6aef2f3154292c83e583f0f0dd80eb1e4cd