

NATIONAL LEGAL METROLOGY EVALUATION CENTRE

Accreditation No: NABL-MET-2026-0982 | Metrology Complex, CSIR Road, New Delhi 110012, India

TYPE EVALUATION TEST REPORT — OIML R 76

Report No: LM/2026/00003 | Evaluation ID: TEST-2026-0002 | Standard: OIML R 76-1:2006

1. INSTRUMENT UNDER EVALUATION

Manufacturer: Avery Weigh-Tronix India Ltd.	Model Name: WeighMaster Bench Scale	Serial Number: SN-2026-WM-98101	Instrument Type: Electronic Platform Scale
Accuracy Class: Class III (Medium)	Max Capacity (Max): 30 kg	Min Capacity (Min): 0.1 kg	Scale Interval (e): 0.01 kg

2. LABORATORY & ENVIRONMENTAL CONDITIONS

Ambient Temperature: 22.50 °C	Relative Humidity: 48.00 %	Atmospheric Pressure: 1013.20 hPa	Standards Used: E2 Class Stainless Steel Mass Set (1 mg to 20 kg)
---	--------------------------------------	---	---

3. METROLOGICAL TEST RESULTS (OIML R 76-1:2006)

• Weighing Performance Test (Clause A.4.4) [Result: PASS]

#	Applied Load	Indication	Changeover (L)	Error (E)	MPE Limit	Status
1	0 kg	0 kg	-	0 kg	±0.005 kg	PASS
2	5 kg	5 kg	-	0 kg	±0.005 kg	PASS
3	10 kg	10.005 kg	0.005	+0.005 kg	±0.01 kg	PASS
4	20 kg	20 kg	-	0 kg	±0.01 kg	PASS
5	30 kg	30 kg	-	0 kg	±0.015 kg	PASS
6	20 kg	20 kg	-	0 kg	±0.01 kg	PASS
7	0 kg	0 kg	-	0 kg	±0.005 kg	PASS

• Repeatability Test (Clause A.4.10) [Result: PASS]

#	Applied Load	Indication	Changeover (L)	Error (E)	MPE Limit	Status
1	15 kg	15 kg	-	0 kg	±0.01 kg	PASS
2	15 kg	15.002 kg	-	+0.002 kg	±0.01 kg	PASS
3	15 kg	15 kg	-	0 kg	±0.01 kg	PASS

Notes: Repeatability spread Delta I = 0.002 kg. MPE limit = 0.01 kg. Result: PASS

• Eccentric Loading Test (Clause A.4.7) [Result: PASS]

#	Applied Load	Indication	Changeover (L)	Error (E)	MPE Limit	Status
1	9.999 kg	9.999 kg	-	0 kg	±0.01 kg	PASS
2	9.999 kg	9.999 kg	-	0 kg	±0.01 kg	PASS
3	9.999 kg	9.999 kg	-	0 kg	±0.01 kg	PASS
4	9.999 kg	9.999 kg	-	0 kg	±0.01 kg	PASS
5	9.999 kg	9.999 kg	-	0 kg	±0.01 kg	PASS

FINAL METROLOGICAL VERDICT: PASS

Based on the results of the evaluation carried out under official OIML R 76-1:2006 requirements, the Non-Automatic Weighing Instrument WeighMaster Bench Scale (SN: SN-2026-WM-98101) SATISFACTORILY COMPLIES with all applicable metrological tolerances.

Tested By (Technician):

Priya Verma (Metrology Engineer)

Date: 2026-09-19

Signature / Verification Stamp

Approved By (Technical Reviewer):

Arun K. Patel (Senior Verification Officer)

Date: 2026-09-19

Official Metrology Seal