

NATIONAL LEGAL METROLOGY EVALUATION CENTRE

Accreditation No: NABL-MET-2026-0982 | Metrology Complex, CSIR Road, New Delhi 110012, India

TYPE EVALUATION TEST REPORT — OIML R 76

Report No: LM/2026/00007 | Evaluation ID: TEST-2026-0008 | Standard: OIML R 76-1:2006_E

1. INSTRUMENT UNDER EVALUATION

Manufacturer: Mettler Toledo Metrology Systems	Model Name: Analytical Precision Balance	Serial Number: SN-2026-MT-44019	Instrument Type: Precision Analytical Balance
Accuracy Class: Class I (Special)	Max Capacity (Max): 220 g	Min Capacity (Min): 0.01 g	Scale Interval (e): 0.001 g

2. LABORATORY & ENVIRONMENTAL CONDITIONS

Ambient Temperature: 22.50 °C	Relative Humidity: 48.00 %	Atmospheric Pressure: 1013.20 hPa	Standards Used: E2 Class Stainless Steel Mass Set (1 mg to 20 kg)
---	--------------------------------------	---	---

3. METROLOGICAL TEST RESULTS (OIML R 76-1:2006)

• Weighing Performance Test (Clause A.4.4) [Result: PASS]

#	Applied Load	Indication	Changeover (L)	Error (E)	MPE Limit	Status
1	0 g	0 g	-	0 g	±0.0005 g	PASS
2	0.5 g	0.5 g	-	0 g	±0.0005 g	PASS
3	1 g	1.0005 g	0.0005 g	+0.0005 g	±0.0005 g	PASS
4	2 g	2 g	-	0 g	±0.0005 g	PASS
5	220 g	220 g	-	0 g	±0.0015 g	PASS
6	2 g	2 g	-	0 g	±0.0005 g	PASS
7	0 g	0 g	-	0 g	±0.0005 g	PASS

• Repeatability Test (Clause A.4.10) [Result: PENDING]

#	Applied Load	Indication	Changeover (L)	Error (E)	MPE Limit	Status
1	110 g	110 g	-	-	-	-
2	110 g	110.0002 g	-	-	-	-
3	110 g	110 g	-	-	-	-

• Eccentric Loading Test (Clause A.4.7) [Result: PENDING]

FINAL METROLOGICAL VERDICT: INCOMPLETE

Based on the results of the evaluation carried out under official OIML R 76-1:2006 requirements, the Non-Automatic Weighing Instrument Analytical Precision Balance (SN: SN-2026-MT-44019) DOES NOT FULLY COMPLY with metrological tolerances or is incomplete.

Tested By (Technician):

Priya Verma (Metrology Engineer)

Date: 2026-09-26

Approved By (Technical Reviewer):

Technical Reviewer

Date: -

Signature / Verification Stamp

Official Metrology Seal