

NATIONAL LEGAL METROLOGY EVALUATION CENTRE

Accreditation No: NABL-MET-2026-0982 | Metrology Complex, CSIR Road, New Delhi 110012, India

TYPE EVALUATION TEST REPORT — OIML R 76

Report No: LM/2026/00010 | Evaluation ID: TEST-2026-0012 | Standard: OIML R 76-1:2006_E

1. INSTRUMENT UNDER EVALUATION

Manufacturer: Mettler Toledo Metrology Systems	Model Name: Analytical Precision Balance	Serial Number: SN-2026-MT-44019	Instrument Type: Precision Analytical Balance
Accuracy Class: Class I (Special)	Max Capacity (Max): 220 g	Min Capacity (Min): 0.01 g	Scale Interval (e): 0.001 g

2. LABORATORY & ENVIRONMENTAL CONDITIONS

Ambient Temperature: 20.00 °C	Relative Humidity: 50.00 %	Atmospheric Pressure: 1013.25 hPa	Standards Used: OIML Standard Weights
---	--------------------------------------	---	---

3. METROLOGICAL TEST RESULTS (OIML R 76-1:2006)

• Weighing Performance Test (Clause A.4.4) [Result: PENDING]

• Repeatability Test (Clause A.4.10) [Result: PASS]

#	Applied Load	Indication	Changeover (L)	Error (E)	MPE Limit	Status
1	25 g	25 g	-	0 g	±0.0005 g	PASS
2	25 g	25.0002 g	-	+0.0002 g	±0.0005 g	PASS
3	25 g	25 g	-	0 g	±0.0005 g	PASS

Notes: Repeatability spread Delta I = 0.0002 g. MPE limit = 0.0005 g. Result: PASS

• Eccentric Loading Test (Clause A.4.7) [Result: PENDING]

FINAL METROLOGICAL VERDICT: INCOMPLETE

Based on the results of the evaluation carried out under official OIML R 76-1:2006 requirements, the Non-Automatic Weighing Instrument Analytical Precision Balance (SN: SN-2026-MT-44019) DOES NOT FULLY COMPLY with metrological tolerances or is incomplete.

Tested By (Technician):

Priya Verma (Metrology Engineer)

Date: 2026-09-30

Approved By (Technical Reviewer):

Arun K. Patel (Senior Verification Officer)

Date: 2026-09-30

Signature / Verification Stamp

Official Metrology Seal