

NATIONAL LEGAL METROLOGY EVALUATION CENTRE

Accreditation No: NABL-MET-2026-0982 | Metrology Complex, CSIR Road, New Delhi 110012, India

TYPE EVALUATION TEST REPORT — OIML R 76

Report Number: LM/2026/00015 | Test ID: TEST-2026-0017 | Generated: 2026-09-30

Manufacturer Avery Weigh-Tronix India Ltd.	Model Name WeighMaster Bench Scale	Serial Number SN-2026-WM-98101	Accuracy Class Class III (Medium)
Max Capacity (Max) 30.000000 kg	Min Capacity (Min) 0.100000 kg	Verification Scale (e) 0.010000 kg	Actual Scale (d) 0.010000 kg

1. Environmental Test Conditions (Clause 3.9)

Ambient Temperature 22.00 °C	Relative Humidity 48.00 %	Atmospheric Pressure 1013.25 hPa	Standards Used E2 Stainless Steel Mass Standards (1 mg - 20 kg)
--	-------------------------------------	--	--

2. Weighing Performance Test (Clause A.4.4) — Status: COMPLETED

Pt #	Direction	Applied (kg)	Indicated (kg)	Error (kg)	MPE (kg)	Verdict
1	INCREASING	0	0	0	±0.005	PASS
2	INCREASING	500	500	-500	±0.005	FAIL
3	INCREASING	1000	1000	-1000	±0.005	FAIL
4	INCREASING	2000	2000	-2000	±0.005	FAIL
5	INCREASING	3000	3000	-3000	±0.005	FAIL

3. Repeatability Test (Clause A.4.10) — Status: PENDING

4. Eccentric Loading Test (Clause A.4.7) — Status: PENDING

Evidence & Supporting Attachments

Attachment Category	File & Description	Visual Evidence / Details
INSTRUMENT PHOTO Size: 0 KB	nawi_test_photo.png Front view of electronic counter scale terminal and load receiver Uploaded by: Dr. Rajesh Sharma (Lab Director)	
NAMEPLATE PHOTO Size: 0 KB	nawi_test_photo.png Metrological rating plate with Class III mark, Max 15kg, and Min 100g Uploaded by: Dr. Rajesh Sharma (Lab Director)	

FINAL METROLOGICAL VERDICT: FAIL

Based on the results of the evaluation carried out under official OIML R 76-1:2006 requirements, the Non-Automatic Weighing Instrument WeighMaster Bench Scale (SN: SN-2026-WM-98101) DOES NOT FULLY COMPLY with metrological tolerances or is incomplete.

OFFICIAL LEGAL METROLOGY DIGITAL SIGNATURE & VERIFICATION SEAL

"I hereby certify under statutory Legal Metrology authority that this evaluation has been verified in strict compliance with OIML R 76-1:2006."

Digitally Signed By Dr. Deepali Sharma (Chief Verification Reviewer) (Chief Metrological Verification Reviewer)	Certificate Reference SIG-LM-2026-989D0F9B Serial: CERT-OIML-2026-IN-0017	Signed Timestamp 2026-09-30 15:57:37.120642 IST Algorithm: SHA256withRSA / OIML Metrology Digital Seal
--	--	---

Cryptographic Seal (SHA-256 Digest): c6efe17f97b9118376a695f70d5d8005c4e563824676203310ff5b648971b212