

NanoSafe Analyzer — Multi-Sample Comparison Report

Comparative Evaluation of ZnO Nanoparticle Cytotoxicity & ISO 10993-5 Biocompatibility


Report Type: Multi-Sample Comparison | Generated: 2026-08-17 05:14 UTC | Samples Compared: 3

1. COMPARATIVE CYTOTOXICITY SUMMARY


Sample / Experiment	Cell Line	Viability	Tox Score	IC50	ISO Verdict
Experiment test 2	HeLa	68.9%	12.1/100	56.2 µg/mL	FAIL — Cytotoxic
Experiment test 2	HeLa	68.9%	12.1/100	56.2 µg/mL	FAIL — Cytotoxic
Experiment test 1	HeLa	71.7%	11.5/100	58.01 µg/mL	FAIL — Cytotoxic

2. MULTI-BIOMARKER PROFILES


Experiment test 2



Experiment test 2



Experiment test 1



3. CLINICAL SUGGESTIONS PER SAMPLE

A. Experiment test 2

CAUTION — MODERATE TOXICITY: Cell viability at 68.9% is below ISO 10993-5 PASS threshold. Formulation requires optimization before biomedical use.

B. Experiment test 2

CAUTION — MODERATE TOXICITY: Cell viability at 68.9% is below ISO 10993-5 PASS threshold. Formulation requires optimization before biomedical use.

C. Experiment test 1

CAUTION — MODERATE TOXICITY: Cell viability at 71.7% is below ISO 10993-5 PASS threshold. Formulation requires optimization before biomedical use.