

# Digital Calibration Certificate

## Administrative Data

### DCC Software

- Software

Name	GEMIMEG tool
Release	v1.3.0

### Core Data

Country Code ISO3166_1	GB
Used Language Code ISO639_1	en
Mandatory Language Code ISO639_1	en
Unique Identifier	DCC2GO-Temperature-DCC-Example
Issuer	calibrationLaboratory
Order no.	2023/04/0001
Performance Date	2023-04-13 to 2023-04-14
Performance Location	laboratory

### Items

• Item	Temperature sensor
Manufacturer	Sensor-Manufacturer
Model	PT100
Issuer	manufacturer
Serial no.	2309823482

### Calibration Laboratory

	EMPIR-DCC2GO
E-Mail	dcc2go@ptb.de
City	Capital City
Post Code	12345
Street	High Street
Street No.	1
Further	<a href="https://www.ptb.de/dcc2go/project">https://www.ptb.de/dcc2go/project</a>

### Person(s) authorizing the report

Name	Jane Bloggs
Main signer	true

### Customer

	Example Company
E-Mail	info@customer.xx
City	Capital City
Post Code	789101

## **Statements**

- Statement

This certificate is consistent with the Calibration and Measurement Capabilities (CMCs) that are included in Appendix C of the Mutual Recognition Arrangement (MRA) drawn up by the International Committee for Weights and Measures (CIPM). Under the MRA, all participating institutes recognize the validity of each other's calibration and measurement certificates for the quantities, ranges and measurement uncertainties specified in Appendix C (for details, see <http://www.bipm.org>).

## **Measurement Results**

### **Measurement results**

#### **Used Methods**

- Used method

Calibration of temperature sensors

#### **Measuring Equipments**

- Measuring Equipment

Pt 100 thermometer

Issuer (Identification): manufacturer

Value (Identification): reference-PT100-002

## **Results**

## Measuring results

Reference value \degreecelsius
33.098
99.971
175.103
250.169
320.004

Indicated measured value probe \degreecelsius
33.17
100.06
175.21
250.16
319.92

Measurement error \kelvin
0.072
0.089
0.107
-0.009
-0.084

Uncertainty List \kelvin	Coverage Factor List	Coverage Probability List	Distribution List
0.061	2	0.95	normal