THE LNM INSTITUTE OF INFORMATION TECHNOLOGY

Central Laboratory Inventory & Procurement Section

SIGNED DEFECT & CONDEMNATION INSPECTION REPORT

Inspection Memo for Damaged / Unserviceable Lab Stock

Memo Reference:	MEMO-CONDEMN-2024-019
Date of Inspection:	05-September-2024
Part Number:	IC-NE555P / DSO-SIG-1102
Item Name:	NE555 Timer IC & Oscilloscope Front Channel
Quantity Condemned:	5 pcs ICs + 2 units DSO (Kharab hoga pr rkha rahega)
Defect Reason:	Over-voltage surge during power stage experiment; retained for scrap audit
Storage Location:	Condemned Equipment Rack B-4, Electronics Lab
Inspecting Staff:	Mukesh Sharma (Lab Administrator) [Signed]
Faculty Incharge:	Prof. R. Gangopadhyay (ECE Dept Head) [Signed]

Verified & Signed By

Lab Administrator / Technician

Approved & Counter-Signed

Head of Department / Faculty Incharge