

Date & Size of Barrage	Test taken with/without hose	RHS		LHS	
		Direct	Indirect	Direct	Indirect
01-01-2021 (500mm)	Barrage without Hose	OL	OL	OL	OL
	Barrage with Hose with water	2900 Ω	900 Ω	5500 Ω	OL
16-09-2021 (500 mm)	Barrage without Hose	OL	OL	OL	OL
	Barrage with Hose with water	3620 Ω	3270 Ω	4200 Ω	OL
13-05-2022 (500 mm)	Barrage without Hose	OL	OL	OL	OL
	Barrage with Hose with water	2022 Ω	1566 Ω	OL	3345 Ω
06-02-2023 (500 mm)	Barrage without Hose	OL	OL	OL	OL
	Barrage with Hose with water	492 Ω	4910 Ω	4063 Ω	8890 Ω
19-06-2023 (500 mm)	Barrage without Hose	OL	OL	OL	OL
	Barrage with Hose with water	1580 Ω	5071 Ω	OL	OL
27-07-2023 (500 mm)	Barrage without Hose	OL	OL	OL	OL
	Barrage with Hose with water	1500 Ω	4791 Ω	OL	OL
6-08-2023 (500 mm)	Barrage without Hose	OL	OL	OL	OL
	Barrage with Hose with water	OL	OL	OL	OL
26-03-2024	Barrage without Hose	OL	OL	OL	OL
	Barrage with Hose with water	OL	OL	OL	OL
30-05-2024 (500 mm/480mm)	Barrage inside the glass	666 Ω		616 Ω	
31-05-2024 500mm	Barrage with Hose with water	OL	OL	OL	OL
1/6/2024 (500 mm/480mm)	Barrage inside the glass	138 Ω		106 Ω	