X-ray Acquisition Statistics Report

View counts

view_code	image_count	percent
AP	147173	39.03
PA	96161	25.5
LATERAL	82853	21.97
LL	35133	9.32
missing	15769	4.18
PA LLD	4	0.0
LAO	3	0.0
RAO	3	0.0
AP AXIAL	2	0.0
AP LLD	2	0.0
AP RLD	2	0.0
XTABLE LATERAL	2	0.0
LPO	1	0.0
PA RLD	1	0.0
SWIMMERS	1	0.0

Detector Types

detector_type_code	image_count
DIRECT	294189
SCINTILLATOR	66692
missing	16229

KVP by detector type - all views

kvp_group	DIRECT	SCINTILLATOR	missing
<90	8749	37	80
=90	76794	4	335
>90 and <110	21258	22	35
=110	65153	22	2

kvp_group	DIRECT	SCINTILLATOR	missing
>110 and <120	369	14	0
=120	82551	66210	0
>120	990	381	8
missing	38325	2	15769

KVP by detector type - AP view

kvp_group	DIRECT	SCINTILLATOR	missing
<90	8654	14	80
=90	76724	3	333
>90 and <110	21159	17	33
=110	1409	0	2
>110 and <120	196	0	0
=120	167	232	0
>120	91	1	0
missing	38058	0	0

KVP by detector type - PA view

kvp_group	DIRECT	SCINTILLATOR	missing
<90	88	18	0
=90	62	0	1
>90 and <110	68	1	2
=110	63539	8	0
>110 and <120	155	7	0
=120	624	31099	0
>120	183	161	4
missing	139	2	0

KVP by detector type - Lateral view

kvp_group	DIRECT	SCINTILLATOR	missing
<90	5	0	0
=90	5	0	0
>90 and <110	29	0	0
=110	197	0	0
>110 and <120	18	0	0
=120	81758	0	0
>120	716	0	0
missing	125	0	0

KVP by detector type - LL view

kvp_group	DIRECT	SCINTILLATOR	missing
<90	0	5	0
=90	0	1	0
>90 and <110	0	4	0
=110	0	14	0
>110 and <120	0	7	0
=120	0	34879	0
>120	0	219	4
missing	0	0	0

KVP by detector type - Missing view

kvp_group	DIRECT	SCINTILLATOR	missing
<90	0	0	0
=90	0	0	0
>90 and <110	0	0	0
=110	0	0	0
>110 and <120	0	0	0
=120	0	0	0
>120	0	0	0
missing	0	0	15769

mAs Standard Deviation by Population

population	mean_exposure_mas	std_exposure_mas
DIRECT, kVp=90	3.41	0.87
DIRECT, kVp=110	3.33	4.26
SCINTILLATOR, kVp=120	9.05	14.22