

X-ray Acquisition Statistics Report

View counts

view_code	image_count	percent
AP	31443	39.11
PA	20452	25.44
LATERAL	17578	21.87
LL	7507	9.34
missing	3409	4.24

Detector Types

detector_type_code	image_count
DIRECT	62611
SCINTILLATOR	14254
missing	3524

KVP by detector type - all views

kvp_group	DIRECT	SCINTILLATOR	missing
<90	1843	5	25
=90	16241	0	83
>90 and <110	4659	8	4
=110	13807	4	1
>110 and <120	73	2	0
=120	17529	14132	0
>120	205	103	2
missing	8254	0	3409

KVP by detector type - AP view

kvp_group	DIRECT	SCINTILLATOR	missing
<90	1829	0	25
=90	16227	0	82
>90 and <110	4638	6	4
=110	295	0	1
>110 and <120	34	0	0
=120	42	49	0
>120	31	1	0
missing	8179	0	0

KVP by detector type - Lateral view

kvp_group	DIRECT	SCINTILLATOR	missing
<90	0	0	0
=90	0	0	0
>90 and <110	4	0	0
=110	37	0	0
>110 and <120	4	0	0
=120	17355	0	0
>120	141	0	0
missing	37	0	0

KVP by detector type - PA view

kvp_group	DIRECT	SCINTILLATOR	missing
<90	14	5	0
=90	14	0	1
>90 and <110	17	0	0
=110	13475	1	0
>110 and <120	35	1	0
=120	132	6636	0
>120	33	49	1
missing	38	0	0

KVP by detector type - LL view

kvp_group	DIRECT	SCINTILLATOR	missing
<90	0	0	0
=90	0	0	0
>90 and <110	0	2	0
=110	0	3	0
>110 and <120	0	1	0
=120	0	7447	0
>120	0	53	1
missing	0	0	0

KVP by detector type - Missing view

kvp_group	DIRECT	SCINTILLATOR	missing
<90	0	0	0
=90	0	0	0
>90 and <110	0	0	0
=110	0	0	0
>110 and <120	0	0	0
=120	0	0	0
>120	0	0	0
missing	0	0	3409

mAs Standard Deviation by Population

population	mean_exposure_mas	std_exposure_mas
DIRECT, kVp=90	3.42	0.95
DIRECT, kVp=110	3.32	3.66
SCINTILLATOR, kVp=120	3.32	3.66