
Calibration is based on the poured in volume ("In") at a ± 20 °C reference temperature. The volume content tolerances conform to accuracy class A, the accuracy limits of the German weights and measures regulations and DIN and ISO specifications.

Typical applications: precise measurement of specified liquid amounts, preparation and storage of standard solutions.

Cat. No.	Capacity (mL)			Neck	Stopper size	Accu- racy limits (mL)	Remark	Pack Unit
246770955	5 W	22	70	9 ± I	10/19	0.04	wide neck	2
246771051	10 W	27	90	9 ± I	10/19	0.04	wide neck	2
246771257	20	39	110	9 ± 1	10/19	0.04		2
246771454	25	40	110	9 ± 1	10/19	0.04		2
246771754	50	50	140	±	12/21	0.06		2
246772459	100	60	170	13 ± 1	12/21	0.1		2
246772553	100	60	170	13 ± 1	14/23	0.1		2
246773258	200	75	210	15.5 ± 1.5	14/23	0.15		2
246773652	250	80	220	15.5 ± 1.5	14/23	0.15		2
246774451	500	100	260	19 ± 2	19/26	0.25		2
246775456	1 000	125	300	23 ± 2	24/29	0.4		2
246776358	2 000	160	370	27.5 ± 2.5	29/32	0.6		2

DURAN® Volumetric Flask, class A, amber, individual certificate

with one graduation mark and ergonomic polyethylene stopper, white printed image, with individual certificate and certificate of conformity









Calibration is based on the poured in volume ("ln") at a ± 20 °C reference temperature. The volume content tolerances conform to accuracy class A, the accuracy limits of the German weights and measures regulations and DIN and ISO specifications.

Typical applications: precise measurement of specified liquid amounts, preparation and storage of standard solutions.

Cat. No.	Capacity (mL)	d (mm)	h (mm)	Neck	Stopper size	Accu- racy limits (mL)	Remark	Pack Unit
246780159	- 1	13	65	7 ± I	7/16	25		2
246780253	2	17	70	7 ± I	7/16	25		2
246780956	5 W	22	70	9 ± I	10/19	0.04	wide neck	2
246781052	10 W	27	90	9 ± I	10/19	0.04	wide neck	2
246781258	20	39	110	9 ± 1	10/19	0.04		2
246781455	25	40	110	9 ± 1	10/19	0.04		2
246781755	50	50	140	±	12/21	0.06		2
246782451	100	60	170	13 ± 1	12/21	0.1		2
246782554	100	60	170	13 ± 1	14/23	0.1		2
246783259	200	75	210	15.5 ± 1.5	14/23	0.15		2
246783653	250	80	220	15.5 ± 1.5	14/23	0.15		2
246784452	500	100	260	19 ± 2	19/26	0.25		2
246785457	1 000	125	300	23 ± 2	24/29	0.4		2
246785551	I 000 W	125	300	27.5 ± 2.5	29/32	0.6	wide neck	2
246786359	2 000	160	370	27.5 ± 2.5	29/32	0.6		2
246787355	5 000	215	475	38 ± 3	34/35	1.2		- 1

DURAN® Volumetric Flask, class A, batch certificate

with one graduation mark and ergonomic polyethylene stopper, blue printed image, with batch certificate and certificate of conformity

