

0.1 Bland-Altman Approach

The issue of whether two measurement methods are comparable to the extent that they can be used interchangeably with sufficient accuracy is encountered frequently in scientific research. Historically comparison of two methods of measurement was carried out by use of matched pairs correlation coefficients or simple bivariate methods. Bland and Altman recognized the inadequacies of these analyses and articulated quite thoroughly the basis on which of which they are unsuitable for comparing two methods of measurement (Altman and Bland, 1983).

As an alternative they proposed a simple statistical methodology specifically appropriate for method comparison studies. They acknowledge that there are other valid methodologies, but argue that a simple approach is preferable to complex approaches, *"especially when the results must be explained to non-statisticians"* (Altman and Bland, 1983).

The scientific question at hand is the correct approach to assessing whether two methods can be used interchangeably. Bland and Altman (1999) expresses this as follows:

We want to know by how much (one) method is likely to differ from the (other), so that if it not enough to cause problems in the mathematical interpretation we can ... use the two interchangeably.

Statisticians Martin Bland and Douglas Altman recognized the inadequacies of these analyzes and articulated quite thoroughly the basis on which of which they are unsuitable for comparing two methods of measurement (Altman and Bland, 1983). Furthermore they proposed their simple methodology specifically constructed for method comparison studies. They acknowledge the opportunity to apply other valid, but complex, methodologies, but argue that a simple approach is preferable, especially when

the results must be ‘explained to non-statisticians’.

In 1986 Bland and Altman published a paper in the Lancet proposing the difference plot for use for method comparison purposes. It has proved highly popular ever since. This is a simple, and widely used, plot of the differences of each data pair, and the corresponding average value. An important requirement is that the two measurement methods use the same scale of measurement.

Bland Altman have recommended the use of graphical techniques to assess agreement. Principally their method is calculating, for each pair of corresponding two methods of measurement of some underlying quantity, with no replicate measurements, the difference and mean. Differences are then plotted against the mean.

Hopkins argued that the bias in a subsequent Bland-Altman plot was due, in part, to using least-squares regression at the calibration phase.

0.1.1 Identity Plot

The first step recommended which the authors argue should be mandatory is construction of a simple scatter plot of the data. The line of equality ($X = Y$) should also be shown, as it is necessary to give the correct interpretation of how both methods compare.

Notwithstanding previous remarks about linear regression, the first step recommended, which the authors argue should be mandatory, is construction of a simple scatter plot of the data. The line of equality ($X = Y$) should also be shown, as it is necessary to give the correct interpretation of how both methods compare. In the case of good agreement, the observations would be distributed closely along the line of equality. A scatter plot of the Grubbs data is shown in Figure 1.1. A visual inspection thereof confirms the previous conclusion that there is an inter method bias present, i.e. Fotobalk device has a tendency to record a lower velocity.




Figure 1: Scatter plot For Fotobalk and Counter Methods.

The authors advise the use of scatter plots to identify outliers, and to determine if there is curvilinearity present. In the region of linearity, simple linear regression may yield results of interest.

Dewitte et al. (2002) notes that scatter plots were very seldom presented in the *Annals of Clinical Biochemistry*. This apparently results from the fact that the 'Instructions for Authors' dissuade the use of regression analysis, which conventionally is accompanied by a scatter plot.

0.2 Bland-Altman Plots

In light of shortcomings associated with scatterplots, Altman and Bland (1983) recommend a further analysis of the data. Firstly case-wise differences of measurements of two methods $d_i = y_{1i} - y_{2i}$ for $i = 1, 2, \dots, n$ on the same subject should be calculated, and then the average of those measurements ($a_i = (y_{1i} + y_{2i})/2$ for $i = 1, 2, \dots, n$). These differences and averages are then plotted.

Altman and Bland (1983) proposes a scatterplot of the case-wise averages and differences of two methods of measurement. This scatterplot has since become widely known as the Bland-Altman plot. Altman and Bland (1983) express the motivation for this plot thusly:

“From this type of plot it is much easier to assess the magnitude of disagreement (both error and bias), spot outliers, and see whether there is any trend, for example an increase in (difference) for high values. This way of plotting the data is a very powerful way of displaying the results of a method comparison study.”

The case wise-averages capture several aspects of the data, such as expressing the range over which the values were taken, and assessing whether the assumptions of constant variance holds. Case-wise averages also allow the case-wise differences to be presented on a two-dimensional plot, with better data visualization qualities than a one dimensional plot.

Bland and Altman (1986) cautions that it would be the difference against either measurement value instead of their average, as the difference relates to both value. This methodology has proved very popular, and the Bland-Altman plots is widely regarded as powerful graphical methodology for making a visual assessment of the data.

The magnitude of the inter-method bias between the two methods is simply the average of the differences \bar{d} . This inter-method bias is represented with a line on the Bland-Altman plot. As the objective of the Bland-Altman plot is to advise on the agreement of two methods, it is the case-wise differences that are also particularly relevant. The variances around this bias is estimated by the standard deviation of these differences S_d .

The magnitude of the inter-method bias between the two methods is simply the average of the differences \bar{d} . The variances around this bias is estimated by the standard deviation of the differences $S(d)$.

This inter-method bias is represented with a line on the Bland-Altman plot. These estimates are only meaningful if there is uniform inter-bias and variability throughout the range of measurements, which can be checked by visual inspection of the plot.

The magnitude of the inter-method bias between the two methods is simply the average of the differences \bar{d} . The variances around this bias is estimated by the standard deviation of the differences $S(d)$. This inter-method bias is represented with a line on the Bland-Altman plot.

0.2.1 Bland-Altman plots for the Grubbs data

These estimates are only meaningful if there is uniform inter-bias and variability throughout the range of measurements, which can be checked by visual inspection of the plot. In the case of Grubbs data the inter-method bias is -0.61 metres per second, and is indicated by the dashed line on Figure 1.2. By inspection of the plot, it is also possible to compare the precision of each method. Noticeably the differences tend to increase as the averages increase.

The Bland-Altman plot for comparing the ‘Fotobalk’ and ‘Counter’ methods, which shall henceforth be referred to as the ‘F vs C’ comparison, is depicted in Figure 1.2,

using data from Table 1.3. The presence and magnitude of the inter-method bias is indicated by the dashed line.

Round	Fotobalk [F]	Counter [C]	Differences [F-C]	Averages [(F+C)/2]
1	793.8	794.6	-0.8	794.2
2	793.1	793.9	-0.8	793.5
3	792.4	793.2	-0.8	792.8
4	794.0	794.0	0.0	794.0
5	791.4	792.2	-0.8	791.8
6	792.4	793.1	-0.7	792.8
7	791.7	792.4	-0.7	792.0
8	792.3	792.8	-0.5	792.5
9	789.6	790.2	-0.6	789.9
10	794.4	795.0	-0.6	794.7
11	790.9	791.6	-0.7	791.2
12	793.5	793.8	-0.3	793.6

Table 1: Fotobalk and Counter methods: differences and averages.

Round	Fotobalk [F]	Terma [T]	Differences [F-T]	Averages [(F+T)/2]
1	793.8	793.2	0.6	793.5
2	793.1	793.3	-0.2	793.2
3	792.4	792.6	-0.2	792.5
4	794.0	793.8	0.2	793.9
5	791.4	791.6	-0.2	791.5
6	792.4	791.6	0.8	792.0
7	791.7	791.6	0.1	791.6
8	792.3	792.4	-0.1	792.3
9	789.6	788.5	1.1	789.0
10	794.4	794.7	-0.3	794.5
11	790.9	791.3	-0.4	791.1
12	793.5	793.5	0.0	793.5

Table 2: Fotobalk and Terma methods: differences and averages.




Figure 2: Bland-Altman plot For Fotobalk and Counter methods.

In Figure 1.3 Bland-Altman plots for the ‘F vs C’ and ‘F vs T’ comparisons are shown, where ‘F vs T’ refers to the comparison of the ‘Fotobalk’ and ‘Terma’ methods. Usage of the Bland-Altman plot can be demonstrate in the contrast between these comparisons. By inspection, there exists a larger inter-method bias in the ‘F vs C’ comparison than in the ‘F vs T’ comparison. Conversely there appears to be less precision in ‘F vs T’ comparison, as indicated by the greater dispersion of covariates.

In Figure 1.3 Bland-Altman plots for the ‘F vs C’ and ‘F vs T’ comparisons are shown, where ‘F vs T’ refers to the comparison of the ‘Fotobalk’ and ‘Terma’ methods. Usage of the Bland-Altman plot can be demonstrate in the contrast between these comparisons.




Figure 3: Bland-Altman plots for Grubbs' F vs C and F vs T comparisons.

By inspection, there exists a larger inter-method bias in the 'F vs C' comparison than in the 'F vs T' comparison. Conversely there appears to be less precision in 'F vs T' comparison, as indicated by the greater dispersion of co-variates.

0.2.2 Bias Detection

Further to this method, the presence of constant bias may be indicated if the average value differences is not equal to zero. Bland and Altman does, however, indicate the indication of absence of bias does not provide sufficient information to allow a judgement as to whether or not one method can be substituted for another.

The dashed line in Figure 2.2 alludes to the inter method bias between the two methods, as mentioned previously. Bland and Altman recommend the estimation of inter method bias by calculating the average of the differences. In the case of Grubbs data the inter method bias is -0.6083 metres per second.




Figure 4: Bland Altman Plot For Fotobalk and Counter Methods

By inspection of the plot, it is also possible to compare the precision of each method. Noticeably the differences tend to increase as the averages increase.

0.2.3 Simulated Data

Figures 1.3 1.4 and 1.5 are three Bland-Altman plots derived from simulated data, each for the purpose of demonstrating how the plot would inform an analyst of trends that would adversely affect use of the recommended methodology. Figure 1.3 demonstrates how the Bland Altman plot would indicate increasing variance of differences over the measurement range. Figure 1.4 is an example of cases where the inter-method bias changes over the measurement range. This is known as proportional bias (Ludbrook, 1997).

Figure 1.4 demonstrates how the Bland-Altman plot would indicate increasing variance of differences over the measurement range. Fitted regression lines, for both the upper and lower half of the plot, has been added to indicate the trend.

Figure 1.5 is an example of cases where the inter-method bias changes over the measurement range. This is known as proportional bias. In both Figures 1.4 and 1.5, the assumptions necessary for further analysis using the limits of agreement are violated.

Figure 1.4 is an example of cases where the inter-method bias changes over the measurement range. This is known as *proportional bias* ?. Both of these cases violate the assumptions necessary for further analysis using limits of agreement ,which shall be discussed later. The plot also can be used to identify outliers. An outlier is an observation that is numerically distant from the rest of the data. Classification thereof is a subjective decision in any analysis, but must be informed by the logic of the formulation. Figure 1.5 is a Bland Altman plot with two conspicuous observations, at the extreme left and right of the plot respectively.

In the Bland-Altman plot, the horizontal displacement of any observation is supported by two independent measurements. Hence any observation , such as the one on the extreme right of figure 1.5, should not be considered an outlier on the basis of a noticeable horizontal displacement from the main cluster. The one on the extreme left should be considered an outlier, as it has a noticeable vertical displacement from the rest of the observations.

Bland and Altman (1999) do not recommend excluding outliers from analyses. However recalculation of the inter-method bias estimate , and further calculations based upon that estimate, are useful for assessing the influence of outliers.(Bland and Altman, 1999) states that "*We usually find that this method of analysis is not too sensitive to one or two large outlying differences.*"

Application of regression techniques to the Bland-Altman plot, and subsequent formal testing for the constant variability of differences is informative. The data set may be divided into two subsets, containing the observations wherein the difference

values are less than and greater than the inter-method bias respectively.

For both of these fits, hypothesis tests for the respective slopes can be performed. While both tests can be considered separately, multiple comparison procedures, such as the Benjamini-Hochberg (Benjamini and Hochberg, 1995) test, should be also be used.




Figure 5: Bland-Altman plot demonstrating the increase of variance over the range.

0.2.4 Adverse features

Estimates for inter-method bias and variance of differences are only meaningful if there is uniform inter-bias and variability throughout the range of measurements. Fulfilment of these assumptions can be checked by visual inspection of the plot. The prototype Bland-Altman plots depicted in Figures 1.4, 1.5 and 1.6 are derived from simulated data, for the purpose of demonstrating how the plot would inform an analyst of features that would adversely affect use of the recommended methodology.




Figure 6: Bland-Altman plot indicating the presence of proportional bias.




Figure 7: Bland-Altman plot indicating the presence of potential outliers.

Figure 1.4 demonstrates how the Bland-Altman plot would indicate increasing variance of differences over the measurement range. Fitted regression lines, for both the upper and lower half of the plot, has been added to indicate the trend. Figure 1.5 is an example of cases where the inter-method bias changes over the measurement range. This is known as proportional bias, and is defined by Ludbrook (1997) as meaning that ‘one method gives values that are higher (or lower) than those from the other by an amount that is proportional to the level of the measured variable’. In both Figures 1.4 and 1.5, the assumptions necessary for further analysis using the limits of agreement are violated.

0.2.5 Variations of The Bland Altman Plot

Variations of the Bland Altman plot is the use of ratios, in the place of differences.

$$D_i = X_i - Y_i \tag{1}$$

Altman and Bland suggest plotting the within subject differences $D = X_1 - X_2$ on the ordinate versus the average of x_1 and x_2 on the abscissa.

0.2.6 Using Bland-Altman Plots

Bland-Altman plots are a powerful graphical methodology for making a visual assessment of the data. Altman and Bland (1983) express the motivation for this plot thusly:

“From this type of plot it is much easier to assess the magnitude of disagreement (both error and bias), spot outliers, and see whether there is any trend, for example an increase in (difference) for high values. This way of plotting the data is a very powerful way of displaying the results of a method comparison study.”

The Bland-Altman plot is simply a scatterplot of the case-wise averages and differences of two methods of measurement. As the objective of the Bland-Altman plot is to advise on the agreement of two methods, it is the case-wise differences that are particularly. Later it will be shown that case-wise differences are the sole component of the next part of the methodology, the limits of agreement.

For creating plots, the case wise-averages fulfil several functions, such as expressing the range over which the values were taken, and assessing whether the assumptions of constant variance holds. Case-wise averages also allow the case-wise differences to be presented on a two-dimensional plot, with better data visualization qualities than a one dimensional plot. Bland and Altman (1986) cautions that it would be the difference against either measurement value instead of their average , as the difference relates to both value.

0.3 Bland Altman Plots

In light of shortcomings associated with scatterplots, Altman and Bland (1983) recommend a further analysis of the data. Firstly differences of measurements of two methods on the same subject should be calculated, and then the average of those measurements (Table 1.1). The averages of the two measurements is considered by Bland and Altman to the best estimate for the unknown true value. Importantly both methods must measure with the same units.

These results are then plotted, with differences on the ordinate and averages on the abscissa (figure 1.2). Altman and Bland (1983) express the motivation for this plot thusly:

”From this type of plot it is much easier to assess the magnitude of disagreement (both error and bias), spot outliers, and see whether there is

any trend, for example an increase in (difference) for high values. This way of plotting the data is a very powerful way of displaying the results of a method comparison study.”

0.4 Conclusions about Existing Methodologies

Scatterplots are recommended by Altman and Bland (1983) for an initial examination of the data, facilitating an initial judgement and helping to identify potential outliers. They are not useful for a thorough examination of the data. O’Brien et al. (1990) notes that data points will tend to cluster around the line of equality, obscuring interpretation.

The Bland Altman methodology is well noted for its ease of use, and can be easily implemented with most software packages. Also it doesn’t require the practitioner to have more than basic statistical training. The plot is quite informative about the variability of the differences over the range of measurements. For example, an inspection of the plot will indicate the ‘fan effect’. They also can be used to detect the presence of an outlier.

Ludbrook (1997, 2002) criticizes these plots on the basis that they presents no information on effect of constant bias or proportional bias. These plots are only practicable when both methods measure in the same units. Hence they are totally unsuitable for conversion problems. The limits of agreement are somewhat arbitrarily constructed. They may or may not be suitable for the data in question. It has been found that the limits given are too wide to be acceptable. There is no guidance on how to deal with outliers. Bland and Altman recognize effect they would have on the limits of agreement, but offer no guidance on how to correct for those effects.

There is no formal testing procedure provided. Rather, it is upon the practitioner

opinion to judge the outcome of the methodology.

Bibliography

- Altman, D. and J. Bland (1983). Measurement in medicine: The analysis of method comparison studies. *Journal of the Royal Statistical Society. Series D (The Statistician)* 32(3), 307–317.
- Benjamini, Y. and Y. Hochberg (1995). Controlling the false discovery rate: A practical and powerful approach to multiple testing. *J. Roy. Statist. Soc. Ser. B* 57(1), 289–300.
- Bland, J. and D. Altman (1986). Statistical methods for assessing agreement between two methods of clinical measurement. *The Lancet* i, 307–310.
- Bland, J. and D. Altman (1999). Measuring agreement in method comparison studies. *Statistical Methods in Medical Research* 8(2), 135–160.
- Dewitte, K., C. Fierens, D. Stckl, and L. M. Thienpont (2002). Application of the Bland Altman plot for interpretation of method-comparison studies: A critical investigation of its practice. *Clinical Chemistry* 48, 799–801.
- Ludbrook, J. (1997). Comparing methods of measurement. *Clinical and Experimental Pharmacology and Physiology* 24, 193–203.
- Ludbrook, J. (2002). Statistical techniques for comparing measurers and methods

of measurement: a critical review. *Clinical and Experimental Pharmacology and Physiology* 29, 527–536.

O'Brien, E., J. Petrie, W. Littler, M. de Swiet, P. L. Padfield, D. Altman, M. Bland, A. Coats, and N. Atkins (1990). The British Hypertension Society protocol for the evaluation of blood pressure measuring devices. *Journal of Hypertension* 8, 607–619.