2_PET_AC_WB_FDG.Adult_PET4

23. januar 2013

Indhold

1 Topogram

1.1 Routine

- mA: 35

- kV: 120

- Topogram length: 1024 mm

- Tube position: Top

1.2 Scan

- mA: 35

- kV: 120

- Delay: 4s

- Topogram length: 1024 mm

- Direction: Craniocaudal

- Tube position: Top

- API: API

- Kernel: 80

- Window: Topogram

2 LavdoseCTWB

2.1 Routine

- Eff. mAs: 50

- kV: 120

- CARE Dose4D: On

- CareDoseType: CareDoseAEC

- CTDlvol: 3.64mGy

- Scan time: 32.530 s

- Delay: 2.000 s

- Slice: 2 mm

- No. of images: Samme som i foerste recon, slet?(y/n)

- Tilt: 0.0 grader

2.2   Scan

- Quality ref. mAs: 50

- Eff. mAs: 50

- kV: 120

- Scan time: 32.530 s

- Rotation time: 0.500 s

- Delay: 2.000 s

- Slice: 2 mm

- Pitch: 0.80

- Direction: Craniocaudal

2.3   Recons

2.3.1   Recon 1

- Series description: tmp LavdoseCTWB RTD

- Slice: 4

- Kernel: B19f

- Window: Abdomen

- Extended FoV: Off

- FoV: 500

- Center X: 0

- Center Y: 0

- Mirroring: None

- Extended CT scale: Standard

- Recon job: Axial

- Recon Axis: Axial

- Image order: Craniocaudal

- Recon increment: 4.000

- No. of images: 245

2.3.2 Recon 2

- Series description: AC lavCT 2 mm

- Slice: 2

- Kernel: B19f

- Window: Abdomen

- Extended FoV: On

- FoV: 780

- Center X: 0

- Center Y: 0

- Mirroring: None

- Extended CT scale: Standard

- Recon job: Axial

- Recon Axis: Axial

- Image order: Craniocaudal

- Recon increment: 2.000

- No. of images: 489

3   Pause

4   PET WB

4.1   Routine

- Isotope: F-18

- Pharm.: Fluorodeoxyglucose

- Inj. Dose: 300000000 Bequerels

- Scan mode: Sinogram

- Scan range: Do not match CT FOV

- No. of beds: Not given. Check recon range

- Scan duration/bed: 2.000000 Minutes

4.2   Scan

- Autoload: On

- Rebinner LUT: Off

- Scan output: Sinogram

- Sinogram mode: false

- Input trigger signal: None

- LLD (keV): n/a

- ULD (keV): n/a

4.3   Recons

4.3.1   Recon 1

- Series description: PET WB Image check HD

- Recon range (bed): 1 to 7

- Output image type: Corrected

- Recon method: Iterative3D

- Iterations: 4

- Subsets: 8

- Image size: 200

- Zoom: 1

- Filter: Gaussian

- FWHM (mm): 4.000000

- Offset X: 0 mm

- Offset Y: 0 mm

- Attenuation correction: Off ()

- Scatter correction: On

- Match CT slice location: On

- Save intermediate data: Off

4.3.2 Recon 2

- Series description: PET WB Corr TrueX+TOF 2 mm

- Recon range (bed): 1 to 7

- Output image type: Corrected

- Recon method: TrueXTOF

- Iterations: 2

- Subsets: 21

- Image size: 400

- Zoom: 1

- Filter: Gaussian

- FWHM (mm): 2.000000

- Offset X: 0 mm

- Offset Y: 0 mm

- Attenuation correction: Off ()

- Scatter correction: On

- Match CT slice location: On

- Save intermediate data: Off