PETCT_Lung.Adult_PET4

3. december 2012

Indhold

1 Topogram

1.1 Routine

- mA: 35

- kV: 120

- Topogram length: 768 mm

- Tube position: Top

1.2 Scan

- mA: 35

- kV: 120

- Delay: 4s

- Topogram length: 768 mm

- Direction: Craniocaudal

- Tube position: Top

- API: API

- Kernel: 80

- Window: Topogram

2 CT Lung

2.1 Routine

- Eff. mAs: 120

- kV: 120

- CARE Dose4D: On

- CareDoseType: CareDoseAEC

- CTDlvol: 9.12mGy

- Scan time: 7.250 s

- Delay: 2.000 s

- Slice: 5 mm

- No. of images: Samme som i foerste recon, slet?(y/n)

- Tilt: 0.0 grader

2.2   Scan

- Quality ref. mAs: 120

- Eff. mAs: 120

- kV: 120

- Scan time: 7.250 s

- Rotation time: 0.500 s

- Delay: 2.000 s

- Slice: 5 mm

- Pitch: 1.40

- Direction: Craniocaudal

2.3   Recons

2.3.1   Recon 1

- Series description: tmp CT Lung RTD

- Slice: 5

- Kernel: B19f

- Window: Mediastinum

- Extended FoV: Off

- FoV: 500

- Center X: 0

- Center Y: 0

- Mirroring: None

- Extended CT scale: Standard

- Recon job: Axial

- Recon Axis: Axial

- Image order: Craniocaudal

- Recon increment: 5.000

- No. of images: 70

2.3.2   Recon 2

- Series description: AC CT Lung

- Slice: 5

- Kernel: B19f

- Window: Mediastinum

- Extended FoV: On

- FoV: 780

- Center X: 0

- Center Y: 0

- Mirroring: None

- Extended CT scale: Standard

- Recon job: Axial

- Recon Axis: Axial

- Image order: Craniocaudal

- Recon increment: 3.000

- No. of images: 116

2.3.3 Recon 3

- Series description: CT Lung 2.0 B31f

- Slice: 2

- Kernel: B31f

- Window: Mediastinum

- Extended FoV: Off

- FoV: 300

- Center X: 0

- Center Y: 0

- Mirroring: None

- Extended CT scale: Standard

- Recon job: Axial

- Recon Axis: Axial

- Image order: Craniocaudal

- Recon increment: 1.500

- No. of images: 232

2.3.4 Recon 4

- Series description: CT LUNG 2.0 B70f

- Slice: 2

- Kernel: B70f

- Window: Lung

- Extended FoV: Off

- FoV: 300

- Center X: 0

- Center Y: 0

- Mirroring: None

- Extended CT scale: Standard

- Recon job: Axial

- Recon Axis: Axial

- Image order: Craniocaudal

- Recon increment: 1.500

- No. of images: 232

3 Pause

4 PET Lung

4.1 Routine

- Isotope: F-18

- Pharm.: Fluorodeoxyglucose

- Inj. Dose: 370000000.000000 Bequerels

- Inj. date (date/month - year): 02/10 - 2008

- Inj. time: 08:48:06

- Scan mode: Sinogram

- Scan range: Match CT FOV

- No. of beds: Not given. Check recon range

- Scan duration/bed: 2.000000 Minutes

4.2 Scan

- Autoload: On

- Rebinner LUT: N/A

- Scan output: Sinogram

- Sinogram mode: false

- Input trigger signal: None

- LLD (keV): N/A

- ULD (keV): N/A

4.3 Recons

4.3.1 Recon 1

- Series description: PET Lung

- Recon range (bed): 1 to 2

- Output image type: Corrected

- Recon method: Iterative3D

- Iterations: 2

- Subsets: 24

- Image size: 200

- Zoom: 1

- Filter: Gaussian

- FWHM (mm): 5

- Offset X: 0 mm

- Offset Y: 0 mm

- Attenuation correction: On (1)

- Scatter correction: On

- Match CT slice location: On

- Save intermediate data: Off