# $3\_PET\_AC\_Brain\_LM\_FDG.Adult$

23. oktober 2012

## Indhold

1	Top	ogram	1
	1.1	Routine	1
	1.2	Scan	1
<b>2</b>	Lav	CT Brain	1
	2.1	Routine	1
	2.2	Scan	2
	2.3	Recons	2
		2.3.1 Recon 1	2
		2.3.2 Recon 2	3
3	Pau	${f se}$	4
4	PET	Γ Brain LM	4
	4.1	Routine	4
	4.2	Scan	4
	4.3	Recons	1
		4.3.1 Recon 1	5

## 1 Topogram

## 1.1 Routine

• mA: 35

• kV: 120

• Topogram length: 256 mm

• Tube position: Lateral

#### 1.2 Scan

• mA: 35

• kV: 120

• Delay: 4s

• Topogram length: 256 mm

• Direction: Craniocaudal

• Tube position: Lateral

• API: API

• Kernel: 20

• Window: Topogram

## 2 Lav CT Brain

## 2.1 Routine

• Eff. mAs: 32

• kV: 120

• CARE Dose4D: Off

• CareDoseType: CareDoseAEC

• CTDlvol: 5.69782mGy

• Scan time: 17.570 s

• Delay: 2.000 s

• Slice: 3 mm

• No. of images: Samme som i foerste recon, slet?(y/n)

• Tilt: 0.0 grader

## 2.2 Scan

• Quality ref. mAs: 350

• Eff. mAs: 32

• kV: 120

• Scan time: 17.570 s

• Rotation time: 0.500 s

• Delay: 2.000 s

• Slice: 3 mm

• Pitch: 0.55

• Direction: Caudocranial

### 2.3 Recons

#### 2.3.1 Recon 1

• Series description: tmp Lav CT Brain RTD

• Slice: 4

• Kernel: H31f

• Window: Cerebrum

• Extended FoV: false

• FoV: 300

• Center X: 0

• Center Y: 0

• Mirroring: None

• Extended CT scale: Standard

• Recon job: Axial

• Recon Axis: Axial

• Image order: Caudocranial

• Recon increment: 4.000

• No. of images: 56

#### 2.3.2 Recon 2

• Series description: AC CT

• Slice: 3

• Kernel: H31f

• Window: Cerebrum

• Extended FoV: false

• FoV: 300

• Center X: 0

• Center Y: 0

• Mirroring: None

• Extended CT scale: Standard

• Recon job: Axial

• Recon Axis: Axial

• Image order: Caudocranial

• Recon increment: 3.000

• No. of images: 74

## 3 Pause

## 4 PET Brain LM

### 4.1 Routine

- Isotope: F-18
- Pharm.: Fluorodeoxyglucose
- $\bullet\,$  Inj. Dose: 185000000.0 Bequerels
- Inj. date (date/month year): 14/05 2012
- Inj. time: 10:40:28
- Scan mode: List mode
- Scan range: Do not match CT FOV
- No. of beds: Not given. Check recon range
- Scan duration/bed: 10.000000 Minutes

#### 4.2 Scan

- Autoload: false
- Rebinner LUT: N/A
- Scan output: List mode
- Sinogram mode: false
- Input trigger signal: None
- LLD (keV): N/A
- ULD (keV): N/A

## 4.3 Recons

#### 4.3.1 Recon 1

- Series description: Brain LM
- Recon range (bed): Automatic/All/dunno
- Output image type: Uncorrected
- Recon method: Iterative3D
- Iterations: 4
- Subsets: 12
- Image size: 400
- Zoom: 1
- Filter: Gaussian
- FWHM (mm): 5
- Offset X: 0 mm
- Offset Y: 0 mm
- Attenuation correction: On (1)
- Scatter correction: false
- Match CT slice location: false
- Save intermediate data: false