# PETCT\_AC\_Brain.Adult

23. oktober 2012

## Indhold

1	Topogram			
	1.1	Routine	1	
	1.2	Scan	1	
<b>2</b>	CT Brain			
	2.1	Routine	1	
	2.2	Scan	2	
	2.3	Recons	2	
		2.3.1 Recon 1	2	
		2.3.2 Recon 2	3	
3	Pau	se	4	
4	PET	Γ Brain	4	
	4.1	Routine	4	
	4.2	Scan	4	
	4.3	Recons	5	
		4.3.1 Recon 1	5	

## 1 Topogram

## 1.1 Routine

• mA: 35

• kV: 120

• Topogram length: 256 mm

• Tube position: Lateral

#### 1.2 Scan

• mA: 35

• kV: 120

• Delay: 4s

• Topogram length: 256 mm

• Direction: Craniocaudal

• Tube position: Lateral

• API: API

• Kernel: 20

• Window: Topogram

## 2 CT Brain

## 2.1 Routine

• Eff. mAs: 170

• kV: 120

• CARE Dose4D: Off

• CareDoseType: CareDoseAEC

 $\bullet$  CTDlvol: 29.597mGy

• Scan time: 20.050 s

• Delay: 2.000 s

• Slice: 3 mm

 $\bullet$  No. of images: Samme som i foerste recon, slet?(y/n)

• Tilt: 0.0 grader

## 2.2 Scan

• Quality ref. mAs: 350

• Eff. mAs: 170

• kV: 120

 $\bullet$  Scan time: 20.050 s

• Rotation time: 1.000 s

• Delay: 2.000 s

• Slice: 3 mm

• Pitch: 1.00

• Direction: Caudocranial

## 2.3 Recons

#### 2.3.1 Recon 1

• Series description: tmp CT Brain RTD

• Slice: 4

• Kernel: H19s

• Window: Cerebrum

• Extended FoV: false

• FoV: 300

• Center X: 0

• Center Y: 0

• Mirroring: None

• Extended CT scale: Standard

• Recon job: Axial

• Recon Axis: Axial

• Image order: Caudocranial

• Recon increment: 4.000

• No. of images: 56

#### 2.3.2 Recon 2

• Series description: AC CT Brain

• Slice: 3

• Kernel: H19s

• Window: Cerebrum

• Extended FoV: false

• FoV: 300

• Center X: 0

• Center Y: 0

• Mirroring: None

• Extended CT scale: Standard

• Recon job: Axial

• Recon Axis: Axial

• Image order: Caudocranial

• Recon increment: 1.500

• No. of images: 148

## 3 Pause

## 4 PET Brain

## 4.1 Routine

- Isotope: F-18
- Pharm.: Fluorodeoxyglucose
- Inj. Dose: 185000000.000000 Bequerels
- Inj. date (date/month year): 02/10 2008
- Inj. time: 07:16:19
- Scan mode: Sinogram
- Scan range: Match CT FOV
- No. of beds: Not given. Check recon range
- Scan duration/bed: 10.000000 Minutes

### 4.2 Scan

- Autoload: true
- Rebinner LUT: N/A
- Scan output: Sinogram
- Sinogram mode: false
- Input trigger signal: None
- LLD (keV): N/A
- $\bullet$  ULD (keV): N/A

## 4.3 Recons

#### 4.3.1 Recon 1

- Series description: PET Brain
- Recon range (bed): Automatic/All/dunno
- Output image type: Corrected
- Recon method: Iterative3D
- Iterations: 4
- Subsets: 24
- Image size: 400
- Zoom: 2.000000
- Filter: Gaussian
- FWHM (mm): 4.000000
- Offset X: 0 mm
- Offset Y: 0 mm
- Attenuation correction: On (1)
- Scatter correction: true
- Match CT slice location: true
- Save intermediate data: false