# PETCT\_Brain.MlAdult\_HD\_PET5

3. december 2012

# Indhold

1	Topogram			
	1.1	Routine	1	
	1.2	Scan	1	
<b>2</b>	CT Brain			
	2.1	Routine	1	
	2.2	Scan	2	
	2.3	Recons	2	
		2.3.1 Recon 1	2	
		2.3.2 Recon 2	3	
3	Pau	se	4	
4	PET	Γ Brain	4	
	4.1	Routine	4	
	4.2	Scan	4	
	4.3	Recons	5	
		4.3.1 Recon 1	5	

# 1 Topogram

# 1.1 Routine

• mA: 35

• kV: 120

• Topogram length: 256 mm

• Tube position: Lateral

#### 1.2 Scan

• mA: 35

• kV: 120

• Delay: 4s

• Topogram length: 256 mm

• Direction: Craniocaudal

• Tube position: Lateral

• API: None

• Kernel: 20s standar

• Window: Topogram Head

# 2 CT Brain

## 2.1 Routine

• Eff. mAs: 380

• kV: 120

• CARE Dose4D: Off

• CareDoseType: CareDoseAEC

• CTDlvol: 59.4292mGy

• Scan time: 16.010 s

• Delay: 4.000 s

• Slice: 3.00 mm

• No. of images: Samme som i foerste recon, slet?(y/n)

• Tilt: 0.0 grader

# 2.2 Scan

• Quality ref. mAs: 380

• Eff. mAs: 380

• kV: 120

 $\bullet$  Scan time: 16.010 s

• Rotation time: 1.000 s

• Delay: 4.000 s

• Slice: 3.00 mm

• Pitch: 0.8000000

• Direction: Caudocranial

## 2.3 Recons

#### 2.3.1 Recon 1

• Series description: AC CT

• Slice: 3.00

• Kernel: H19s PET very smooth

• Window: Cerebrum

• Extended FoV: Off

• FoV: 300

• Center X: 0

• Center Y: 0

• Mirroring: None

• Extended CT scale: Standard

• Recon job: Axial

• Recon Axis: Axial

• Image order: Caudocranial

• Recon increment: 1.500

• No. of images: 148

#### 2.3.2 Recon 2

• Series description: CT Brain 3.0 H31s

• Slice: 3.00

• Kernel: H31s medium smooth +

• Window: Cerebrum

• Extended FoV: Off

• FoV: 250

• Center X: 0

• Center Y: 0

• Mirroring: None

• Extended CT scale: Standard

• Recon job: Axial

• Recon Axis: Axial

• Image order: Caudocranial

• Recon increment: 1.500

• No. of images: 148

## 3 Pause

# 4 PET Brain

## 4.1 Routine

• Isotope: F-18

• Pharm.: FDG

• Inj. Dose: 5 Millicuries

 $\bullet\,$  Inj. date (date/month - year): 30/10 - 2006

• Inj. time: 08:50:00

• Scan mode: Sinogram

• Scan range: Match CT FOV

• No. of beds: 1

• Scan duration/bed: 10 Minutes

#### 4.2 Scan

• Autoload: On

• Rebinner LUT: Off

• Scan output: Sinogram

• Sinogram mode: Trues

• Input trigger signal: None

• LLD (keV): 435

• ULD (keV): 650

## 4.3 Recons

#### 4.3.1 Recon 1

- Series description: PET Brain
- Recon range (bed): Automatic/All/dunno
- Output image type: Corrected
- Recon method: TrueX
- Iterations: 3
- Subsets: 21
- Image size: 336
- Zoom: 2
- Filter: Gaussian
- FWHM (mm): 4
- Offset X: 0 mm
- Offset Y: 0 mm
- Attenuation correction: On (1)
- Scatter correction: On
- Match CT slice location: On
- Save intermediate data: Off